


# 2,4,6-Tribromophenol

Response Ratio




$$\text{Response} = 1.873\text{e-}001 * \text{Amt} - 9.062\text{e-}002$$

Coef of Det (r^2) = 0.996666 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN062921.M

Calibration Table Last Updated: Tue Jun 29 19:13:09 2021


# 4,6-Dinitro-2-methylphenol



$R = 1.715e-003 A^2 + 5.263e-003 A - 2.882e-003$   
 Coef of Det ( $r^2$ ) = 0.999845    Curve Fit: Quadratic  
 Method Name:    Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN062921.M  
 Calibration Table Last Updated: Tue Jun 29 19:13:09 2021

# Atrazine

Response Ratio



$$\text{Response} = 2.220\text{e-}001 * \text{Amt} - 9.423\text{e-}002$$


Coef of Det ( $r^2$ ) = 0.997986 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN062921.M

Calibration Table Last Updated: Tue Jun 29 19:13:09 2021

# Pentachlorophenol


Response Ratio



$R = 3.795e-003 A^2 + 2.975e-002 A - 1.916e-002$   
 Coef of Det ( $r^2$ ) = 0.999575    Curve Fit: Quadratic  
 Method Name:    Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN062921.M  
 Calibration Table Last Updated: Tue Jun 29 19:13:09 2021

# Bis(2-ethylhexyl)phthalate

Response Ratio



$$\text{Response} = 7.822\text{e-}001 * \text{Amt} - 5.118\text{e-}001$$

Coef of Det ( $r^2$ ) = 0.990015 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN062921.M

Calibration Table Last Updated: Tue Jun 29 19:13:09 2021