

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062921\
 Data File : BN015187.D
 Acq On : 29 Jun 2021 14:50
 Operator : CG/JU
 Sample : SSTDICC0.1
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1

Quant Time: Jun 29 16:43:10 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN062921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 29 16:42:28 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.732	152	16897	0.400	ng	# 0.00
7) Naphthalene-d8	10.505	136	75259	0.400	ng	# 0.00
13) Acenaphthene-d10	14.346	164	34775	0.400	ng	0.00
19) Phenanthrene-d10	17.091	188	60165	0.400	ng	0.00
29) Chrysene-d12	21.303	240	54221	0.400	ng	# 0.00
35) Perylene-d12	23.557	264	56184	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.343	112	4056	0.078	ng	0.00
5) Phenol-d6	6.902	99	5948	0.090	ng	0.00
8) Nitrobenzene-d5	8.870	82	6654	0.117	ng	0.00
11) 2-Methylnaphthalene-d10	12.088	152	12614	0.098	ng	0.00
14) 2,4,6-Tribromophenol	15.843	330	648	0.040	ng	0.00
15) 2-Fluorobiphenyl	12.975	172	15208	0.115	ng	0.00
27) Fluoranthene-d10	19.138	212	16512	0.087	ng	0.00
31) Terphenyl-d14	19.744	244	11068	0.082	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.352	88	553	0.078	ng	# 1
6) bis(2-Chloroethyl)ether	7.169	93	5651	0.109	ng	# 82
9) Naphthalene	10.543	128	21917	0.104	ng	98
10) Hexachlorobutadiene	10.847	225	4468	0.107	ng	# 98
12) 2-Methylnaphthalene	12.164	142	13685	0.096	ng	# 90
16) Acenaphthylene	14.067	152	16468	0.094	ng	98
17) Acenaphthene	14.409	154	10796	0.100	ng	96
18) Fluorene	15.399	166	12061	0.091	ng	99
21) 4-Bromophenyl-phenylether	16.288	248	3867	0.099	ng	# 77
22) Hexachlorobenzene	16.410	284	5115	0.125	ng	# 91
23) Atrazine	16.556	200	1519	0.044	ng	# 92
25) Phenanthrene	17.140	178	20093	0.109	ng	99
26) Anthracene	17.225	178	15137	0.084	ng	99
28) Fluoranthene	19.168	202	19342	0.093	ng	# 97
30) Pyrene	19.530	202	18383	0.086	ng	99
32) Benzo(a)anthracene	21.292	228	16353	0.080	ng	98
33) Chrysene	21.345	228	21107	0.108	ng	97
34) Bis(2-ethylhexyl)phtha...	21.228	149	4460	0.031	ng	# 84
36) Indeno(1,2,3-cd)pyrene	25.820	276	20869	0.111	ng	# 89
37) Benzo(b)fluoranthene	22.886	252	20845	0.099	ng	# 91
38) Benzo(k)fluoranthene	22.927	252	22986	0.113	ng	# 95
39) Benzo(a)pyrene	23.461	252	19663	0.106	ng	# 81
40) Dibenzo(a,h)anthracene	25.837	278	16361	0.108	ng	# 91
41) Benzo(g,h,i)perylene	26.508	276	17041	0.105	ng	# 89


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062921\
Data File : BN015187.D
Acq On : 29 Jun 2021 14:50
Operator : CG/JU
Sample : SSTDIC0.1
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1

Quant Time: Jun 29 16:43:10 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN062921.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 29 16:42:28 2021
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.732 min Scan# 483
 Delta R.T. -0.000 min
 Lab File: BN015187.D
 Acq: 29 Jun 2021 14:50




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1


Tgt Ion:152 Resp: 16897
 Ion Ratio Lower Upper
 152 100
 150 156.5 119.0 178.6
 115 61.1 62.2 93.4#



#2
 1,4-Dioxane
 Concen: 0.078 ng
 RT: 3.352 min Scan# 8
 Delta R.T. -0.000 min
 Lab File: BN015187.D
 Acq: 29 Jun 2021 14:50

Tgt Ion: 88 Resp: 553
 Ion Ratio Lower Upper
 88 100
 43 38.0 12.9 19.3#
 58 167.3 38.1 57.1#









#4
2-Fluorophenol
Concen: 0.078 ng
RT: 5.343 min Scan# 224
Delta R.T. -0.000 min
Lab File: BN015187.D
Acq: 29 Jun 2021 14:50

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1

Tgt Ion: 112 Resp: 4056




Ion	Ratio	Lower	Upper
112	100		
64	56.9	39.0	58.6
63	30.4	20.6	30.8




#5
Phenol-d6
Concen: 0.090 ng
RT: 6.902 min Scan# 393
Delta R.T. -0.000 min
Lab File: BN015187.D
Acq: 29 Jun 2021 14:50

Tgt Ion: 99 Resp: 5948

Ion	Ratio	Lower	Upper
99	100		
42	18.0	8.9	13.3#
71	31.9	35.6	53.4#









#6
bis(2-Chloroethyl)ether
Concen: 0.109 ng
RT: 7.169 min Scan# 422
Delta R.T. -0.000 min
Lab File: BN015187.D
Acq: 29 Jun 2021 14:50




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion: 93 Resp: 5651
Ion Ratio Lower Upper
93 100
63 77.0 46.6 70.0#
95 32.4 28.6 42.8



#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.505 min Scan# 719
Delta R.T. -0.000 min
Lab File: BN015187.D
Acq: 29 Jun 2021 14:50

Tgt Ion:136 Resp: 75259
Ion Ratio Lower Upper
136 100
137 11.0 9.8 14.8
54 5.4 5.4 8.2#
68 4.4 4.9 7.3#









#8
Nitrobenzene-d5
Concen: 0.117 ng
RT: 8.870 min Scan# 590
Delta R.T. -0.000 min
Lab File: BN015187.D
Acq: 29 Jun 2021 14:50




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion: 82 Resp: 6654
Ion Ratio Lower Upper
82 100
128 36.2 31.6 47.4
54 53.2 29.0 43.4#



#9
Naphthalene
Concen: 0.104 ng
RT: 10.543 min Scan# 722
Delta R.T. -0.000 min
Lab File: BN015187.D
Acq: 29 Jun 2021 14:50

Tgt Ion: 128 Resp: 21917
Ion Ratio Lower Upper
128 100
129 11.2 9.1 13.7
127 13.6 11.8 17.6









#10
Hexachlorobutadiene
Concen: 0.107 ng
RT: 10.847 min Scan# 746
Delta R.T. -0.000 min
Lab File: BN015187.D
Acq: 29 Jun 2021 14:50




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:225 Resp: 4468
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.4 50.2 75.2



#11
2-Methylnaphthalene-d10
Concen: 0.098 ng
RT: 12.088 min Scan# 927
Delta R.T. -0.000 min
Lab File: BN015187.D
Acq: 29 Jun 2021 14:50

Tgt Ion:152 Resp: 12614
Ion Ratio Lower Upper
152 100
151 20.4 16.7 25.1









#12
2-Methylnaphthalene
Concen: 0.096 ng
RT: 12.164 min Scan# 944
Delta R.T. -0.000 min
Lab File: BN015187.D
Acq: 29 Jun 2021 14:50




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:142 Resp: 13685
Ion Ratio Lower Upper
142 100
141 87.5 72.3 108.5
115 35.0 40.7 61.1#



#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.346 min Scan# 1166
Delta R.T. -0.000 min
Lab File: BN015187.D
Acq: 29 Jun 2021 14:50

Tgt Ion:164 Resp: 34775
Ion Ratio Lower Upper
164 100
162 103.4 83.8 125.6
160 53.0 37.4 56.0









#14
2,4,6-Tribromophenol
Concen: 0.040 ng
RT: 15.843 min Scan# 1284
Delta R.T. -0.000 min
Lab File: BN015187.D
Acq: 29 Jun 2021 14:50




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:330 Resp: 648
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 36.1 22.4 33.6#



#15
2-Fluorobiphenyl
Concen: 0.115 ng
RT: 12.975 min Scan# 1058
Delta R.T. -0.000 min
Lab File: BN015187.D
Acq: 29 Jun 2021 14:50

Tgt Ion:172 Resp: 15208
Ion Ratio Lower Upper
172 100
171 33.8 30.6 45.8
170 21.8 20.2 30.4









#16
Acenaphthylene
Concen: 0.094 ng
RT: 14.067 min Scan# 1144
Delta R.T. -0.000 min
Lab File: BN015187.D
Acq: 29 Jun 2021 14:50




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:152 Resp: 16468
Ion Ratio Lower Upper
152 100
151 19.4 16.7 25.1
153 12.8 10.5 15.7



#17
Acenaphthene
Concen: 0.100 ng
RT: 14.409 min Scan# 1171
Delta R.T. -0.000 min
Lab File: BN015187.D
Acq: 29 Jun 2021 14:50

Tgt Ion:154 Resp: 10796
Ion Ratio Lower Upper
154 100
153 114.4 90.9 136.3
152 65.6 45.8 68.8









#18
Fluorene
Concen: 0.091 ng
RT: 15.399 min Scan# 1249
Delta R.T. -0.000 min
Lab File: BN015187.D
Acq: 29 Jun 2021 14:50




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:	166	Resp:	12061
Ion Ratio	Lower	Upper	
166	100		
165	97.9	79.2	118.8
167	13.5	10.7	16.1



#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.091 min Scan# 1386
Delta R.T. -0.000 min
Lab File: BN015187.D
Acq: 29 Jun 2021 14:50

Tgt Ion:	188	Resp:	60165
Ion Ratio	Lower	Upper	
188	100		
94	0.0	0.0	0.0
80	13.0	9.0	13.4









#21
4-Bromophenyl-phenylether
Concen: 0.099 ng
RT: 16.288 min Scan# 1320
Delta R.T. -0.000 min
Lab File: BN015187.D
Acq: 29 Jun 2021 14:50




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:	248	Resp:	3867
Ion	Ratio	Lower	Upper
248	100		
250	92.9	86.2	129.4
141	72.8	36.2	54.2#



#22
Hexachlorobenzene
Concen: 0.125 ng
RT: 16.410 min Scan# 1330
Delta R.T. -0.000 min
Lab File: BN015187.D
Acq: 29 Jun 2021 14:50

Tgt Ion:	284	Resp:	5115
Ion	Ratio	Lower	Upper
284	100		
142	35.9	22.2	33.2#
249	31.1	26.6	39.8









#23
Atrazine
Concen: 0.044 ng
RT: 16.556 min Scan# 1342
Delta R.T. -0.000 min
Lab File: BN015187.D
Acq: 29 Jun 2021 14:50




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion: 200 Resp: 1519
Ion Ratio Lower Upper
200 100
173 34.3 21.7 32.5#
215 50.6 43.1 64.7



#25
Phenanthrene
Concen: 0.109 ng
RT: 17.140 min Scan# 1390
Delta R.T. -0.000 min
Lab File: BN015187.D
Acq: 29 Jun 2021 14:50

Tgt Ion: 178 Resp: 20093
Ion Ratio Lower Upper
178 100
176 19.0 15.6 23.4
179 15.6 12.1 18.1









#26
 Anthracene
 Concen: 0.084 ng
 RT: 17.225 min Scan# 1397
 Delta R.T. -0.000 min
 Lab File: BN015187.D
 Acq: 29 Jun 2021 14:50




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1


Tgt Ion:178 Resp: 15137
 Ion Ratio Lower Upper
 178 100
 176 18.1 15.1 22.7
 179 15.2 12.2 18.2



#27
 Fluoranthene-d10
 Concen: 0.087 ng
 RT: 19.138 min Scan# 1688
 Delta R.T. 0.005 min
 Lab File: BN015187.D
 Acq: 29 Jun 2021 14:50

Tgt Ion:212 Resp: 16512
 Ion Ratio Lower Upper
 212 100
 106 14.0 7.0 10.6#
 104 7.9 4.1 6.1#









#28
Fluoranthene
Concen: 0.093 ng
RT: 19.168 min Scan# 1694
Delta R.T. 0.005 min
Lab File: BN015187.D
Acq: 29 Jun 2021 14:50




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:202 Resp: 19342
Ion Ratio Lower Upper
202 100
101 11.3 6.6 10.0#
203 17.0 13.9 20.9



#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.303 min Scan# 1985
Delta R.T. -0.000 min
Lab File: BN015187.D
Acq: 29 Jun 2021 14:50

Tgt Ion:240 Resp: 54221
Ion Ratio Lower Upper
240 100
120 13.4 4.2 6.4#
236 26.9 21.4 32.2









#30
Pyrene
Concen: 0.086 ng
RT: 19.530 min Scan# 1767
Delta R.T. -0.000 min
Lab File: BN015187.D
Acq: 29 Jun 2021 14:50




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:	202	Resp:	18383
Ion	Ratio	Lower	Upper
202	100		
200	20.7	16.9	25.3
203	17.6	13.9	20.9



#31
Terphenyl-d14
Concen: 0.082 ng
RT: 19.744 min Scan# 1810
Delta R.T. -0.000 min
Lab File: BN015187.D
Acq: 29 Jun 2021 14:50

Tgt Ion:	244	Resp:	11068
Ion	Ratio	Lower	Upper
244	100		
212	8.2	6.3	9.5
122	12.4	6.2	9.2#









#32
Benzo(a)anthracene
Concen: 0.080 ng
RT: 21.292 min Scan# 1984
Delta R.T. -0.000 min
Lab File: BN015187.D
Acq: 29 Jun 2021 14:50




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:228 Resp: 16353
Ion Ratio Lower Upper
228 100
226 26.8 23.0 34.6
229 19.7 15.7 23.5



#33
Chrysene
Concen: 0.108 ng
RT: 21.345 min Scan# 1989
Delta R.T. -0.000 min
Lab File: BN015187.D
Acq: 29 Jun 2021 14:50

Tgt Ion:228 Resp: 21107
Ion Ratio Lower Upper
228 100
226 29.1 25.0 37.6
229 20.1 15.9 23.9









#34
Bis(2-ethylhexyl)phthalate
Concen: 0.031 ng
RT: 21.228 min Scan# 1978
Delta R.T. -0.000 min
Lab File: BN015187.D
Acq: 29 Jun 2021 14:50




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:149 Resp: 4460
Ion Ratio Lower Upper
149 100
167 26.0 28.9 43.3#
279 3.3 4.9 7.3#



#35
Perylene-d12
Concen: 0.400 ng
RT: 23.557 min Scan# 2425
Delta R.T. -0.000 min
Lab File: BN015187.D
Acq: 29 Jun 2021 14:50

Tgt Ion:264 Resp: 56184
Ion Ratio Lower Upper
264 100
260 25.2 20.6 31.0
265 24.5 21.4 32.2









#36
Indeno(1,2,3-cd)pyrene
Concen: 0.111 ng
RT: 25.820 min Scan# 3086
Delta R.T. 0.003 min
Lab File: BN015187.D
Acq: 29 Jun 2021 14:50




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:276 Resp: 20869
Ion Ratio Lower Upper
276 100
138 24.8 10.0 15.0#
277 24.9 20.4 30.6



#37
Benzo(b)fluoranthene
Concen: 0.099 ng
RT: 22.886 min Scan# 2229
Delta R.T. 0.003 min
Lab File: BN015187.D
Acq: 29 Jun 2021 14:50

Tgt Ion:252 Resp: 20845
Ion Ratio Lower Upper
252 100
253 23.4 18.5 27.7
125 19.0 5.8 8.6#









#38
Benzo(k)fluoranthene
Concen: 0.113 ng
RT: 22.927 min Scan# 2241
Delta R.T. -0.000 min
Lab File: BN015187.D
Acq: 29 Jun 2021 14:50




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:252 Resp: 22986
Ion Ratio Lower Upper
252 100
253 22.8 19.1 28.7
125 12.4 5.4 8.0#



#39
Benzo(a)pyrene
Concen: 0.106 ng
RT: 23.461 min Scan# 2397
Delta R.T. 0.003 min
Lab File: BN015187.D
Acq: 29 Jun 2021 14:50

Tgt Ion:252 Resp: 19663
Ion Ratio Lower Upper
252 100
253 25.0 19.4 29.2
125 32.6 7.4 11.2#









#40
Dibenzo(a,h)anthracene
Concen: 0.108 ng
RT: 25.837 min Scan# 3091
Delta R.T. 0.003 min
Lab File: BN015187.D
Acq: 29 Jun 2021 14:50

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1

Tgt Ion:278 Resp: 16361
Ion Ratio Lower Upper
278 100
139 19.3 9.0 13.4#
279 24.5 21.0 31.4



#41
Benzo(g,h,i)perylene
Concen: 0.105 ng
RT: 26.508 min Scan# 3287
Delta R.T. 0.003 min
Lab File: BN015187.D
Acq: 29 Jun 2021 14:50

Tgt Ion:276 Resp: 17041
Ion Ratio Lower Upper
276 100
277 24.2 19.7 29.5
138 24.9 9.8 14.6#

