

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062922\
 Data File : BN020568.D
 Acq On : 29 Jun 2022 14:15
 Operator : CG/JU
 Sample : PB145836BL
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK836

Quant Time: Jun 29 18:52:53 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN062422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 24 00:17:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.737	152	3624	0.400	ng/ul	-0.02
4) Naphthalene-d8	10.517	136	11786	0.400	ng/ul	-0.02
9) Acenaphthene-d10	14.366	164	6880	0.400	ng/ul	-0.01
13) Phenanthrene-d10	17.110	188	13365	0.400	ng/ul	0.00
17) Chrysene-d12	21.309	240	8353	0.400	ng/ul	0.00
23) Perylene-d12	23.583	264	8015	0.400	ng/ul	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.193	96	25063	5.970	ng/ul	-0.01
6) 2-Methylnaphthalene-d10	12.123	152	5522	0.286	ng/ul	0.00
18) Fluoranthene-d10	19.144	212	11286	0.410	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062922\
Data File : BN020568.D
Acq On : 29 Jun 2022 14:15
Operator : CG/JU
Sample : PB145836BL
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK836

Quant Time: Jun 29 18:52:53 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN062422.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 24 00:17:34 2022
Response via : Initial Calibration

