

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062922\
 Data File : BN020571.D
 Acq On : 29 Jun 2022 16:04
 Operator : CG/JU
 Sample : N3355-04
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 EY0C5

Quant Time: Jun 29 18:59:18 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN062422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 24 00:17:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.737	152	3414	0.400	ng/ul	-0.02
4) Naphthalene-d8	10.518	136	11404	0.400	ng/ul	-0.02
9) Acenaphthene-d10	14.368	164	6822	0.400	ng/ul	-0.01
13) Phenanthrene-d10	17.108	188	13506	0.400	ng/ul	-0.01
17) Chrysene-d12	21.307	240	10519	0.400	ng/ul	0.00
23) Perylene-d12	23.581	264	9748	0.400	ng/ul	#-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.193	96	19320	4.885	ng/ul	-0.01
6) 2-Methylnaphthalene-d10	12.118	152	6409	0.342	ng/ul	0.00
18) Fluoranthene-d10	19.142	212	14577	0.420	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062922\
Data File : BN020571.D
Acq On : 29 Jun 2022 16:04
Operator : CG/JU
Sample : N3355-04
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
EY0C5

Quant Time: Jun 29 18:59:18 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN062422.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 24 00:17:34 2022
Response via : Initial Calibration

