

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN063021\
 Data File : BN015241.D
 Acq On : 01 Jul 2021 02:28
 Operator : CG/JU
 Sample : M2783-03DL 2X
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RE134D2-20210617DL

Manual Integrations
 APPROVED

mohammad
 7/1/2021 2:54:17 PM

Quant Time: Jul 01 03:00:26 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN063021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 30 15:55:37 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.732	152	26191	0.400	ng	# 0.00
7) Naphthalene-d8	10.492	136	119535	0.400	ng	# 0.00
13) Acenaphthene-d10	14.346	164	62115	0.400	ng	0.00
19) Phenanthrene-d10	17.091	188	120554	0.400	ng	0.00
29) Chrysene-d12	21.292	240	99538	0.400	ng	#-0.01
35) Perylene-d12	23.537	264	84441	0.400	ng	-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.362	112	5194	0.078	ng	0.00
5) Phenol-d6	6.911	99	3527	0.042	ng	0.00
8) Nitrobenzene-d5	8.870	82	13185	0.195	ng	0.00
11) 2-Methylnaphthalene-d10	12.084	152	31281	0.158	ng	0.00
14) 2,4,6-Tribromophenol	15.843	330	2648	0.159	ng	0.00
15) 2-Fluorobiphenyl	12.963	172	47880	0.190	ng	0.00
27) Fluoranthene-d10	19.128	212	59771	0.168	ng	0.00
31) Terphenyl-d14	19.734	244	55055	0.236	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.352	88	48962m	3.987	ng	Qvalue
34) Bis(2-ethylhexyl)phtha...	21.218	149	9755	0.087	ng	# 87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN063021\
Data File : BN015241.D
Acq On : 01 Jul 2021 02:28
Operator : CG/JU
Sample : M2783-03DL 2X
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE134D2-20210617DL

Manual Integrations
APPROVED

mohammad
7/1/2021 2:54:17 PM

Quant Time: Jul 01 03:00:26 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN063021.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 30 15:55:37 2021
Response via : Initial Calibration

