

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN063021\
 Data File : BN015252.D
 Acq On : 01 Jul 2021 10:16
 Operator : CG/JU
 Sample : M2783-15
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RE132D1-20210618

Manual Integrations
 APPROVED

mohammad
 7/1/2021 2:54:33 PM

Quant Time: Jul 01 11:40:55 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN063021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 01 09:06:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.732	152	22561	0.400	ng	# 0.00
7) Naphthalene-d8	10.492	136	103497	0.400	ng	# 0.00
13) Acenaphthene-d10	14.346	164	53663	0.400	ng	0.00
19) Phenanthrene-d10	17.091	188	104592	0.400	ng	0.00
29) Chrysene-d12	21.292	240	86833	0.400	ng	#-0.01
35) Perylene-d12	23.540	264	71947	0.400	ng	-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.371	112	8403	0.146	ng	0.00
5) Phenol-d6	6.911	99	5579	0.077	ng	0.00
8) Nitrobenzene-d5	8.870	82	25413	0.438	ng	0.00
11) 2-Methylnaphthalene-d10	12.088	152	55972	0.327	ng	0.00
14) 2,4,6-Tribromophenol	15.843	330	5573	0.461	ng	0.00
15) 2-Fluorobiphenyl	12.963	172	82744	0.380	ng	0.00
27) Fluoranthene-d10	19.128	212	115438	0.374	ng	0.00
31) Terphenyl-d14	19.734	244	101258	0.498	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.361	88	26496m	2.505	ng	Qvalue
34) Bis(2-ethylhexyl)phtha...	21.218	149	18471	0.323	ng	# 87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN063021\
Data File : BN015252.D
Acq On : 01 Jul 2021 10:16
Operator : CG/JU
Sample : M2783-15
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE132D1-20210618

Manual Integrations
APPROVED

mohammad
7/1/2021 2:54:33 PM

Quant Time: Jul 01 11:40:55 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN063021.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 01 09:06:31 2021
Response via : Initial Calibration

