


2,4,6-Tribromophenol

Response Ratio




$$\text{Response} = 1.723\text{e-}001 * \text{Amt} - 9.453\text{e-}002$$

Coef of Det (r^2) = 0.995659 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN063021.M

Calibration Table Last Updated: Thu Jul 01 09:06:31 2021


4,6-Dinitro-2-methylphenol



$R = 1.034e-003 A^2 + 5.917e-003 A - 3.054e-003$
 Coef of Det (r^2) = 0.999530 Curve Fit: Quadratic
 Method Name: Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN063021.M
 Calibration Table Last Updated: Thu Jul 01 09:06:31 2021

Atrazine

Response Ratio



$$\text{Response} = 2.192\text{e-}001 * \text{Amt} - 9.464\text{e-}002$$


Coef of Det (r^2) = 0.998178 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN063021.M

Calibration Table Last Updated: Thu Jul 01 09:06:31 2021

Pentachlorophenol


Response Ratio



$R = 3.120e-003 A^2 + 4.325e-002 A - 2.733e-002$
 Coef of Det (r^2) = 0.998804 Curve Fit: Quadratic
 Method Name: Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN063021.M
 Calibration Table Last Updated: Thu Jul 01 09:06:31 2021

Bis(2-ethylhexyl)phthalate

Response Ratio



$R = 3.154e-002 A^2 + 4.381e-001 A - 1.621e-001$
 Coef of Det (r^2) = 0.998672 Curve Fit: Quadratic
 Method Name: Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN063021.M
 Calibration Table Last Updated: Thu Jul 01 09:06:31 2021