

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN063021\
 Data File : BN015263.D
 Acq On : 01 Jul 2021 16:54
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC

Quant Time: Jul 01 17:27:25 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN063021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 01 09:06:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.735	152	20335	0.400	ng	# 0.00
7) Naphthalene-d8	10.495	136	90393	0.400	ng	0.00
13) Acenaphthene-d10	14.345	164	47501	0.400	ng	0.00
19) Phenanthrene-d10	17.090	188	90306	0.400	ng	0.00
29) Chrysene-d12	21.291	240	76687	0.400	ng	#-0.01
35) Perylene-d12	23.539	264	66071	0.400	ng	-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.356	112	22891	0.442	ng	0.00
5) Phenol-d6	6.905	99	27989	0.429	ng	0.00
8) Nitrobenzene-d5	8.873	82	23178	0.458	ng	0.00
11) 2-Methylnaphthalene-d10	12.087	152	58497	0.391	ng	0.00
14) 2,4,6-Tribromophenol	15.842	330	5648	0.496	ng	0.00
15) 2-Fluorobiphenyl	12.962	172	74710	0.387	ng	0.00
27) Fluoranthene-d10	19.132	212	102026	0.383	ng	0.00
31) Terphenyl-d14	19.733	244	74904	0.417	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.355	88	4116	0.432	ng	# 62
3) n-Nitrosodimethylamine	3.696	42	7095	0.397	ng	# 77
6) bis(2-Chloroethyl)ether	7.163	93	26686	0.416	ng	# 85
9) Naphthalene	10.546	128	103113	0.398	ng	97
10) Hexachlorobutadiene	10.837	225	19458	0.398	ng	# 99
12) 2-Methylnaphthalene	12.158	142	67514	0.406	ng	# 90
16) Acenaphthylene	14.066	152	82090	0.390	ng	98
17) Acenaphthene	14.408	154	57994	0.393	ng	98
18) Fluorene	15.398	166	69950	0.389	ng	99
20) 4,6-Dinitro-2-methylph...	15.487	198	231	0.331	ng	97
21) 4-Bromophenyl-phenylether	16.287	248	21824	0.381	ng	# 86
22) Hexachlorobenzene	16.409	284	24857	0.382	ng	# 92
23) Atrazine	16.555	200	13992	0.455	ng	94
24) Pentachlorophenol	16.750	266	3588	0.563	ng	100
25) Phenanthrene	17.127	178	113645	0.391	ng	99
26) Anthracene	17.224	178	95583	0.383	ng	99
28) Fluoranthene	19.162	202	125322	0.401	ng	# 97
30) Pyrene	19.525	202	123593	0.424	ng	99
32) Benzo(a)anthracene	21.280	228	101844	0.387	ng	96
33) Chrysene	21.334	228	113318	0.395	ng	97
34) Bis(2-ethylhexyl)phtha...	21.217	149	40691	0.573	ng	# 87
36) Indeno(1,2,3-cd)pyrene	25.798	276	83304	0.352	ng	# 88
37) Benzo(b)fluoranthene	22.865	252	106346	0.411	ng	# 93
38) Benzo(k)fluoranthene	22.909	252	107308	0.388	ng	# 94
39) Benzo(a)pyrene	23.443	252	92096	0.389	ng	# 86
40) Dibenzo(a,h)anthracene	25.812	278	64042	0.345	ng	# 91
41) Benzo(g,h,i)perylene	26.486	276	71499	0.349	ng	# 89


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN063021\
Data File : BN015263.D
Acq On : 01 Jul 2021 16:54
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 37 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Quant Time: Jul 01 17:27:25 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN063021.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 01 09:06:31 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.735 min Scan# 505
Delta R.T. 0.003 min
Lab File: BN015263.D
Acq: 01 Jul 2021 16:54




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion:152 Resp: 20335
Ion Ratio Lower Upper
152 100
150 154.2 119.0 178.6
115 59.7 62.2 93.4#




#2
1,4-Dioxane
Concen: 0.432 ng
RT: 3.355 min Scan# 30
Delta R.T. 0.003 min
Lab File: BN015263.D
Acq: 01 Jul 2021 16:54

Tgt Ion: 88 Resp: 4116
Ion Ratio Lower Upper
88 100
43 23.3 12.9 19.3#
58 77.7 38.1 57.1#



Abundance Scan 45 (3.693 min): BN015219.D\data.ms (-39)





#3
n-Nitrosodimethylamine
Concen: 0.397 ng
RT: 3.696 min Scan# 67
Delta R.T. 0.003 min
Lab File: BN015263.D
Acq: 01 Jul 2021 16:54


Tgt Ion: 42 Resp: 7095
Ion Ratio Lower Upper
42 100
74 152.7 149.5 224.3
44 7.9 4.2 6.2#

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Abundance Scan 67 (3.696 min): BN015263.D\data.ms



Abundance Scan 67 (3.696 min): BN015263.D\data.ms (-56)





Abundance Scan 226 (5.362 min): BN015219.D\data.ms (-21)




#4
2-Fluorophenol
Concen: 0.442 ng
RT: 5.356 min Scan# 247
Delta R.T. -0.006 min
Lab File: BN015263.D
Acq: 01 Jul 2021 16:54


Tgt Ion: 112 Resp: 22891
Ion Ratio Lower Upper
112 100
64 56.2 39.0 58.6
63 27.9 20.6 30.8

Abundance Scan 247 (5.356 min): BN015263.D\data.ms



Abundance Scan 247 (5.356 min): BN015263.D\data.ms (-23)









#5
Phenol-d6
Concen: 0.429 ng
RT: 6.905 min Scan# 415
Delta R.T. 0.003 min
Lab File: BN015263.D
Acq: 01 Jul 2021 16:54




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion: 99 Resp: 27989
Ion Ratio Lower Upper
99 100
42 16.8 8.9 13.3#
71 30.1 35.6 53.4#



#6
bis(2-Chloroethyl)ether
Concen: 0.416 ng
RT: 7.163 min Scan# 443
Delta R.T. -0.006 min
Lab File: BN015263.D
Acq: 01 Jul 2021 16:54

Tgt Ion: 93 Resp: 26686
Ion Ratio Lower Upper
93 100
63 74.6 46.6 70.0#
95 32.9 28.6 42.8









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.495 min Scan# 740
Delta R.T. 0.003 min
Lab File: BN015263.D
Acq: 01 Jul 2021 16:54




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:136	Resp:	90393
Ion Ratio	Lower	Upper
136	100	
137	11.0	9.8 14.8
54	7.1	5.4 8.2
68	6.0	4.9 7.3



#8
Nitrobenzene-d5
Concen: 0.458 ng
RT: 8.873 min Scan# 612
Delta R.T. 0.003 min
Lab File: BN015263.D
Acq: 01 Jul 2021 16:54

Tgt Ion: 82	Resp:	23178
Ion Ratio	Lower	Upper
82	100	
128	45.8	31.6 47.4
54	44.4	29.0 43.4









#9
Naphthalene
Concen: 0.398 ng
RT: 10.546 min Scan# 744
Delta R.T. 0.003 min
Lab File: BN015263.D
Acq: 01 Jul 2021 16:54




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:128 Resp: 103113
Ion Ratio Lower Upper
128 100
129 10.9 9.1 13.7
127 12.8 11.8 17.6



#10
Hexachlorobutadiene
Concen: 0.398 ng
RT: 10.837 min Scan# 767
Delta R.T. 0.003 min
Lab File: BN015263.D
Acq: 01 Jul 2021 16:54

Tgt Ion:225 Resp: 19458
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.8 50.2 75.2









#11
2-Methylnaphthalene-d10
Concen: 0.391 ng
RT: 12.087 min Scan# 950
Delta R.T. -0.001 min
Lab File: BN015263.D
Acq: 01 Jul 2021 16:54




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:152 Resp: 58497
Ion Ratio Lower Upper
152 100
151 20.4 16.7 25.1



#12
2-Methylnaphthalene
Concen: 0.406 ng
RT: 12.158 min Scan# 966
Delta R.T. -0.001 min
Lab File: BN015263.D
Acq: 01 Jul 2021 16:54

Tgt Ion:142 Resp: 67514
Ion Ratio Lower Upper
142 100
141 88.9 72.3 108.5
115 33.6 40.7 61.1#









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.345 min Scan# 1189
Delta R.T. -0.001 min
Lab File: BN015263.D
Acq: 01 Jul 2021 16:54




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:164 Resp: 47501
Ion Ratio Lower Upper
164 100
162 101.2 83.8 125.6
160 45.6 37.4 56.0



#14
2,4,6-Tribromophenol
Concen: 0.496 ng
RT: 15.842 min Scan# 1307
Delta R.T. -0.001 min
Lab File: BN015263.D
Acq: 01 Jul 2021 16:54

Tgt Ion:330 Resp: 5648
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 35.5 22.4 33.6#









#15
2-Fluorobiphenyl
Concen: 0.387 ng
RT: 12.962 min Scan# 1080
Delta R.T. -0.001 min
Lab File: BN015263.D
Acq: 01 Jul 2021 16:54




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:172	Resp:	74710
Ion Ratio	Lower	Upper
172	100	
171	35.1	30.6 45.8
170	23.9	20.2 30.4



#16
Acenaphthylene
Concen: 0.390 ng
RT: 14.066 min Scan# 1167
Delta R.T. -0.001 min
Lab File: BN015263.D
Acq: 01 Jul 2021 16:54

Tgt Ion:152	Resp:	82090
Ion Ratio	Lower	Upper
152	100	
151	19.4	16.7 25.1
153	13.0	10.5 15.7









#17
Acenaphthene
Concen: 0.393 ng
RT: 14.408 min Scan# 1194
Delta R.T. -0.001 min
Lab File: BN015263.D
Acq: 01 Jul 2021 16:54




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:154 Resp: 57994
Ion Ratio Lower Upper
154 100
153 112.7 90.9 136.3
152 55.2 45.8 68.8



#18
Fluorene
Concen: 0.389 ng
RT: 15.398 min Scan# 1272
Delta R.T. -0.001 min
Lab File: BN015263.D
Acq: 01 Jul 2021 16:54

Tgt Ion:166 Resp: 69950
Ion Ratio Lower Upper
166 100
165 97.8 79.2 118.8
167 13.5 10.7 16.1









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.090 min Scan# 1409
Delta R.T. -0.001 min
Lab File: BN015263.D
Acq: 01 Jul 2021 16:54




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:188 Resp: 90306
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 10.2 9.0 13.4



#20
4,6-Dinitro-2-methylphenol
Concen: 0.331 ng
RT: 15.487 min Scan# 1279
Delta R.T. -0.001 min
Lab File: BN015263.D
Acq: 01 Jul 2021 16:54

Tgt Ion:198 Resp: 231
Ion Ratio Lower Upper
198 100
182 41.3 34.5 51.7
77 0.0 0.0 0.0









#21
4-Bromophenyl-phenylether
Concen: 0.381 ng
RT: 16.287 min Scan# 1343
Delta R.T. -0.001 min
Lab File: BN015263.D
Acq: 01 Jul 2021 16:54




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:248 Resp: 21824
Ion Ratio Lower Upper
248 100
250 95.9 86.2 129.4
141 59.2 36.2 54.2#



#22
Hexachlorobenzene
Concen: 0.382 ng
RT: 16.409 min Scan# 1353
Delta R.T. -0.001 min
Lab File: BN015263.D
Acq: 01 Jul 2021 16:54

Tgt Ion:284 Resp: 24857
Ion Ratio Lower Upper
284 100
142 34.7 22.2 33.2#
249 30.7 26.6 39.8









#23
Atrazine
Concen: 0.455 ng
RT: 16.555 min Scan# 1365
Delta R.T. -0.001 min
Lab File: BN015263.D
Acq: 01 Jul 2021 16:54




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:	200	Resp:	13992
Ion Ratio	Lower	Upper	
200	100		
173	27.8	21.7	32.5
215	48.1	43.1	64.7



#24
Pentachlorophenol
Concen: 0.563 ng
RT: 16.750 min Scan# 1381
Delta R.T. -0.001 min
Lab File: BN015263.D
Acq: 01 Jul 2021 16:54

Tgt Ion:	266	Resp:	3588
Ion Ratio	Lower	Upper	
266	100		
264	63.1	51.0	76.4
268	63.2	50.6	76.0









#25
Phenanthrene
Concen: 0.391 ng
RT: 17.127 min Scan# 1412
Delta R.T. -0.001 min
Lab File: BN015263.D
Acq: 01 Jul 2021 16:54




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:178 Resp: 113645
Ion Ratio Lower Upper
178 100
176 18.8 15.6 23.4
179 15.5 12.1 18.1



#26
Anthracene
Concen: 0.383 ng
RT: 17.224 min Scan# 1420
Delta R.T. -0.001 min
Lab File: BN015263.D
Acq: 01 Jul 2021 16:54

Tgt Ion:178 Resp: 95583
Ion Ratio Lower Upper
178 100
176 18.3 15.1 22.7
179 15.3 12.2 18.2









#27
Fluoranthene-d10
Concen: 0.383 ng
RT: 19.132 min Scan# 1710
Delta R.T. -0.001 min
Lab File: BN015263.D
Acq: 01 Jul 2021 16:54




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:212 Resp: 102026
Ion Ratio Lower Upper
212 100
106 14.2 7.0 10.6#
104 7.8 4.1 6.1#



#28
Fluoranthene
Concen: 0.401 ng
RT: 19.162 min Scan# 1716
Delta R.T. -0.001 min
Lab File: BN015263.D
Acq: 01 Jul 2021 16:54

Tgt Ion:202 Resp: 125322
Ion Ratio Lower Upper
202 100
101 11.9 6.6 10.0#
203 17.3 13.9 20.9









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.291 min Scan# 2007
Delta R.T. -0.012 min
Lab File: BN015263.D
Acq: 01 Jul 2021 16:54




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:240 Resp: 76687
Ion Ratio Lower Upper
240 100
120 11.3 4.2 6.4#
236 26.5 21.4 32.2



#30
Pyrene
Concen: 0.424 ng
RT: 19.525 min Scan# 1789
Delta R.T. -0.005 min
Lab File: BN015263.D
Acq: 01 Jul 2021 16:54

Tgt Ion:202 Resp: 123593
Ion Ratio Lower Upper
202 100
200 20.5 16.9 25.3
203 17.3 13.9 20.9









#31
Terphenyl-d14
Concen: 0.417 ng
RT: 19.733 min Scan# 1831
Delta R.T. -0.006 min
Lab File: BN015263.D
Acq: 01 Jul 2021 16:54




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:244 Resp: 74904
Ion Ratio Lower Upper
244 100
212 7.7 6.3 9.5
122 13.5 6.2 9.2#



#32
Benzo(a)anthracene
Concen: 0.387 ng
RT: 21.280 min Scan# 2006
Delta R.T. -0.001 min
Lab File: BN015263.D
Acq: 01 Jul 2021 16:54

Tgt Ion:228 Resp: 101844
Ion Ratio Lower Upper
228 100
226 25.9 23.0 34.6
229 20.0 15.7 23.5









#33
Chrysene
Concen: 0.395 ng
RT: 21.334 min Scan# 2011
Delta R.T. -0.001 min
Lab File: BN015263.D
Acq: 01 Jul 2021 16:54

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion:228 Resp: 113318




Ion	Ratio	Lower	Upper
228	100		
226	28.5	25.0	37.6
229	19.9	15.9	23.9




#34
Bis(2-ethylhexyl)phthalate
Concen: 0.573 ng
RT: 21.217 min Scan# 2000
Delta R.T. -0.011 min
Lab File: BN015263.D
Acq: 01 Jul 2021 16:54

Tgt Ion:149 Resp: 40691


Ion	Ratio	Lower	Upper
149	100		
167	27.9	28.9	43.3#
279	3.5	4.9	7.3#








#35
Perylene-d12
Concen: 0.400 ng
RT: 23.539 min Scan# 2443
Delta R.T. -0.011 min
Lab File: BN015263.D
Acq: 01 Jul 2021 16:54

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC







Tgt Ion:264 Resp: 66071
Ion Ratio Lower Upper
264 100
260 24.8 20.6 31.0
265 22.9 21.4 32.2



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.352 ng
RT: 25.798 min Scan# 3103
Delta R.T. -0.015 min
Lab File: BN015263.D
Acq: 01 Jul 2021 16:54


Tgt Ion:276 Resp: 83304
Ion Ratio Lower Upper
276 100
138 25.7 10.0 15.0#
277 24.9 20.4 30.6







#37
Benzo(b)fluoranthene
Concen: 0.411 ng
RT: 22.865 min Scan# 2246
Delta R.T. -0.011 min
Lab File: BN015263.D
Acq: 01 Jul 2021 16:54


Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC





Tgt Ion:252 Resp: 106346
Ion Ratio Lower Upper
252 100
253 22.1 18.5 27.7
125 14.4 5.8 8.6#




#38
Benzo(k)fluoranthene
Concen: 0.388 ng
RT: 22.909 min Scan# 2259
Delta R.T. -0.011 min
Lab File: BN015263.D
Acq: 01 Jul 2021 16:54



Tgt Ion:252 Resp: 107308
Ion Ratio Lower Upper
252 100
253 22.0 19.1 28.7
125 11.1 5.4 8.0#





#39

Benzo(a)pyrene


Concen: 0.389 ng

RT: 23.443 min Scan# 2415




Delta R.T. -0.011 min

Lab File: BN015263.D

Acq: 01 Jul 2021 16:54



Tgt Ion:252	Resp:	92096
Ion Ratio	Lower	Upper
252	100	
253	24.0	19.4 29.2
125	26.8	7.4 11.2#



#40

Dibenzo(a,h)anthracene


Concen: 0.345 ng

RT: 25.812 min Scan# 3107



Delta R.T. -0.015 min


Lab File: BN015263.D

Acq: 01 Jul 2021 16:54



Tgt Ion:278	Resp:	64042
Ion Ratio	Lower	Upper
278	100	
139	19.1	9.0 13.4#
279	23.9	21.0 31.4





#41
 Benzo(g,h,i)perylene
 Concen: 0.349 ng
 RT: 26.486 min Scan# 3304
 Delta R.T. -0.011 min
 Lab File: BN015263.D
 Acq: 01 Jul 2021 16:54

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC

Tgt Ion:	276	Resp:	71499
Ion	Ratio	Lower	Upper
276	100		
277	23.7	19.7	29.5
138	23.8	9.8	14.6#

