


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN063021\
 Data File : BN015226.D
 Acq On : 30 Jun 2021 17:23
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN063021.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Jul 01 09:06:31 2021






AutoFind: Scans 2458, 2459, 2460; Background Corrected with Scan 2451

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	34.1	55405	PASS
68	69	0.00	2	1.6	927	PASS
69	198	0.00	100	36.1	58685	PASS
70	69	0.00	2	0.2	117	PASS
127	198	10	80	47.0	76488	PASS
197	198	0.00	2	0.3	485	PASS
198	198	100	100	100.0	162635	PASS
199	198	5	9	7.0	11354	PASS
275	198	10	60	25.2	40904	PASS
365	198	1	100	3.0	4863	PASS
441	198	0.01	100	13.6	22061	PASS
442	442	50	100	100.0	138056	PASS
443	442	15	24	20.0	27555	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN063021\
Data File : BN015226.D
Acq On : 30 Jun 2021 17:23
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Jul 01 04:58:31 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 23 10:10:58 2021
Response via : Initial Calibration



TIC: BN015226.D\data.ms

(70) Pentachlorophenol (C)

16.745min (-0.006) 19072.41 ng




response 130517

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	59.06
264.00	61.60	65.11
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN063021\
Data File : BN015226.D
Acq On : 30 Jun 2021 17:23
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Jul 01 04:58:31 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 23 10:10:58 2021
Response via : Initial Calibration



TIC: BN015226.D\data.ms

(77) Benzidine

19.351min (-0.006) 0.00 ng

response 974581

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	14.38
183.00	13.20	11.66
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
6/30/2021	BNA_N	<u>BN015226.D</u>
Compound Name	Response	Retention Time
DDT	379531	20.598
DDD	6812	20.157
DDE	704	19.645
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
7516	387047	1.94