

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN063022\
 Data File : BN020587.D
 Acq On : 30 Jun 2022 11:27
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD0.4386

Quant Time: Jun 30 11:55:55 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN062422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 24 00:17:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.737	152	3705	0.400	ng/ul	-0.02
4) Naphthalene-d8	10.518	136	11294	0.400	ng/ul	-0.02
9) Acenaphthene-d10	14.368	164	6224	0.400	ng/ul	-0.01
13) Phenanthrene-d10	17.112	188	12350	0.400	ng/ul	0.00
17) Chrysene-d12	21.306	240	8807	0.400	ng/ul	0.00
23) Perylene-d12	23.586	264	6800	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.193	96	1891	0.441	ng/ul	-0.01
6) 2-Methylnaphthalene-d10	12.113	152	6718	0.362	ng/ul	-0.02
18) Fluoranthene-d10	19.141	212	12704	0.437	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.227	88	1936	0.442	ng/ul#	91
5) Naphthalene	10.568	128	12374	0.355	ng/ul	99
7) 2-Methylnaphthalene	12.190	142	7563	0.324	ng/ul	99
8) 1-Methylnaphthalene	12.405	142	8179	0.370	ng/ul	100
10) Acenaphthylene	14.086	152	10256	0.348	ng/ul	99
11) Acenaphthene	14.428	153	8406	0.356	ng/ul	99
12) Fluorene	15.418	166	9275	0.331	ng/ul	99
14) Pentachlorophenol	16.782	266	538	0.216	ng/ul	97
15) Phenanthrene	17.150	178	15164	0.353	ng/ul	99
16) Anthracene	17.247	178	12304	0.332	ng/ul	99
19) Fluoranthene	19.174	202	15694	0.390	ng/ul	99
20) Pyrene	19.536	202	15797	0.379	ng/ul	99
21) Benzo(a)anthracene	21.292	228	10229	0.323	ng/ul	100
22) Chrysene	21.341	228	12683	0.360	ng/ul	99
24) Benzo(b)fluoranthene	22.896	252	11032	0.351	ng/ul	99
25) Benzo(k)fluoranthene	22.943	252	11241	0.354	ng/ul	99
26) Benzo(a)pyrene	23.487	252	10118	0.354	ng/ul	99
27) Indeno(1,2,3-cd)pyrene	25.920	276	10771	0.378	ng/ul	97
28) Dibenzo(a,h)anthracene	25.933	278	8546	0.372	ng/ul	100
29) Benzo(g,h,i)perylene	26.631	276	9642	0.395	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN063022\
Data File : BN020587.D
Acq On : 30 Jun 2022 11:27
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTD0.4386

Quant Time: Jun 30 11:55:55 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN062422.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 24 00:17:34 2022
Response via : Initial Calibration

