

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN063022\
 Data File : BN020589.D
 Acq On : 30 Jun 2022 12:40
 Operator : CG/JU
 Sample : PB145665BL
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK665

Quant Time: Jun 30 15:39:22 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN062422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 24 00:17:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.737	152	4030	0.400	ng/ul	-0.02
4) Naphthalene-d8	10.518	136	12600	0.400	ng/ul	-0.01
9) Acenaphthene-d10	14.368	164	7139	0.400	ng/ul	-0.01
13) Phenanthrene-d10	17.107	188	14163	0.400	ng/ul	-0.01
17) Chrysene-d12	21.306	240	9242	0.400	ng/ul	0.00
23) Perylene-d12	23.577	264	9452	0.400	ng/ul	#-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.193	96	29483	6.315	ng/ul	-0.01
6) 2-Methylnaphthalene-d10	12.124	152	5815	0.281	ng/ul	0.00
18) Fluoranthene-d10	19.141	212	12606	0.414	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN063022\
Data File : BN020589.D
Acq On : 30 Jun 2022 12:40
Operator : CG/JU
Sample : PB145665BL
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK665

Quant Time: Jun 30 15:39:22 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN062422.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 24 00:17:34 2022
Response via : Initial Calibration

