

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN070220\
 Data File : BN011095.D
 Acq On : 02 Jul 2020 20:41
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jul 02 21:13:35 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN063020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 30 14:12:13 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	106	0.00
2	1,4-Dioxane	0.564	0.497	11.9	91	0.00
3	n-Nitrosodimethylamine	0.285	0.268	6.0	105	0.00
4 S	2-Fluorophenol	0.995	0.901	9.4	94	0.00
5 S	Phenol-d6	1.144	1.045	8.7	92	0.00
6	bis(2-Chloroethyl)ether	1.039	0.884	14.9	82	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	95	0.00
8 S	Nitrobenzene-d5	0.340	0.344	-1.2	96	0.00
9	Naphthalene	1.194	1.173	1.8	94	0.00
10	Hexachlorobutadiene	0.303	0.345	-13.9	100	0.00
11 SURR	2-Methylnaphthalene-d10	0.692	0.666	3.8	88	0.00
12	2-Methylnaphthalene	0.799	0.771	3.5	91	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	104	0.00
14 S	2,4,6-Tribromophenol	0.168	0.158	6.0	96	-0.01
15 S	2-Fluorobiphenyl	1.884	1.775	5.8	95	0.00
16	Acenaphthylene	2.006	1.917	4.4	99	0.00
17	Acenaphthene	1.359	1.310	3.6	99	0.00
18	Fluorene	1.698	1.592	6.2	96	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	99	-0.01
20	4,6-Dinitro-2-methylphenol	0.043	0.039	9.3	100	0.00
21	4-Bromophenyl-phenylether	0.266	0.256	3.8	94	0.00
22	Hexachlorobenzene	0.313	0.320	-2.2	99	0.00
23	Atrazine	0.188	0.183	2.7	94	0.00
24	Pentachlorophenol	0.084	0.083	1.2	99	0.00
25	Phenanthrene	1.295	1.287	0.6	96	0.00
26	Anthracene	1.072	1.020	4.9	92	-0.01
27 SURR	Fluoranthene-d10	1.210	1.193	1.4	97	0.00
28	Fluoranthene	1.457	1.439	1.2	97	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	88	-0.01
30	Pyrene	1.492	1.557	-4.4	96	0.00
31 S	Terphenyl-d14	1.041	1.225	-17.7	107	0.00
32	Benzo(a)anthracene	1.337	1.277	4.5	83	0.00
33	Chrysene	1.534	1.610	-5.0	93	0.00
34	Bis(2-ethylhexyl)phthalate	0.540	0.546	-1.1	89	0.00
35	Indeno(1,2,3-cd)pyrene	1.604	1.733	-8.0	97	-0.01
36 I	Perylene-d12	1.000	1.000	0.0	111	0.00
37	Benzo(b)fluoranthene	1.603	1.432	10.7	96	0.00
38	Benzo(k)fluoranthene	1.716	1.707	0.5	99	0.00
39 C	Benzo(a)pyrene	1.382	1.349	2.4	107	0.00
40	Dibenzo(a,h)anthracene	1.416	1.189	16.0	98	0.00
41	Benzo(g,h,i)perylene	1.560	1.347	13.7	99	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN070220\
Data File : BN011095.D
Acq On : 02 Jul 2020 20:41
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jul 02 21:13:35 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN063020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 30 14:12:13 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0