

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN070221\
 Data File : BN015292.D
 Acq On : 02 Jul 2021 12:41
 Operator : CG/JU
 Sample : PB137376BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB137376BL

Quant Time: Jul 02 13:39:34 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN063021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 01 09:06:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.725	152	23362	0.40	ng	# 0.00
7) Naphthalene-d8	10.495	136	103761	0.40	ng	# 0.00
13) Acenaphthene-d10	14.345	164	51377	0.40	ng	0.00
19) Phenanthrene-d10	17.091	188	94635	0.40	ng	0.00
29) Chrysene-d12	21.291	240	53014	0.40	ng	#-0.01
35) Perylene-d12	23.546	264	38071	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.365	112	21533	0.36	ng	0.00
5) Phenol-d6	6.905	99	23361	0.31	ng	0.00
8) Nitrobenzene-d5	8.861	82	23579	0.41	ng	0.00
11) 2-Methylnaphthalene-d10	12.078	152	57069	0.33	ng	0.00
14) 2,4,6-Tribromophenol	15.842	330	3129	0.36	ng	0.00
15) 2-Fluorobiphenyl	12.962	172	78549	0.38	ng	0.00
27) Fluoranthene-d10	19.128	212	84556	0.30	ng	0.00
31) Terphenyl-d14	19.733	244	67209	0.54	ng	0.00

Target Compounds	Qvalue
-----	-----

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN070221\
Data File : BN015292.D
Acq On : 02 Jul 2021 12:41
Operator : CG/JU
Sample : PB137376BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB137376BL

Quant Time: Jul 02 13:39:34 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN063021.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 01 09:06:31 2021
Response via : Initial Calibration

