

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN070221\
 Data File : BN015333.D
 Acq On : 03 Jul 2021 14:25
 Operator : CG/JU
 Sample : M2803-08
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DUP04-20210618

Manual Integrations
 APPROVED

mohammad
 7/6/2021 10:04:14 AM

Quant Time: Jul 05 05:49:55 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN063021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 01 09:06:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.725	152	25804	0.40	ng	# 0.00
7) Naphthalene-d8	10.495	136	120638	0.40	ng	# 0.00
13) Acenaphthene-d10	14.345	164	65962	0.40	ng	0.00
19) Phenanthrene-d10	17.090	188	130142	0.40	ng	# 0.00
29) Chrysene-d12	21.291	240	93484	0.40	ng	-0.01
35) Perylene-d12	23.532	264	73518	0.40	ng	-0.02
System Monitoring Compounds						
4) 2-Fluorophenol	5.365	112	11297	0.17	ng	0.00
5) Phenol-d6	6.905	99	7566	0.09	ng	0.00
8) Nitrobenzene-d5	8.860	82	32116	0.48	ng	0.00
11) 2-Methylnaphthalene-d10	12.083	152	68712	0.34	ng	0.00
14) 2,4,6-Tribromophenol	15.829	330	7602	0.49	ng	-0.01
15) 2-Fluorobiphenyl	12.962	172	97460	0.36	ng	0.00
27) Fluoranthene-d10	19.122	212	138276	0.36	ng	-0.01
31) Terphenyl-d14	19.728	244	112012	0.51	ng	-0.01
Target Compounds						Qvalue
2) 1,4-Dioxane	3.355	88	48849m	4.04	ng	
6) bis(2-Chloroethyl)ether	7.163	93	3551	0.04	ng	# 84
24) Pentachlorophenol	16.750	266	50	0.25	ng	# 69
34) Bis(2-ethylhexyl)phtha...	21.206	149	35541	0.46	ng	# 87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN070221\
Data File : BN015333.D
Acq On : 03 Jul 2021 14:25
Operator : CG/JU
Sample : M2803-08
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DUP04-20210618

Manual Integrations
APPROVED

mohammad
7/6/2021 10:04:14 AM

Quant Time: Jul 05 05:49:55 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN063021.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 01 09:06:31 2021
Response via : Initial Calibration

