

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN070522\
 Data File : BN020663.D
 Acq On : 05 Jul 2022 16:21
 Operator : CG/JU
 Sample : N3457-11DL 2X
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RE103D1-20220620DL

Quant Time: Jul 06 01:03:14 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN062622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jul 02 03:10:12 2022
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Christian Giraldo 07/06/2022
 Supervised By :Jagrut Upadhyay 07/06/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.833	152	4750	0.400	ng	0.00
7) Naphthalene-d8	10.626	136	13550	0.400	ng	# 0.00
13) Acenaphthene-d10	14.474	164	7770	0.400	ng	0.00
19) Phenanthrene-d10	17.226	188	15005	0.400	ng	# 0.00
29) Chrysene-d12	21.440	240	9206	0.400	ng	# 0.00
35) Perylene-d12	23.790	264	6956	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.428	112	597	0.045	ng	0.01
5) Phenol-d6	7.132	99	262m	0.017	ng	0.10
8) Nitrobenzene-d5	9.035	82	833	0.076	ng	0.04
11) 2-Methylnaphthalene-d10	12.242	152	2755	0.119	ng	0.02
14) 2,4,6-Tribromophenol	15.995	330	169	0.058	ng	0.02
15) 2-Fluorobiphenyl	13.116	172	3622	0.113	ng	0.01
27) Fluoranthene-d10	19.274	212	5926	0.133	ng	0.00
31) Terphenyl-d14	19.877	244	3923	0.178	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.254	88	14386	2.311	ng	95
34) Bis(2-ethylhexyl)phtha...	21.351	149	481	0.026	ng	# 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN070522\
Data File : BN020663.D
Acq On : 05 Jul 2022 16:21
Operator : CG/JU
Sample : N3457-11DL 2X
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE103D1-20220620DL

Quant Time: Jul 06 01:03:14 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN062622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Jul 02 03:10:12 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 07/06/2022
Supervised By :Jagrut Upadhyay 07/06/2022

