

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN070524\
 Data File : BN032530.D
 Acq On : 06 Jul 2024 05:10
 Operator : MA/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD0.4252

Quant Time: Jul 06 05:36:08 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN062724.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 03 03:26:42 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.599	152	4293	0.400	ng/ul	0.00
4) Naphthalene-d8	10.375	136	13225	0.400	ng/ul #	0.00
9) Acenaphthene-d10	14.243	164	7608	0.400	ng/ul	-0.01
13) Phenanthrene-d10	17.010	188	14955	0.400	ng/ul	0.00
17) Chrysene-d12	21.210	240	14013	0.400	ng/ul	0.00
23) Perylene-d12	23.426	264	16843	0.400	ng/ul #	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.139	96	2308	0.484	ng/ul	0.00
6) 2-Methylnaphthalene-d10	11.981	152	7193	0.363	ng/ul	0.00
18) Fluoranthene-d10	19.044	212	14804	0.351	ng/ul	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.173	88	2441	0.449	ng/ul#	88
5) Naphthalene	10.425	128	13966	0.399	ng/ul	98
7) 2-Methylnaphthalene	12.058	142	8328	0.359	ng/ul	99
8) 1-Methylnaphthalene	12.272	142	9163	0.381	ng/ul	100
10) Acenaphthylene	13.970	152	13263	0.419	ng/ul	98
11) Acenaphthene	14.307	153	9749	0.397	ng/ul	98
12) Fluorene	15.311	166	10274	0.351	ng/ul	100
14) Pentachlorophenol	16.690	266	1332	0.394	ng/ul	99
15) Phenanthrene	17.053	178	15427	0.376	ng/ul	99
16) Anthracene	17.150	178	13913	0.380	ng/ul	98
19) Fluoranthene	19.077	202	18740	0.351	ng/ul	98
20) Pyrene	19.439	202	19901	0.357	ng/ul	98
21) Benzo(a)anthracene	21.196	228	19045	0.377	ng/ul	98
22) Chrysene	21.248	228	20374	0.367	ng/ul	98
24) Benzo(b)fluoranthene	22.759	252	21447	0.293	ng/ul	93
25) Benzo(k)fluoranthene	22.803	252	22114	0.311	ng/ul#	93
26) Benzo(a)pyrene	23.332	252	20913	0.411	ng/ul	94
27) Indeno(1,2,3-cd)pyrene	25.665	276	26363	0.331	ng/ul	99
28) Dibenzo(a,h)anthracene	25.668	278	21000	0.355	ng/ul	94
29) Benzo(g,h,i)perylene	26.353	276	21743	0.342	ng/ul	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN070524\
Data File : BN032530.D
Acq On : 06 Jul 2024 05:10
Operator : MA/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTD0.4252

Quant Time: Jul 06 05:36:08 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN062724.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jul 03 03:26:42 2024
Response via : Initial Calibration

