

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN070620\
 Data File : BN011097.D
 Acq On : 06 Jul 2020 08:43
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jul 06 10:53:56 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN063020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 06 10:52:12 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	116	0.00
2	1,4-Dioxane	0.564	0.558	1.1	111	0.00
3	n-Nitrosodimethylamine	0.285	0.297	-4.2	126	0.00
4 S	2-Fluorophenol	0.995	1.024	-2.9	116	0.00
5 S	Phenol-d6	1.144	1.244	-8.7	120	0.00
6	bis(2-Chloroethyl)ether	1.039	1.038	0.1	104	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	114	0.00
8 S	Nitrobenzene-d5	0.340	0.377	-10.9	126	0.00
9	Naphthalene	1.194	1.200	-0.5	115	0.00
10	Hexachlorobutadiene	0.303	0.309	-2.0	107	0.00
11 SURR	2-Methylnaphthalene-d10	0.692	0.699	-1.0	110	0.00
12	2-Methylnaphthalene	0.799	0.829	-3.8	117	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	115	0.00
14 S	2,4,6-Tribromophenol	0.168	0.180	-7.1	121	0.00
15 S	2-Fluorobiphenyl	1.884	1.950	-3.5	116	0.00
16	Acenaphthylene	2.006	2.000	0.3	115	0.00
17	Acenaphthene	1.359	1.378	-1.4	116	0.00
18	Fluorene	1.698	1.712	-0.8	114	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	115	0.00
20	4,6-Dinitro-2-methylphenol	0.043	0.043	0.0	125	0.00
21	4-Bromophenyl-phenylether	0.266	0.283	-6.4	120	0.00
22	Hexachlorobenzene	0.313	0.319	-1.9	115	0.00
23	Atrazine	0.188	0.184	2.1	109	0.00
24	Pentachlorophenol	0.084	0.102	-21.4	140	0.00
25	Phenanthrene	1.295	1.333	-2.9	115	0.00
26	Anthracene	1.072	1.077	-0.5	113	0.00
27 SURR	Fluoranthene-d10	1.210	1.227	-1.4	115	0.00
28	Fluoranthene	1.457	1.477	-1.4	115	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	103	0.00
30	Pyrene	1.492	1.725	-15.6	124	0.00
31 S	Terphenyl-d14	1.041	1.175	-12.9	119	0.00
32	Benzo(a)anthracene	1.337	1.392	-4.1	105	0.00
33	Chrysene	1.534	1.533	0.1	103	0.00
34	Bis(2-ethylhexyl)phthalate	0.540	0.563	-4.3	107	0.00
35	Indeno(1,2,3-cd)pyrene	1.604	1.296	19.2	85	0.00
36 I	Perylene-d12	1.000	1.000	0.0	104	0.00
37	Benzo(b)fluoranthene	1.603	1.589	0.9	99	0.00
38	Benzo(k)fluoranthene	1.716	1.810	-5.5	98	0.00
39 C	Benzo(a)pyrene	1.382	1.386	-0.3	103	0.00
40	Dibenzo(a,h)anthracene	1.416	0.976	31.1#	75	0.00
41	Benzo(g,h,i)perylene	1.560	1.411	9.6	97	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN070620\
Data File : BN011097.D
Acq On : 06 Jul 2020 08:43
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jul 06 10:53:56 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN063020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jul 06 10:52:12 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0