

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN070720\
 Data File : BN011108.D
 Acq On : 07 Jul 2020 08:49
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jul 07 10:48:22 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN063020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 07 10:47:09 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	136	0.00
2	1,4-Dioxane	0.400	0.325	18.8	107	0.00
3	n-Nitrosodimethylamine	0.400	0.356	11.0	127	0.00
4 S	2-Fluorophenol	0.400	0.360	10.0	119	0.00
5 S	Phenol-d6	0.400	0.375	6.3	121	0.00
6	bis(2-Chloroethyl)ether	0.400	0.402	-0.5	124	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	122	0.00
8 S	Nitrobenzene-d5	0.400	0.423	-5.7	129	0.00
9	Naphthalene	0.400	0.392	2.0	120	0.00
10	Hexachlorobutadiene	0.400	0.405	-1.3	113	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.409	-2.2	119	0.00
12	2-Methylnaphthalene	0.400	0.395	1.3	119	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	120	0.00
14 S	2,4,6-Tribromophenol	0.400	0.421	-5.2	125	0.00
15 S	2-Fluorobiphenyl	0.400	0.420	-5.0	123	0.00
16	Acenaphthylene	0.400	0.401	-0.3	120	0.00
17	Acenaphthene	0.400	0.404	-1.0	120	0.00
18	Fluorene	0.400	0.410	-2.5	121	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	132	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.387	3.3	126	0.00
21	4-Bromophenyl-phenylether	0.400	0.384	4.0	124	0.00
22	Hexachlorobenzene	0.400	0.380	5.0	123	0.00
23	Atrazine	0.400	0.361	9.8	115	0.00
24	Pentachlorophenol	0.400	0.473	-18.2	158	0.00
25	Phenanthrene	0.400	0.393	1.8	127	0.00
26	Anthracene	0.400	0.368	8.0	118	0.00
27 SURR	Fluoranthene-d10	0.400	0.377	5.8	122	0.00
28	Fluoranthene	0.400	0.375	6.3	122	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	111	0.00
30	Pyrene	0.400	0.423	-5.7	122	0.00
31 S	Terphenyl-d14	0.400	0.427	-6.7	122	0.00
32	Benzo(a)anthracene	0.400	0.384	4.0	104	0.00
33	Chrysene	0.400	0.426	-6.5	118	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.415	-3.7	115	0.00
35	Indeno(1,2,3-cd)pyrene	0.400	0.366	8.5	104	0.00
36 I	Perylene-d12	0.400	0.400	0.0	114	0.00
37	Benzo(b)fluoranthene	0.400	0.399	0.3	110	0.00
38	Benzo(k)fluoranthene	0.400	0.419	-4.7	107	0.00
39 C	Benzo(a)pyrene	0.400	0.406	-1.5	115	0.00
40	Dibenzo(a,h)anthracene	0.400	0.317	20.8	95	0.00
41	Benzo(g,h,i)perylene	0.400	0.431	-7.7	127	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN070720\
Data File : BN011108.D
Acq On : 07 Jul 2020 08:49
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jul 07 10:48:22 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN063020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 07 10:47:09 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0