

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN070720\
 Data File : BN011117.D
 Acq On : 07 Jul 2020 16:57
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jul 07 17:27:42 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN063020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 07 14:39:07 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	83	0.00
2	1,4-Dioxane	0.564	0.518	8.2	74	-0.02
3	n-Nitrosodimethylamine	0.285	0.300	-5.3	92	-0.02
4 S	2-Fluorophenol	0.995	0.940	5.5	76	0.00
5 S	Phenol-d6	1.144	1.171	-2.4	81	0.00
6	bis(2-Chloroethyl)ether	1.039	1.035	0.4	75	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	76	0.00
8 S	Nitrobenzene-d5	0.340	0.365	-7.4	82	0.00
9	Naphthalene	1.194	1.188	0.5	77	0.00
10	Hexachlorobutadiene	0.303	0.327	-7.9	76	0.00
11 SURR	2-Methylnaphthalene-d10	0.692	0.679	1.9	72	0.00
12	2-Methylnaphthalene	0.799	0.783	2.0	74	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	78	0.00
14 S	2,4,6-Tribromophenol	0.168	0.186	-10.7	85	0.00
15 S	2-Fluorobiphenyl	1.884	1.925	-2.2	78	0.00
16	Acenaphthylene	2.006	2.020	-0.7	78	0.00
17	Acenaphthene	1.359	1.370	-0.8	78	0.00
18	Fluorene	1.698	1.736	-2.2	78	0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	80	0.00
20	4,6-Dinitro-2-methylphenol	0.043	0.048	-11.6	99	0.00
21	4-Bromophenyl-phenylether	0.266	0.267	-0.4	79	0.00
22	Hexachlorobenzene	0.313	0.333	-6.4	83	0.00
23	Atrazine	0.188	0.192	-2.1	79	0.01
24	Pentachlorophenol	0.084	0.089	-6.0	85	0.00
25	Phenanthrene	1.295	1.299	-0.3	78	0.00
26	Anthracene	1.072	1.058	1.3	77	0.00
27 SURR	Fluoranthene-d10	1.210	1.248	-3.1	81	0.00
28	Fluoranthene	1.457	1.473	-1.1	80	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	73	0.00
30	Pyrene	1.492	1.564	-4.8	80	0.00
31 S	Terphenyl-d14	1.041	1.097	-5.4	79	0.00
32	Benzo(a)anthracene	1.337	1.276	4.6	68	0.00
33	Chrysene	1.534	1.611	-5.0	77	0.00
34	Bis(2-ethylhexyl)phthalate	0.540	0.581	-7.6	79	-0.01
35	Indeno(1,2,3-cd)pyrene	1.604	1.665	-3.8	77	0.00
36 I	Perylene-d12	1.000	1.000	0.0	78	0.00
37	Benzo(b)fluoranthene	1.603	1.526	4.8	72	0.00
38	Benzo(k)fluoranthene	1.716	1.803	-5.1	74	0.00
39 C	Benzo(a)pyrene	1.382	1.391	-0.7	78	0.00
40	Dibenzo(a,h)anthracene	1.416	1.280	9.6	74	0.00
41	Benzo(g,h,i)perylene	1.560	1.575	-1.0	82	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN070720\
Data File : BN011117.D
Acq On : 07 Jul 2020 16:57
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jul 07 17:27:42 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN063020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 07 14:39:07 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0