

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN070822\
 Data File : BN020695.D
 Acq On : 08 Jul 2022 17:06
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jul 09 00:09:54 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN062622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jul 02 03:10:12 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	148	0.00
2	1,4-Dioxane	0.400	0.349	12.8	135	0.00
3	n-Nitrosodimethylamine	0.400	0.281	29.8#	111	0.00
4 S	2-Fluorophenol	0.400	0.268	33.0#	102	0.00
5 S	Phenol-d6	0.400	0.292	27.0#	113	0.00
6	bis(2-Chloroethyl)ether	0.400	0.315	21.3	123	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	150	0.00
8 S	Nitrobenzene-d5	0.400	0.320	20.0	131	-0.01
9	Naphthalene	0.400	0.366	8.5	143	-0.01
10	Hexachlorobutadiene	0.400	0.361	9.8	139	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.344	14.0	136	0.00
12	2-Methylnaphthalene	0.400	0.348	13.0	136	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	145	0.00
14 S	2,4,6-Tribromophenol	0.400	0.242	39.5#	102	0.00
15 S	2-Fluorobiphenyl	0.400	0.371	7.3	137	-0.01
16	Acenaphthylene	0.400	0.330	17.5	131	0.00
17	Acenaphthene	0.400	0.353	11.8	133	0.00
18	Fluorene	0.400	0.338	15.5	129	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	138	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.350	12.5	89	0.00
21	4-Bromophenyl-phenylether	0.400	0.370	7.5	137	-0.01
22	Hexachlorobenzene	0.400	0.383	4.3	142	0.00
23	Atrazine	0.400	0.281	29.8#	107	0.00
24	Pentachlorophenol	0.400	0.367	8.3	116	0.00
25	Phenanthrene	0.400	0.348	13.0	128	0.00
26	Anthracene	0.400	0.306	23.5	116	0.00
27 SURR	Fluoranthene-d10	0.400	0.312	22.0	116	0.00
28	Fluoranthene	0.400	0.323	19.3	120	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	102	0.00
30	Pyrene	0.400	0.440	-10.0	115	0.00
31 S	Terphenyl-d14	0.400	0.418	-4.5	109	0.00
32	Benzo(a)anthracene	0.400	0.294	26.5#	80	0.00
33	Chrysene	0.400	0.373	6.8	98	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.371	7.3	107	0.00
35 I	Perylene-d12	0.400	0.400	0.0	81	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.327	18.3	69	0.00
37	Benzo(b)fluoranthene	0.400	0.388	3.0	82	0.00
38	Benzo(k)fluoranthene	0.400	0.360	10.0	78	0.00
39 C	Benzo(a)pyrene	0.400	0.353	11.8	75	0.00
40	Dibenzo(a,h)anthracene	0.400	0.324	19.0	69	0.00
41	Benzo(g,h,i)perylene	0.400	0.332	17.0	69	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4