

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN070824\
 Data File : BN032551.D
 Acq On : 08 Jul 2024 14:55
 Operator : MA/JU
 Sample : PB161896BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB161896BL

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 07/09/2024
 Supervised By :mohammad ahmed 07/09/2024

Quant Time: Jul 08 15:20:09 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN061024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 08 15:11:21 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.603	152	3901m	0.400	ng	0.02
7) Naphthalene-d8	10.410	136	11738m	0.400	ng	0.05
13) Acenaphthene-d10	14.253	164	6698	0.400	ng	0.03
19) Phenanthrene-d10	17.016	188	15792m	0.400	ng	0.02
29) Chrysene-d12	21.229	240	11180	0.400	ng	# 0.03
35) Perylene-d12	23.440	264	10718	0.400	ng	# 0.02
System Monitoring Compounds						
4) 2-Fluorophenol	5.219	112	3262	0.294	ng	0.02
5) Phenol-d6	6.873	99	4691m	0.323	ng	0.09
8) Nitrobenzene-d5	8.873	82	3696m	0.379	ng	0.13
11) 2-Methylnaphthalene-d10	12.017	152	7423m	0.397	ng	0.06
14) 2,4,6-Tribromophenol	15.800	330	649	0.202	ng	0.05
15) 2-Fluorobiphenyl	12.895	172	10132m	0.399	ng	0.05
27) Fluoranthene-d10	19.049	212	15644	0.364	ng	0.02
31) Terphenyl-d14	19.648	244	11386	0.417	ng	0.01

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN070824\
Data File : BN032551.D
Acq On : 08 Jul 2024 14:55
Operator : MA/JU
Sample : PB161896BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB161896BL

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/09/2024
Supervised By :mohammad ahmed 07/09/2024

Quant Time: Jul 08 15:20:09 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN061024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jul 08 15:11:21 2024
Response via : Initial Calibration

