

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN070919\
 Data File : BN006771.D
 Acq On : 09 Jul 2019 17:41
 Operator : HP/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jul 10 04:19:02 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN070919.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 09 13:11:11 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	88	0.00
2	1,4-Dioxane	0.657	0.637	3.0	86	0.00
3	n-Nitrosodimethylamine	0.510	0.295	42.2#	58	0.00
4 S	2-Fluorophenol	0.889	0.649	27.0#	65	0.02
5 S	Phenol-d6	1.223	0.912	25.4#	69	0.02
6	bis(2-Chloroethyl)ether	1.361	0.900	33.9#	62	0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	87	0.00
8 S	Nitrobenzene-d5	0.273	0.207	24.2	68	0.03
9	Naphthalene	1.071	1.016	5.1	77	0.00
10	Hexachlorobutadiene	0.232	0.229	1.3	78	0.00
11 SURR	2-Methylnaphthalene-d10	0.598	0.541	9.5	74	0.00
12	2-Methylnaphthalene	0.685	0.629	8.2	76	0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	87	0.00
14 S	2,4,6-Tribromophenol	0.187	0.143	23.5	68	0.01
15 S	2-Fluorobiphenyl	1.490	1.373	7.9	76	0.01
16	Acenaphthylene	1.889	1.695	10.3	76	0.00
17	Acenaphthene	1.268	1.220	3.8	77	0.00
18	Fluorene	1.571	1.434	8.7	74	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	85	0.01
20	4-Bromophenyl-phenylether	0.226	0.200	11.5	71	0.01
21	Hexachlorobenzene	0.295	0.290	1.7	77	0.00
22	Pentachlorophenol	0.039	0.022	43.6#	51	0.01
23	Phenanthrene	1.109	1.028	7.3	74	0.01
24	Anthracene	0.961	0.818	14.9	71	0.00
25 SURR	Fluoranthene-d10	1.208	1.096	9.3	72	0.00
26	Fluoranthene	1.392	1.297	6.8	73	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	82	0.00
28	Pvrene	1.390	1.351	2.8	73	0.00
29 S	Terphenyl-d14	0.881	0.830	5.8	72	0.00
30	Benzo(a)anthracene	1.233	1.080	12.4	69	-0.01
31	Chrysene	1.382	1.362	1.4	73	-0.01
32	Bis(2-ethylhexyl)phthalate	0.574	0.420	26.8#	63	-0.01
33	Indeno(1,2,3-cd)pyrene	1.691	1.726	-2.1	73	0.00
34 I	Perylene-d12	1.000	1.000	0.0	83	0.00
35	Benzo(b)fluoranthene	1.257	1.155	8.1	74	0.00
36	Benzo(k)fluoranthene	1.411	1.386	1.8	75	0.00
37 C	Benzo(a)pyrene	1.219	1.122	8.0	71	0.00
38	Dibenzo(a,h)anthracene	1.229	1.172	4.6	73	0.00
39	Benzo(a,h,i)perylene	1.288	1.254	2.6	73	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN070919\
Data File : BN006771.D
Acq On : 09 Jul 2019 17:41
Operator : HP/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jul 10 04:19:02 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN070919.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 09 13:11:11 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			