

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN070919\
 Data File : BN006771.D
 Acq On : 09 Jul 2019 17:41
 Operator : HP/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jul 10 04:19:02 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN070919.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 09 13:11:11 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	88	0.00
2	1,4-Dioxane	0.400	0.388	3.0	86	0.00
3	n-Nitrosodimethylamine	0.400	0.382	4.5	58	0.00
4 S	2-Fluorophenol	0.400	0.292	27.0#	65	0.02
5 S	Phenol-d6	0.400	0.298	25.5#	69	0.02
6	bis(2-Chloroethyl)ether	0.400	0.264	34.0#	62	0.02
7 I	Naphthalene-d8	0.400	0.400	0.0	87	0.00
8 S	Nitrobenzene-d5	0.400	0.303	24.3	68	0.03
9	Naphthalene	0.400	0.379	5.3	77	0.00
10	Hexachlorobutadiene	0.400	0.395	1.3	78	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.362	9.5	74	0.00
12	2-Methylnaphthalene	0.400	0.367	8.3	76	0.01
13 I	Acenaphthene-d10	0.400	0.400	0.0	87	0.00
14 S	2,4,6-Tribromophenol	0.400	0.306	23.5	68	0.01
15 S	2-Fluorobiphenyl	0.400	0.369	7.8	76	0.01
16	Acenaphthylene	0.400	0.359	10.3	76	0.00
17	Acenaphthene	0.400	0.385	3.8	77	0.00
18	Fluorene	0.400	0.365	8.8	74	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	85	0.01
20	4-Bromophenyl-phenylether	0.400	0.353	11.8	71	0.01
21	Hexachlorobenzene	0.400	0.393	1.8	77	0.00
22	Pentachlorophenol	0.400	0.364	9.0	51	0.01
23	Phenanthrene	0.400	0.371	7.3	74	0.01
24	Anthracene	0.400	0.341	14.8	71	0.00
25 SURR	Fluoranthene-d10	0.400	0.363	9.3	72	0.00
26	Fluoranthene	0.400	0.373	6.8	73	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	82	0.00
28	Pvrene	0.400	0.389	2.8	73	0.00
29 S	Terphenyl-d14	0.400	0.377	5.8	72	0.00
30	Benzo(a)anthracene	0.400	0.350	12.5	69	-0.01
31	Chrysene	0.400	0.394	1.5	73	-0.01
32	Bis(2-ethylhexyl)phthalate	0.400	0.293	26.8#	63	-0.01
33	Indeno(1,2,3-cd)pyrene	0.400	0.408	-2.0	73	0.00
34 I	Perylene-d12	0.400	0.400	0.0	83	0.00
35	Benzo(b)fluoranthene	0.400	0.367	8.3	74	0.00
36	Benzo(k)fluoranthene	0.400	0.393	1.8	75	0.00
37 C	Benzo(a)pyrene	0.400	0.368	8.0	71	0.00
38	Dibenzo(a,h)anthracene	0.400	0.382	4.5	73	0.00
39	Benzo(a,h,i)perylene	0.400	0.389	2.8	73	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN070919\
Data File : BN006771.D
Acq On : 09 Jul 2019 17:41
Operator : HP/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jul 10 04:19:02 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN070919.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 09 13:11:11 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				