

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN070924\
 Data File : BN032564.D
 Acq On : 09 Jul 2024 00:03
 Operator : MA/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN070924

Quant Time: Jul 09 02:40:47 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN070924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 09 02:39:48 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.588	152	4389	0.400	ng	0.00
7) Naphthalene-d8	10.378	136	13799	0.400	ng	0.00
13) Acenaphthene-d10	14.231	164	9647	0.400	ng	0.00
19) Phenanthrene-d10	17.004	188	21027	0.400	ng	0.00
29) Chrysene-d12	21.211	240	16744	0.400	ng	0.00
35) Perylene-d12	23.426	264	16900	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.212	112	4048	0.365	ng	0.00
5) Phenol-d6	6.815	99	5256	0.371	ng	0.00
8) Nitrobenzene-d5	8.798	82	3442	0.334	ng	0.00
11) 2-Methylnaphthalene-d10	11.983	152	8497	0.390	ng	0.00
14) 2,4,6-Tribromophenol	15.763	330	1443	0.330	ng	0.00
15) 2-Fluorobiphenyl	12.863	172	13143	0.370	ng	0.00
27) Fluoranthene-d10	19.045	212	20256	0.370	ng	0.00
31) Terphenyl-d14	19.653	244	15216	0.372	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.161	88	2288	0.396	ng	99
3) n-Nitrosodimethylamine	3.486	42	1536	0.366	ng	97
6) bis(2-Chloroethyl)ether	7.047	93	3521	0.370	ng	99
9) Naphthalene	10.421	128	14605	0.387	ng	99
10) Hexachlorobutadiene	10.667	225	2925	0.390	ng	# 99
12) 2-Methylnaphthalene	12.062	142	9747	0.386	ng	100
16) Acenaphthylene	13.953	152	14563	0.351	ng	99
17) Acenaphthene	14.296	154	10843	0.376	ng	100
18) Fluorene	15.300	166	14200	0.375	ng	100
20) 4,6-Dinitro-2-methylph...	15.493	198	498	0.350	ng	98
21) 4-Bromophenyl-phenylether	16.197	248	4798	0.384	ng	97
22) Hexachlorobenzene	16.296	284	5332	0.388	ng	99
23) Atrazine	16.482	200	3331	0.342	ng	98
24) Pentachlorophenol	16.693	266	1368	0.386	ng	98
25) Phenanthrene	17.041	178	21278	0.373	ng	100
26) Anthracene	17.140	178	16219	0.353	ng	99
28) Fluoranthene	19.072	202	25141	0.364	ng	100
30) Pyrene	19.439	202	26362	0.387	ng	100
32) Benzo(a)anthracene	21.202	228	18368	0.332	ng	100
33) Chrysene	21.247	228	26567	0.404	ng	100
34) Bis(2-ethylhexyl)phtha...	21.095	149	11875	0.316	ng	100
36) Indeno(1,2,3-cd)pyrene	25.665	276	23746	0.368	ng	100
37) Benzo(b)fluoranthene	22.759	252	19810	0.352	ng	99
38) Benzo(k)fluoranthene	22.803	252	25257	0.389	ng	99
39) Benzo(a)pyrene	23.335	252	19522	0.366	ng	99
40) Dibenzo(a,h)anthracene	25.680	278	18807	0.367	ng	99
41) Benzo(g,h,i)perylene	26.349	276	20928	0.375	ng	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN070924\
Data File : BN032564.D
Acq On : 09 Jul 2024 00:03
Operator : MA/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
ICVBN070924

Quant Time: Jul 09 02:40:47 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN070924.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 09 02:39:48 2024
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.588 min Scan# 61
 Delta R.T. 0.000 min
 Lab File: BN032564.D
 Acq: 09 Jul 2024 00:03




Instrument :
 BNA_N
 ClientSampleId :
 ICVBN070924


Tgt Ion:152 Resp: 4389
 Ion Ratio Lower Upper
 152 100
 150 152.4 121.6 182.4
 115 66.4 53.1 79.7



#2
 1,4-Dioxane
 Concen: 0.396 ng
 RT: 3.161 min Scan# 25
 Delta R.T. 0.000 min
 Lab File: BN032564.D
 Acq: 09 Jul 2024 00:03

Tgt Ion: 88 Resp: 2288
 Ion Ratio Lower Upper
 88 100
 43 25.1 20.4 30.6
 58 63.4 51.0 76.6









#3
n-Nitrosodimethylamine
Concen: 0.366 ng
RT: 3.486 min Scan# 70
Delta R.T. 0.000 min
Lab File: BN032564.D
Acq: 09 Jul 2024 00:03




Instrument :
BNA_N
ClientSampleId :
ICVBN070924


Tgt Ion: 42 Resp: 1536
Ion Ratio Lower Upper
42 100
74 161.8 126.8 190.2
44 9.0 8.7 13.1



#4
2-Fluorophenol
Concen: 0.365 ng
RT: 5.212 min Scan# 309
Delta R.T. 0.000 min
Lab File: BN032564.D
Acq: 09 Jul 2024 00:03

Tgt Ion: 112 Resp: 4048
Ion Ratio Lower Upper
112 100
64 50.2 41.2 61.8
63 27.7 21.6 32.4









#5
Phenol-d6
Concen: 0.371 ng
RT: 6.815 min Scan# 51
Delta R.T. -0.007 min
Lab File: BN032564.D
Acq: 09 Jul 2024 00:03




Instrument :
BNA_N
ClientSampleId :
ICVBN070924


Tgt Ion: 99 Resp: 5256
Ion Ratio Lower Upper
99 100
42 12.9 10.6 16.0
71 30.6 25.2 37.8



#6
bis(2-Chloroethyl)ether
Concen: 0.370 ng
RT: 7.047 min Scan# 563
Delta R.T. 0.008 min
Lab File: BN032564.D
Acq: 09 Jul 2024 00:03

Tgt Ion: 93 Resp: 3521
Ion Ratio Lower Upper
93 100
63 67.7 54.1 81.1
95 39.2 30.5 45.7









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.378 min Scan# 91
Delta R.T. 0.000 min
Lab File: BN032564.D
Acq: 09 Jul 2024 00:03




Instrument :
BNA_N
ClientSampleId :
ICVBN070924


Tgt Ion:	136	Resp:	13799
Ion	Ratio	Lower	Upper
136	100		
137	12.4	10.0	15.0
54	7.9	6.6	9.8
68	7.3	5.8	8.8



#8
Nitrobenzene-d5
Concen: 0.334 ng
RT: 8.798 min Scan# 784
Delta R.T. 0.000 min
Lab File: BN032564.D
Acq: 09 Jul 2024 00:03

Tgt Ion:	82	Resp:	3442
Ion	Ratio	Lower	Upper
82	100		
128	56.0	43.4	65.2
54	50.0	40.2	60.2





#9

Naphthalene

Concen: 0.387 ng

RT: 10.421 min Scan# 91

Delta R.T. 0.000 min

Lab File: BN032564.D

Acq: 09 Jul 2024 00:03





Instrument :

BNA_N

ClientSampleId :

ICVBN070924

Tgt Ion:128	Resp:	14605
Ion Ratio	Lower	Upper
128	100	
129	12.4	10.2 15.4
127	14.7	11.4 17.0



#10

Hexachlorobutadiene

Concen: 0.390 ng




RT: 10.667 min Scan# 959


Delta R.T. 0.000 min

Lab File: BN032564.D

Acq: 09 Jul 2024 00:03

Tgt Ion:225	Resp:	2925
Ion Ratio	Lower	Upper
225	100	
223	0.0	0.0 0.0
227	63.9	50.6 76.0









#11
2-Methylnaphthalene-d10
Concen: 0.390 ng
RT: 11.983 min Scan# 1163
Delta R.T. 0.000 min
Lab File: BN032564.D
Acq: 09 Jul 2024 00:03




Instrument :
BNA_N
ClientSampleId :
ICVBN070924


Tgt Ion:152 Resp: 8497
Ion Ratio Lower Upper
152 100
151 20.3 16.4 24.6



#12
2-Methylnaphthalene
Concen: 0.386 ng
RT: 12.062 min Scan# 1184
Delta R.T. 0.004 min
Lab File: BN032564.D
Acq: 09 Jul 2024 00:03

Tgt Ion:142 Resp: 9747
Ion Ratio Lower Upper
142 100
141 88.6 71.0 106.6
115 37.6 30.5 45.7









#13
 Acenaphthene-d10
 Concen: 0.400 ng
 RT: 14.231 min Scan# 1480
 Delta R.T. 0.000 min
 Lab File: BN032564.D
 Acq: 09 Jul 2024 00:03




Instrument :
 BNA_N
 ClientSampleId :
 ICVBN070924


Tgt Ion:164 Resp: 9647
 Ion Ratio Lower Upper
 164 100
 162 105.1 83.2 124.8
 160 49.9 38.6 57.8



#14
 2,4,6-Tribromophenol
 Concen: 0.330 ng
 RT: 15.763 min Scan# 1620
 Delta R.T. 0.000 min
 Lab File: BN032564.D
 Acq: 09 Jul 2024 00:03

Tgt Ion:330 Resp: 1443
 Ion Ratio Lower Upper
 330 100
 332 95.8 77.7 116.5
 141 37.1 31.8 47.8





#15

2-Fluorobiphenyl

Concen: 0.370 ng

RT: 12.863 min Scan# 1352

Delta R.T. 0.000 min

Lab File: BN032564.D

Acq: 09 Jul 2024 00:03

Instrument :

BNA_N

ClientSampleId :

ICVBN070924





Tgt Ion:172 Resp: 13143

Ion Ratio Lower Upper

172 100

171 35.9 28.5 42.7

170 24.4 19.2 28.8



#16

Acenaphthylene

Concen: 0.351 ng

RT: 13.953 min Scan# 1454

Delta R.T. 0.000 min

Lab File: BN032564.D

Acq: 09 Jul 2024 00:03




Tgt Ion:152 Resp: 14563


Ion Ratio Lower Upper

152 100

151 19.8 15.6 23.4

153 12.9 10.5 15.7









#17
Acenaphthene
Concen: 0.376 ng
RT: 14.296 min Scan# 1486
Delta R.T. 0.000 min
Lab File: BN032564.D
Acq: 09 Jul 2024 00:03




Instrument :
BNA_N
ClientSampleId :
ICVBN070924


Tgt Ion:154 Resp: 10843
Ion Ratio Lower Upper
154 100
153 113.9 90.9 136.3
152 55.5 44.4 66.6



#18
Fluorene
Concen: 0.375 ng
RT: 15.300 min Scan# 1580
Delta R.T. 0.000 min
Lab File: BN032564.D
Acq: 09 Jul 2024 00:03

Tgt Ion:166 Resp: 14200
Ion Ratio Lower Upper
166 100
165 98.2 78.7 118.1
167 13.5 10.7 16.1









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.004 min Scan# 1720
Delta R.T. 0.000 min
Lab File: BN032564.D
Acq: 09 Jul 2024 00:03




Instrument :
BNA_N
ClientSampleId :
ICVBN070924


Tgt Ion:188 Resp: 21027
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 10.3 7.8 11.6



#20
4,6-Dinitro-2-methylphenol
Concen: 0.350 ng
RT: 15.493 min Scan# 1598
Delta R.T. -0.000 min
Lab File: BN032564.D
Acq: 09 Jul 2024 00:03

Tgt Ion:198 Resp: 498
Ion Ratio Lower Upper
198 100
51 158.0 123.4 185.2
105 98.0 79.5 119.3









#21
4-Bromophenyl-phenylether
Concen: 0.384 ng
RT: 16.197 min Scan# 1655
Delta R.T. 0.000 min
Lab File: BN032564.D
Acq: 09 Jul 2024 00:03




Instrument :
BNA_N
ClientSampleId :
ICVBN070924


Tgt Ion:248 Resp: 4798
Ion Ratio Lower Upper
248 100
250 95.9 80.2 120.2
141 68.5 55.7 83.5



#22
Hexachlorobenzene
Concen: 0.388 ng
RT: 16.296 min Scan# 1663
Delta R.T. 0.000 min
Lab File: BN032564.D
Acq: 09 Jul 2024 00:03

Tgt Ion:284 Resp: 5332
Ion Ratio Lower Upper
284 100
142 34.6 27.6 41.4
249 28.6 23.6 35.4





#23

Atrazine

Concen: 0.342 ng

RT: 16.482 min Scan# 1678

Delta R.T. 0.000 min

Lab File: BN032564.D

Acq: 09 Jul 2024 00:03

Instrument :

BNA_N

ClientSampleId :

ICVBN070924





Tgt Ion:200 Resp: 3331

Ion Ratio Lower Upper

200 100

173 30.0 22.3 33.5

215 51.5 41.0 61.6



#24

Pentachlorophenol

Concen: 0.386 ng

RT: 16.693 min Scan# 1695

Delta R.T. 0.000 min

Lab File: BN032564.D

Acq: 09 Jul 2024 00:03




Tgt Ion:266 Resp: 1368


Ion Ratio Lower Upper

266 100

264 61.5 49.6 74.4

268 62.7 52.2 78.2









#25
 Phenanthrene
 Concen: 0.373 ng
 RT: 17.041 min Scan# 1723
 Delta R.T. 0.000 min
 Lab File: BN032564.D
 Acq: 09 Jul 2024 00:03




Instrument :
 BNA_N
 ClientSampleId :
 ICVBN070924


Tgt Ion:	178	Resp:	21278
Ion Ratio	Lower	Upper	
178	100		
176	18.9	15.2	22.8
179	15.3	12.2	18.4



#26
 Anthracene
 Concen: 0.353 ng
 RT: 17.140 min Scan# 1731
 Delta R.T. 0.000 min
 Lab File: BN032564.D
 Acq: 09 Jul 2024 00:03

Tgt Ion:	178	Resp:	16219
Ion Ratio	Lower	Upper	
178	100		
176	18.2	15.0	22.4
179	15.3	11.9	17.9





#27

Fluoranthene-d10

Concen: 0.370 ng

RT: 19.045 min Scan# 1957

Delta R.T. 0.000 min

Lab File: BN032564.D

Acq: 09 Jul 2024 00:03

Instrument :

BNA_N

ClientSampleId :

ICVBN070924





Tgt Ion:212 Resp: 20256

Ion Ratio Lower Upper

212 100

106 14.8 11.8 17.8

104 8.7 6.9 10.3



#28

Fluoranthene

Concen: 0.364 ng

RT: 19.072 min Scan# 1963

Delta R.T. 0.000 min

Lab File: BN032564.D

Acq: 09 Jul 2024 00:03




Tgt Ion:202 Resp: 25141


Ion Ratio Lower Upper

202 100

101 11.8 9.5 14.3

203 17.4 13.8 20.6





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.211 min Scan# 2402

Delta R.T. 0.000 min

Lab File: BN032564.D

Acq: 09 Jul 2024 00:03

Instrument :

BNA_N

ClientSampleId :

ICVBN070924





Tgt Ion:240 Resp: 16744

Ion Ratio Lower Upper

240 100

120 16.5 13.0 19.4

236 28.6 23.1 34.7



#30

Pyrene

Concen: 0.387 ng

RT: 19.439 min Scan# 2042

Delta R.T. 0.000 min

Lab File: BN032564.D

Acq: 09 Jul 2024 00:03




Tgt Ion:202 Resp: 26362


Ion Ratio Lower Upper

202 100

200 20.8 16.6 24.8

203 17.8 14.4 21.6









#31
 Terphenyl-d14
 Concen: 0.372 ng
 RT: 19.653 min Scan# 2088
 Delta R.T. 0.005 min
 Lab File: BN032564.D
 Acq: 09 Jul 2024 00:03




Instrument :
 BNA_N
 ClientSampleId :
 ICVBN070924


Tgt Ion:244 Resp: 15216
 Ion Ratio Lower Upper
 244 100
 212 9.0 7.2 10.8
 122 15.1 12.1 18.1



#32
 Benzo(a)anthracene
 Concen: 0.332 ng
 RT: 21.202 min Scan# 2401
 Delta R.T. 0.000 min
 Lab File: BN032564.D
 Acq: 09 Jul 2024 00:03

Tgt Ion:228 Resp: 18368
 Ion Ratio Lower Upper
 228 100
 226 27.4 21.7 32.5
 229 20.1 16.0 24.0





#33

Chrysene

Concen: 0.404 ng

RT: 21.247 min Scan# 2406

Delta R.T. 0.000 min

Lab File: BN032564.D

Acq: 09 Jul 2024 00:03

Instrument :

BNA_N

ClientSampleId :

ICVBN070924





Tgt Ion:228 Resp: 26567

Ion Ratio Lower Upper

228 100

226 30.2 24.1 36.1

229 20.1 16.1 24.1



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.316 ng

RT: 21.095 min Scan# 2389

Delta R.T. -0.009 min

Lab File: BN032564.D

Acq: 09 Jul 2024 00:03




Tgt Ion:149 Resp: 11875


Ion Ratio Lower Upper

149 100

167 26.7 21.6 32.4


279 3.9 3.0 4.6








#35
 Perylene-d12
 Concen: 0.400 ng
 RT: 23.426 min Scan# 21
 Delta R.T. 0.000 min
 Lab File: BN032564.D
 Acq: 09 Jul 2024 00:03

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN070924







Tgt Ion:264 Resp: 16900
 Ion Ratio Lower Upper
 264 100
 260 25.8 21.0 31.6
 265 95.9 76.6 115.0



#36
 Indeno(1,2,3-cd)pyrene
 Concen: 0.368 ng
 RT: 25.665 min Scan# 3649
 Delta R.T. -0.003 min
 Lab File: BN032564.D
 Acq: 09 Jul 2024 00:03

Tgt Ion:276 Resp: 23746
 Ion Ratio Lower Upper
 276 100
 138 26.6 21.4 32.0
 277 25.2 19.9 29.9









#37
Benzo(b)fluoranthene
Concen: 0.352 ng
RT: 22.759 min Scan# 2655
Delta R.T. 0.000 min
Lab File: BN032564.D
Acq: 09 Jul 2024 00:03




Instrument :
BNA_N
ClientSampleId :
ICVBN070924


Tgt Ion:252 Resp: 19810
Ion Ratio Lower Upper
252 100
253 28.5 22.2 33.4
125 19.4 15.4 23.0



#38
Benzo(k)fluoranthene
Concen: 0.389 ng
RT: 22.803 min Scan# 2670
Delta R.T. -0.003 min
Lab File: BN032564.D
Acq: 09 Jul 2024 00:03


Tgt Ion:252 Resp: 25257
Ion Ratio Lower Upper
252 100
253 28.0 21.9 32.9
125 19.5 15.3 22.9








#39
Benzo(a)pyrene
Concen: 0.366 ng
RT: 23.335 min Scan# 21
Delta R.T. -0.003 min
Lab File: BN032564.D
Acq: 09 Jul 2024 00:03


Instrument :
BNA_N
ClientSampleId :
ICVBN070924





Tgt Ion:252 Resp: 19522
Ion Ratio Lower Upper
252 100
253 32.6 25.5 38.3
125 25.8 20.7 31.1




#40
Dibenzo(a,h)anthracene
Concen: 0.367 ng
RT: 25.680 min Scan# 3654
Delta R.T. -0.003 min
Lab File: BN032564.D
Acq: 09 Jul 2024 00:03




Tgt Ion:278 Resp: 18807
Ion Ratio Lower Upper
278 100
139 21.6 17.8 26.6
279 30.7 24.4 36.6





#41
Benzo(g,h,i)perylene
Concen: 0.375 ng
RT: 26.349 min Scan# 3883
Delta R.T. 0.000 min
Lab File: BN032564.D
Acq: 09 Jul 2024 00:03

Instrument :
BNA_N
ClientSampleId :
ICVBN070924



Tgt Ion:	276	Resp:	20928
Ion	Ratio	Lower	Upper
276	100		
277	26.1	21.5	32.3
138	24.6	21.4	32.0

