

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN070924\
 Data File : BN032580.D
 Acq On : 09 Jul 2024 10:25
 Operator : MA/JU
 Sample : P3167-08
 Misc :
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 VPB196-HYD-20240703

Quant Time: Jul 09 12:02:09 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN070924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 09 02:39:48 2024
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 07/10/2024
 Supervised By :mohammad ahmed 07/11/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.588	152	4318	0.400	ng	0.00
7) Naphthalene-d8	10.389	136	13228	0.400	ng	0.01
13) Acenaphthene-d10	14.242	164	10062	0.400	ng	0.01
19) Phenanthrene-d10	17.016	188	22246	0.400	ng	0.01
29) Chrysene-d12	21.211	240	17711	0.400	ng	# 0.00
35) Perylene-d12	23.431	264	10573	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.227	112	551	0.050	ng	0.01
5) Phenol-d6	6.801	99	115	0.008	ng	-0.02
8) Nitrobenzene-d5	8.830	82	3395m	0.344	ng	0.03
11) 2-Methylnaphthalene-d10	11.990	152	74	0.004	ng	0.00
14) 2,4,6-Tribromophenol	15.775	330	1237	0.271	ng	0.01
15) 2-Fluorobiphenyl	12.874	172	11279	0.304	ng	0.01
27) Fluoranthene-d10	19.045	212	11867	0.205	ng	0.00
31) Terphenyl-d14	19.635	244	19398	0.449	ng	-0.01
Target Compounds						
2) 1,4-Dioxane	3.161	88	425	0.075	ng	# 75
34) Bis(2-ethylhexyl)phtha...	21.095	149	2125	0.053	ng	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN070924\
Data File : BN032580.D
Acq On : 09 Jul 2024 10:25
Operator : MA/JU
Sample : P3167-08
Misc :
ALS Vial : 36 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
VPB196-HYD-20240703

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 07/10/2024
Supervised By :mohammad ahmed 07/11/2024

Quant Time: Jul 09 12:02:09 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN070924.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 09 02:39:48 2024
Response via : Initial Calibration

