

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN070924\
 Data File : BN032590.D
 Acq On : 09 Jul 2024 16:27
 Operator : MA/JU
 Sample : P3173-02
 Misc :
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 VPB189-HYD-20240703

Quant Time: Jul 09 17:05:12 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN070924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 09 02:39:48 2024
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 07/10/2024
 Supervised By :mohammad ahmed 07/11/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.588	152	3740	0.400	ng	0.00
7) Naphthalene-d8	10.378	136	11014	0.400	ng	# 0.00
13) Acenaphthene-d10	14.242	164	7286	0.400	ng	0.01
19) Phenanthrene-d10	17.016	188	13936	0.400	ng	# 0.01
29) Chrysene-d12	21.220	240	11615	0.400	ng	# 0.00
35) Perylene-d12	23.440	264	12250	0.400	ng	# 0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.212	112	1046	0.111	ng	0.00
5) Phenol-d6	6.801	99	2	0.000	ng	-0.02
8) Nitrobenzene-d5	8.820	82	3348m	0.407	ng	0.02
11) 2-Methylnaphthalene-d10	11.998	152	97	0.006	ng	0.02
14) 2,4,6-Tribromophenol	15.775	330	977	0.296	ng	0.01
15) 2-Fluorobiphenyl	12.874	172	10092	0.376	ng	0.01
27) Fluoranthene-d10	19.045	212	11610	0.320	ng	0.00
31) Terphenyl-d14	19.639	244	12328	0.435	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.161	88	4570	0.928	ng	# 91
34) Bis(2-ethylhexyl)phtha...	21.104	149	1737	0.067	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN070924\
Data File : BN032590.D
Acq On : 09 Jul 2024 16:27
Operator : MA/JU
Sample : P3173-02
Misc :
ALS Vial : 46 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
VPB189-HYD-20240703

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 07/10/2024
Supervised By :mohammad ahmed 07/11/2024

Quant Time: Jul 09 17:05:12 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN070924.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 09 02:39:48 2024
Response via : Initial Calibration

