

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN070924\
 Data File : BN032602.D
 Acq On : 10 Jul 2024 00:24
 Operator : MA/JU
 Sample : P3167-08RE
 Misc :
 ALS Vial : 60 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 VPB196-HYD-20240703RE

Quant Time: Jul 10 01:03:59 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN070924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 09 02:39:48 2024
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 07/10/2024
 Supervised By :mohammad ahmed 07/11/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.588	152	5114	0.400	ng	0.00
7) Naphthalene-d8	10.389	136	15939	0.400	ng	0.01
13) Acenaphthene-d10	14.242	164	11709	0.400	ng	0.01
19) Phenanthrene-d10	17.004	188	25762	0.400	ng	# 0.00
29) Chrysene-d12	21.211	240	18748	0.400	ng	0.00
35) Perylene-d12	23.428	264	10330	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.227	112	583	0.045	ng	0.01
5) Phenol-d6	6.801	99	152	0.009	ng	-0.02
8) Nitrobenzene-d5	8.851	82	4061m	0.341	ng	0.05
11) 2-Methylnaphthalene-d10	11.967	152	1628	0.065	ng	-0.02
14) 2,4,6-Tribromophenol	15.775	330	1316	0.248	ng	0.01
15) 2-Fluorobiphenyl	12.874	172	13574	0.315	ng	0.01
27) Fluoranthene-d10	19.044	212	12884	0.192	ng	0.00
31) Terphenyl-d14	19.634	244	21296	0.465	ng	-0.01
Target Compounds						Qvalue
2) 1,4-Dioxane	3.161	88	492	0.073	ng	# 71
34) Bis(2-ethylhexyl)phtha...	21.094	149	2048	0.049	ng	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN070924\
Data File : BN032602.D
Acq On : 10 Jul 2024 00:24
Operator : MA/JU
Sample : P3167-08RE
Misc :
ALS Vial : 60 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
VPB196-HYD-20240703RE

Quant Time: Jul 10 01:03:59 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN070924.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 09 02:39:48 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/10/2024
Supervised By :mohammad ahmed 07/11/2024

