

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN070924\
 Data File : BN032564.D
 Acq On : 09 Jul 2024 00:03
 Operator : MA/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN070924

Quant Time: Jul 09 02:40:47 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN070924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 09 02:39:48 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	106	0.00
2	1,4-Dioxane	0.400	0.396	1.0	107	0.00
3	n-Nitrosodimethylamine	0.400	0.366	8.5	102	0.00
4 S	2-Fluorophenol	0.400	0.365	8.8	104	0.00
5 S	Phenol-d6	0.400	0.371	7.3	105	0.00
6	bis(2-Chloroethyl)ether	0.400	0.370	7.5	104	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	104	0.00
8 S	Nitrobenzene-d5	0.400	0.334	16.5	102	0.00
9	Naphthalene	0.400	0.387	3.3	104	0.00
10	Hexachlorobutadiene	0.400	0.390	2.5	103	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.390	2.5	103	0.00
12	2-Methylnaphthalene	0.400	0.386	3.5	103	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	101	0.00
14 S	2,4,6-Tribromophenol	0.400	0.330	17.5	96	0.00
15 S	2-Fluorobiphenyl	0.400	0.370	7.5	103	0.00
16	Acenaphthylene	0.400	0.351	12.3	101	0.00
17	Acenaphthene	0.400	0.376	6.0	102	0.00
18	Fluorene	0.400	0.375	6.3	101	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	100	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.350	12.5	91	0.00
21	4-Bromophenyl-phenylether	0.400	0.384	4.0	100	0.00
22	Hexachlorobenzene	0.400	0.388	3.0	102	0.00
23	Atrazine	0.400	0.342	14.5	92	0.00
24	Pentachlorophenol	0.400	0.386	3.5	92	0.00
25	Phenanthrene	0.400	0.373	6.8	100	0.00
26	Anthracene	0.400	0.353	11.8	96	0.00
27 SURR	Fluoranthene-d10	0.400	0.370	7.5	98	0.00
28	Fluoranthene	0.400	0.364	9.0	97	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	96	0.00
30	Pyrene	0.400	0.387	3.3	97	0.00
31 S	Terphenyl-d14	0.400	0.372	7.0	96	0.00
32	Benzo(a)anthracene	0.400	0.332	17.0	84	0.00
33	Chrysene	0.400	0.404	-1.0	102	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.316	21.0	88	0.00
35 I	Perylene-d12	0.400	0.400	0.0	98	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.368	8.0	102	0.00
37	Benzo(b)fluoranthene	0.400	0.352	12.0	94	0.00
38	Benzo(k)fluoranthene	0.400	0.389	2.8	97	0.00
39 C	Benzo(a)pyrene	0.400	0.366	8.5	95	0.00
40	Dibenzo(a,h)anthracene	0.400	0.367	8.3	103	0.00
41	Benzo(g,h,i)perylene	0.400	0.375	6.3	103	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

ClientSampleId :

ICVBN070924