

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN070924\
 Data File : BN032592.D
 Acq On : 09 Jul 2024 17:39
 Operator : MA/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jul 10 01:01:14 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN070924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 09 02:39:48 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	130	0.00
2	1,4-Dioxane	0.400	0.363	9.3	121	0.00
3	n-Nitrosodimethylamine	0.400	0.359	10.3	123	0.00
4 S	2-Fluorophenol	0.400	0.378	5.5	132	0.00
5 S	Phenol-d6	0.400	0.369	7.8	129	0.00
6	bis(2-Chloroethyl)ether	0.400	0.359	10.3	123	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	132	-0.01
8 S	Nitrobenzene-d5	0.400	0.341	14.8	133	-0.01
9	Naphthalene	0.400	0.389	2.8	134	-0.01
10	Hexachlorobutadiene	0.400	0.382	4.5	128	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.394	1.5	133	-0.01
12	2-Methylnaphthalene	0.400	0.392	2.0	134	-0.01
13 I	Acenaphthene-d10	0.400	0.400	0.0	129	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.335	16.3	124	0.00
15 S	2-Fluorobiphenyl	0.400	0.371	7.3	132	-0.01
16	Acenaphthylene	0.400	0.349	12.8	128	0.00
17	Acenaphthene	0.400	0.374	6.5	129	-0.01
18	Fluorene	0.400	0.370	7.5	127	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	120	-0.01
20	4,6-Dinitro-2-methylphenol	0.400	0.405	-1.3	150	-0.01
21	4-Bromophenyl-phenylether	0.400	0.392	2.0	124	-0.01
22	Hexachlorobenzene	0.400	0.390	2.5	124	-0.01
23	Atrazine	0.400	0.336	16.0	110	-0.01
24	Pentachlorophenol	0.400	0.405	-1.3	121	-0.01
25	Phenanthrene	0.400	0.380	5.0	123	0.00
26	Anthracene	0.400	0.354	11.5	116	0.00
27 SURR	Fluoranthene-d10	0.400	0.360	10.0	115	0.00
28	Fluoranthene	0.400	0.354	11.5	114	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	101	0.00
30	Pyrene	0.400	0.431	-7.7	114	0.00
31 S	Terphenyl-d14	0.400	0.410	-2.5	111	0.00
32	Benzo(a)anthracene	0.400	0.349	12.8	93	0.00
33	Chrysene	0.400	0.404	-1.0	108	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.366	8.5	107	0.00
35 I	Perylene-d12	0.400	0.400	0.0	101	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.389	2.8	111	0.00
37	Benzo(b)fluoranthene	0.400	0.364	9.0	101	0.00
38	Benzo(k)fluoranthene	0.400	0.403	-0.8	104	0.00
39 C	Benzo(a)pyrene	0.400	0.375	6.3	101	0.00
40	Dibenzo(a,h)anthracene	0.400	0.387	3.3	112	-0.01
41	Benzo(g,h,i)perylene	0.400	0.399	0.3	113	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4