

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN070924\
 Data File : BN032557.D
 Acq On : 08 Jul 2024 19:15
 Operator : MA/JU
 Sample : SSTDICC0.1
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :

Manual Integrations
 APPROVED

Reviewed By :Jagrut
 Upadhyay

Quant Time: Jul 09 02:32:05 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN070924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 09 02:30:59 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)	
Internal Standards							07/09/2024
1) 1,4-Dichlorobenzene-d4	7.595	152	3317	0.400	ng	0.00	Supervised By :mohammad
7) Naphthalene-d8	10.400	136	10784	0.400	ng	# 0.02	07/11/2024
13) Acenaphthene-d10	14.242	164	8240	0.400	ng	0.01	
19) Phenanthrene-d10	17.016	188	18438	0.400	ng	# 0.01	
29) Chrysene-d12	21.220	240	15575	0.400	ng	0.00	
35) Perylene-d12	23.431	264	15702	0.400	ng	0.00	
System Monitoring Compounds							
4) 2-Fluorophenol	5.219	112	824	0.098	ng	0.00	
5) Phenol-d6	6.880	99	942m	0.088	ng	0.06	
8) Nitrobenzene-d5	8.862	82	876m	0.114	ng	0.06	
11) 2-Methylnaphthalene-d10	12.028	152	1632	0.096	ng	0.05	
14) 2,4,6-Tribromophenol	15.787	330	308	0.082	ng	0.02	
15) 2-Fluorobiphenyl	12.895	172	2567	0.085	ng	0.03	
27) Fluoranthene-d10	19.054	212	4586	0.095	ng	0.00	
31) Terphenyl-d14	19.658	244	3851	0.101	ng	0.00	
Target Compounds							Qvalue
2) 1,4-Dioxane	3.161	88	562	0.129	ng		96
3) n-Nitrosodimethylamine	3.500	42	327	0.103	ng	#	95
6) bis(2-Chloroethyl)ether	7.068	93	403	0.129	ng	#	72
9) Naphthalene	10.453	128	2874	0.098	ng	#	78
10) Hexachlorobutadiene	10.677	225	646	0.110	ng	#	100
12) 2-Methylnaphthalene	12.100	142	1881	0.095	ng	#	92
16) Acenaphthylene	13.975	152	3095	0.087	ng		99
17) Acenaphthene	14.306	154	2317	0.094	ng		98
18) Fluorene	15.322	166	2977	0.092	ng		99
21) 4-Bromophenyl-phenylether	16.222	248	1068	0.098	ng		97
22) Hexachlorobenzene	16.309	284	1193	0.099	ng		98
23) Atrazine	16.495	200	818	0.096	ng	#	87
25) Phenanthrene	17.053	178	4749	0.095	ng		99
26) Anthracene	17.165	178	2984	0.074	ng		98
28) Fluoranthene	19.086	202	5887	0.097	ng		99
30) Pyrene	19.449	202	6295	0.099	ng		99
32) Benzo(a)anthracene	21.211	228	4351	0.084	ng		94
33) Chrysene	21.256	228	6748	0.110	ng		96
34) Bis(2-ethylhexyl)phtha...	21.103	149	3913	0.112	ng		99
36) Indeno(1,2,3-cd)pyrene	25.694	276	5565	0.093	ng		97
37) Benzo(b)fluoranthene	22.771	252	4820	0.092	ng	#	56
38) Benzo(k)fluoranthene	22.814	252	5819	0.097	ng	#	56
39) Benzo(a)pyrene	23.352	252	5004	0.101	ng	#	39
40) Dibenzo(a,h)anthracene	25.706	278	4423	0.093	ng	#	56
41) Benzo(g,h,i)perylene	26.367	276	4942	0.095	ng	#	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN070924\
Data File : BN032557.D
Acq On : 08 Jul 2024 19:15
Operator : MA/JU
Sample : SSTDICC0.1
Misc :
ALS Vial : 12 Sample Multiplier: 1


Instrument :
BNA_N
ClientSampleId :

Quant Time: Jul 09 02:32:05 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN070924.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 09 02:30:59 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay






#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.595 min Scan# 61
Delta R.T. 0.007 min
Lab File: BN032557.D
Acq: 08 Jul 2024 19:15

Instrument :

BNA_N


ClientSampleId :



Tgt Ion:152 Resp: 331
Ion Ratio Lower Upper
152 100
150 147.1 121.6 182.4
115 69.1 53.1 79.7


Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay






07/09/2024
Supervised By :mohammad
ahmed


07/11/2024



#2
1,4-Dioxane
Concen: 0.129 ng
RT: 3.161 min Scan# 25
Delta R.T. -0.000 min
Lab File: BN032557.D
Acq: 08 Jul 2024 19:15


Tgt Ion: 88 Resp: 562
Ion Ratio Lower Upper
88 100
43 23.8 20.4 30.6
58 60.9 51.0 76.6





#3
n-Nitrosodimethylamine
Concen: 0.103 ng
RT: 3.500 min Scan# 70
Delta R.T. 0.014 min
Lab File: BN032557.D
Acq: 08 Jul 2024 19:15



Instrument :
BNA_N
ClientSampleId :



Tgt Ion: 42 Resp: 32
Ion Ratio Lower Upper
42 100
74 152.6 126.8 190.2
44 16.5 8.7 13.1


Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay






07/09/2024
Supervised By :mohammad
ahmed


07/11/2024



#4
2-Fluorophenol
Concen: 0.098 ng
RT: 5.219 min Scan# 310
Delta R.T. 0.007 min
Lab File: BN032557.D
Acq: 08 Jul 2024 19:15


Tgt Ion:112 Resp: 824
Ion Ratio Lower Upper
112 100
64 50.5 41.2 61.8
63 23.5 21.6 32.4





#5
Phenol-d6
Concen: 0.088 ng m
RT: 6.880 min Scan# 540
Delta R.T. 0.058 min
Lab File: BN032557.D
Acq: 08 Jul 2024 19:15



Instrument :
BNA_N
ClientSampleId :



Tgt Ion: 99 Resp: 942
Ion Ratio Lower Upper
99 100
42 1.7 10.6 16.0#
71 0.0 25.2 37.8#


Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay






07/09/2024
Supervised By :mohammad
ahmed


07/11/2024



#6
bis(2-Chloroethyl)ether
Concen: 0.129 ng
RT: 7.068 min Scan# 566
Delta R.T. 0.029 min
Lab File: BN032557.D
Acq: 08 Jul 2024 19:15


Tgt Ion: 93 Resp: 403
Ion Ratio Lower Upper
93 100
63 95.8 54.1 81.1#
95 47.6 30.5 45.7#





#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.400 min Scan# 91
Delta R.T. 0.022 min
Lab File: BN032557.D
Acq: 08 Jul 2024 19:15



Instrument :
BNA_N
ClientSampleId :



Tgt Ion:136 Resp: 10784
Ion Ratio Lower Upper
136 100
137 13.6 10.0 15.0
54 10.4 6.6 9.8
68 9.8 5.8 8.8#


Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay






07/09/2024
Supervised By :mohammad
ahmed


07/11/2024



#8
Nitrobenzene-d5
Concen: 0.114 ng m
RT: 8.862 min Scan# 790
Delta R.T. 0.064 min
Lab File: BN032557.D
Acq: 08 Jul 2024 19:15


Tgt Ion: 82 Resp: 876
Ion Ratio Lower Upper
82 100
128 86.0 43.4 65.2#
54 72.0 40.2 60.2#





#9
Naphthalene
Concen: 0.098 ng
RT: 10.453 min Scan# 91
Delta R.T. 0.032 min
Lab File: BN032557.D
Acq: 08 Jul 2024 19:15



Instrument :
BNA_N
ClientSampleId :



Tgt Ion:128 Resp: 2874
Ion Ratio Lower Upper
128 100
129 22.9 10.2 15.4
127 22.1 11.4 17.0


Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay






07/09/2024
Supervised By :mohammad
ahmed


07/11/2024



#10
Hexachlorobutadiene
Concen: 0.110 ng
RT: 10.677 min Scan# 960
Delta R.T. 0.011 min
Lab File: BN032557.D
Acq: 08 Jul 2024 19:15


Tgt Ion:225 Resp: 646
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.0 50.6 76.0





#11
2-Methylnaphthalene-d10
Concen: 0.096 ng
RT: 12.028 min Scan# 1163
Delta R.T. 0.045 min
Lab File: BN032557.D
Acq: 08 Jul 2024 19:15



Instrument :
BNA_N
ClientSampleId :



Tgt Ion:152 Resp: 1632
Ion Ratio Lower Upper
152 100
151 12.9 16.4 24.6


Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




07/09/2024
Supervised By :mohammad
ahmed



07/11/2024




#12
2-Methylnaphthalene
Concen: 0.095 ng
RT: 12.100 min Scan# 1194
Delta R.T. 0.042 min
Lab File: BN032557.D
Acq: 08 Jul 2024 19:15




Tgt Ion:142 Resp: 1881
Ion Ratio Lower Upper
142 100
141 89.0 71.0 106.6
115 54.4 30.5 45.7#





#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.242 min Scan# 1480
Delta R.T. 0.011 min
Lab File: BN032557.D
Acq: 08 Jul 2024 19:15



Instrument :
BNA_N
ClientSampleId :



Tgt Ion:164 Resp: 8240
Ion Ratio Lower Upper
164 100
162 103.4 83.2 124.8
160 48.9 38.6 57.8


Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




07/09/2024
Supervised By :mohammad
ahmed



07/11/2024




#14
2,4,6-Tribromophenol
Concen: 0.082 ng
RT: 15.787 min Scan# 1622
Delta R.T. 0.025 min
Lab File: BN032557.D
Acq: 08 Jul 2024 19:15




Tgt Ion:330 Resp: 308
Ion Ratio Lower Upper
330 100
332 97.1 77.7 116.5
141 42.2 31.8 47.8





#15
2-Fluorobiphenyl
Concen: 0.085 ng
RT: 12.895 min Scan# 1355
Delta R.T. 0.032 min
Lab File: BN032557.D
Acq: 08 Jul 2024 19:15



Instrument :
BNA_N
ClientSampleId :



Tgt Ion:172 Resp: 256
Ion Ratio Lower Upper
172 100
171 40.0 28.5 42.7
170 29.7 19.2 28.8


Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




07/09/2024
Supervised By :mohammad
ahmed



07/11/2024




#16
Acenaphthylene
Concen: 0.087 ng
RT: 13.975 min Scan# 1456
Delta R.T. 0.021 min
Lab File: BN032557.D
Acq: 08 Jul 2024 19:15




Tgt Ion:152 Resp: 3095
Ion Ratio Lower Upper
152 100
151 18.8 15.6 23.4
153 13.1 10.5 15.7





#17
Acenaphthene
Concen: 0.094 ng
RT: 14.306 min Scan# 1487
Delta R.T. 0.011 min
Lab File: BN032557.D
Acq: 08 Jul 2024 19:15



Instrument :
BNA_N
ClientSampleId :



Tgt Ion:154 Resp: 231
Ion Ratio Lower Upper
154 100
153 116.7 90.9 136.3
152 56.7 44.4 66.6


Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




07/09/2024
Supervised By :mohammad
ahmed



07/11/2024




#18
Fluorene
Concen: 0.092 ng
RT: 15.322 min Scan# 1582
Delta R.T. 0.021 min
Lab File: BN032557.D
Acq: 08 Jul 2024 19:15




Tgt Ion:166 Resp: 2977
Ion Ratio Lower Upper
166 100
165 97.8 78.7 118.1
167 13.5 10.7 16.1





#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.016 min Scan# 1720
Delta R.T. 0.012 min
Lab File: BN032557.D
Acq: 08 Jul 2024 19:15



Instrument :
BNA_N
ClientSampleId :



Tgt Ion:188 Resp: 18438
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 11.8 7.8 11.6


Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




07/09/2024
Supervised By :mohammad
ahmed



07/11/2024




#21
4-Bromophenyl-phenylether
Concen: 0.098 ng
RT: 16.222 min Scan# 1657
Delta R.T. 0.025 min
Lab File: BN032557.D
Acq: 08 Jul 2024 19:15




Tgt Ion:248 Resp: 1068
Ion Ratio Lower Upper
248 100
250 99.7 80.2 120.2
141 74.5 55.7 83.5





#22
 Hexachlorobenzene
 Concen: 0.099 ng
 RT: 16.309 min Scan# 1664
 Delta R.T. 0.012 min
 Lab File: BN032557.D
 Acq: 08 Jul 2024 19:15


Instrument :
 BNA_N
 ClientSampleId :



Tgt Ion: 284 Resp: 119


Ion	Ratio	Lower	Upper
284	100		
142	35.3	27.6	41.4
249	27.7	23.6	35.4

Manual Integrations
APPROVED
 Reviewed By :Jagrut Upadhyay




07/09/2024
 Supervised By :mohammad ahmed

07/11/2024






#23
 Atrazine
 Concen: 0.096 ng
 RT: 16.495 min Scan# 1679
 Delta R.T. 0.012 min
 Lab File: BN032557.D
 Acq: 08 Jul 2024 19:15



Tgt Ion: 200 Resp: 818


Ion	Ratio	Lower	Upper
200	100		
173	39.0	22.3	33.5
215	57.1	41.0	61.6





#25
Phenanthrene
Concen: 0.095 ng
RT: 17.053 min Scan# 1724
Delta R.T. 0.012 min
Lab File: BN032557.D
Acq: 08 Jul 2024 19:15



Instrument :
BNA_N
ClientSampleId :



Tgt Ion:178 Resp: 4749
Ion Ratio Lower Upper
178 100
176 19.0 15.2 22.8
179 15.8 12.2 18.4


Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




07/09/2024
Supervised By :mohammad
ahmed



07/11/2024




#26
Anthracene
Concen: 0.074 ng
RT: 17.165 min Scan# 1733
Delta R.T. 0.025 min
Lab File: BN032557.D
Acq: 08 Jul 2024 19:15



Tgt Ion:178 Resp: 2984
Ion Ratio Lower Upper
178 100
176 19.7 15.0 22.4
179 15.7 11.9 17.9







#27
Fluoranthene-d10
Concen: 0.095 ng
RT: 19.054 min Scan# 1957
Delta R.T. 0.009 min
Lab File: BN032557.D
Acq: 08 Jul 2024 19:15

Instrument :
BNA_N
ClientSampleId :

Tgt Ion:212 Resp: 4580
Ion Ratio Lower Upper
212 100
106 15.6 11.8 17.8
104 9.4 6.9 10.3



Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay






07/09/2024
Supervised By :mohammad
ahmed


07/11/2024



#28
Fluoranthene
Concen: 0.097 ng
RT: 19.086 min Scan# 1966
Delta R.T. 0.014 min
Lab File: BN032557.D
Acq: 08 Jul 2024 19:15

Tgt Ion:202 Resp: 5887
Ion Ratio Lower Upper
202 100
101 12.0 9.5 14.3
203 18.1 13.8 20.6





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.220 min Scan# 2403

Delta R.T. 0.009 min

Lab File: BN032557.D

Acq: 08 Jul 2024 19:15

Instrument :

BNA_N



ClientSampleId :

Tgt Ion:240	Resp:	1557
Ion Ratio	Lower	Upper
240	100	
120	17.6	13.0 19.4
236	28.8	23.1 34.7

Manual Integrations



APPROVED

Reviewed By :Jagrut
Upadhyay



07/09/2024
Supervised By :mohammad
ahmed

07/11/2024



#30

Pyrene

Concen: 0.099 ng




RT: 19.449 min Scan# 2044


Delta R.T. 0.009 min

Lab File: BN032557.D

Acq: 08 Jul 2024 19:15

Tgt Ion:202	Resp:	6295
Ion Ratio	Lower	Upper
202	100	
200	20.7	16.6 24.8
203	19.0	14.4 21.6






#31
 Terphenyl-d14
 Concen: 0.101 ng
 RT: 19.658 min Scan# 2089
 Delta R.T. 0.009 min
 Lab File: BN032557.D
 Acq: 08 Jul 2024 19:15

Instrument :

BNA_N


ClientSampleId :



Tgt Ion:244 Resp: 385
 Ion Ratio Lower Upper
 244 100
 212 11.6 7.2 10.8
 122 18.0 12.1 18.1


Manual Integrations
 APPROVED

Reviewed By :Jagrut
 Upadhyay




07/09/2024
 Supervised By :mohammad
 ahmed



07/11/2024




#32
 Benzo(a)anthracene
 Concen: 0.084 ng
 RT: 21.211 min Scan# 2402
 Delta R.T. 0.009 min
 Lab File: BN032557.D
 Acq: 08 Jul 2024 19:15




Tgt Ion:228 Resp: 4351
 Ion Ratio Lower Upper
 228 100
 226 30.3 21.7 32.5
 229 22.8 16.0 24.0





#33
Chrysene
Concen: 0.110 ng
RT: 21.256 min Scan# 2407
Delta R.T. 0.009 min
Lab File: BN032557.D
Acq: 08 Jul 2024 19:15



Instrument :
BNA_N
ClientSampleId :



Tgt Ion:228 Resp: 6748
Ion Ratio Lower Upper
228 100
226 31.8 24.1 36.1
229 22.7 16.1 24.1


Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




07/09/2024
Supervised By :mohammad
ahmed



07/11/2024




#34
Bis(2-ethylhexyl)phthalate
Concen: 0.112 ng
RT: 21.103 min Scan# 2390
Delta R.T. -0.000 min
Lab File: BN032557.D
Acq: 08 Jul 2024 19:15



Tgt Ion:149 Resp: 3913
Ion Ratio Lower Upper
149 100
167 27.5 21.6 32.4
279 4.1 3.0 4.6






#35
 Perylene-d12
 Concen: 0.400 ng
 RT: 23.431 min Scan# 21
 Delta R.T. 0.006 min
 Lab File: BN032557.D
 Acq: 08 Jul 2024 19:15

Instrument :

BNA_N



ClientSampleId :



Tgt Ion:264 Resp: 1570
 Ion Ratio Lower Upper
 264 100
 260 26.7 21.0 31.6
 265 104.1 76.6 115.0


Manual Integrations
 APPROVED

Reviewed By :Jagrut
 Upadhyay




07/09/2024
 Supervised By :mohammad
 ahmed



07/11/2024




#36
 Indeno(1,2,3-cd)pyrene
 Concen: 0.093 ng
 RT: 25.694 min Scan# 3659
 Delta R.T. 0.026 min
 Lab File: BN032557.D
 Acq: 08 Jul 2024 19:15



Tgt Ion:276 Resp: 5565
 Ion Ratio Lower Upper
 276 100
 138 24.5 21.4 32.0
 277 24.2 19.9 29.9





#37

Benzo(b)fluoranthene

Concen: 0.092 ng

RT: 22.771 min Scan# 2659

Delta R.T. 0.012 min


Lab File: BN032557.D

Acq: 08 Jul 2024 19:15

Instrument :

BN032557.D

ClientSampleId :



Tgt Ion:252 Resp: 4820

Ion Ratio Lower Upper

252 100

253 45.5 22.2 33.4


125 45.8 15.4 23.0

Manual Integrations

APPROVED

Reviewed By :Jagrut


Upadhyay



07/09/2024

Supervised By :mohammad
ahmed

07/11/2024



#38

Benzo(k)fluoranthene

Concen: 0.097 ng

RT: 22.814 min Scan# 2674

Delta R.T. 0.009 min

Lab File: BN032557.D

Acq: 08 Jul 2024 19:15



Tgt Ion:252 Resp: 5819


Ion Ratio Lower Upper

252 100

253 45.4 21.9 32.9#


125 44.9 15.3 22.9#





#39
 Benzo(a)pyrene
 Concen: 0.101 ng
 RT: 23.352 min Scan# 2853
 Delta R.T. 0.015 min
 Lab File: BN032557.D
 Acq: 08 Jul 2024 19:15


Instrument :
 BNA_N
 ClientSampleId :



Tgt Ion: 252 Resp: 5004



Ion	Ratio	Lower	Upper
252	100		
253	59.9	25.5	38.3
125	64.0	20.7	31.1

Manual Integrations
APPROVED
 Reviewed By :Jagrut Upadhyay



07/09/2024
 Supervised By :mohammad ahmed




07/11/2024




#40
 Dibenzo(a,h)anthracene
 Concen: 0.093 ng
 RT: 25.706 min Scan# 3663
 Delta R.T. 0.023 min
 Lab File: BN032557.D
 Acq: 08 Jul 2024 19:15

Tgt Ion: 278 Resp: 4423

Ion	Ratio	Lower	Upper
278	100		
139	41.3	17.8	26.6
279	56.3	24.4	36.6






#41
Benzo(g,h,i)perylene
Concen: 0.095 ng
RT: 26.367 min Scan# 3889
Delta R.T. 0.017 min
Lab File: BN032557.D
Acq: 08 Jul 2024 19:15

Instrument :

BNA_N



ClientSampleId :



Tgt Ion:276 Resp: 494
Ion Ratio Lower Upper
276 100
277 35.5 21.5 32.3
138 38.6 21.4 32.0

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay



07/09/2024
Supervised By :mohammad
ahmed

07/11/2024