

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN071020\
 Data File : BN011138.D
 Acq On : 09 Jul 2020 18:38
 Operator : CG/JU
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDC00.4

Quant Time: Jul 09 19:07:55 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN063020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 09 15:33:27 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	116	0.00
2	1,4-Dioxane	0.400	0.346	13.5	97	0.00
3	n-Nitrosodimethylamine	0.400	0.352	12.0	106	0.00
4 S	2-Fluorophenol	0.400	0.351	12.3	99	0.00
5 S	Phenol-d6	0.400	0.418	-4.5	115	0.00
6	bis(2-Chloroethyl)ether	0.400	0.382	4.5	100	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	98	0.00
8 S	Nitrobenzene-d5	0.400	0.416	-4.0	102	0.00
9	Naphthalene	0.400	0.398	0.5	98	0.00
10	Hexachlorobutadiene	0.400	0.407	-1.7	92	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.388	3.0	91	0.00
12	2-Methylnaphthalene	0.400	0.388	3.0	94	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	96	0.00
14 S	2,4,6-Tribromophenol	0.400	0.416	-4.0	99	0.00
15 S	2-Fluorobiphenyl	0.400	0.411	-2.7	97	0.00
16	Acenaphthylene	0.400	0.412	-3.0	99	0.00
17	Acenaphthene	0.400	0.404	-1.0	97	0.00
18	Fluorene	0.400	0.418	-4.5	99	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	99	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.521	-30.3#	140	0.00
21	4-Bromophenyl-phenylether	0.400	0.401	-0.3	97	0.00
22	Hexachlorobenzene	0.400	0.400	0.0	97	0.00
23	Atrazine	0.400	0.398	0.5	95	0.01
24	Pentachlorophenol	0.400	0.432	-8.0	108	0.00
25	Phenanthrene	0.400	0.395	1.3	95	0.00
26	Anthracene	0.400	0.396	1.0	95	0.00
27 SURR	Fluoranthene-d10	0.400	0.403	-0.8	98	0.00
28	Fluoranthene	0.400	0.406	-1.5	99	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	89	0.00
30	Pyrene	0.400	0.416	-4.0	97	0.00
31 S	Terphenyl-d14	0.400	0.421	-5.2	97	0.00
32	Benzo(a)anthracene	0.400	0.376	6.0	82	0.00
33	Chrysene	0.400	0.434	-8.5	97	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.425	-6.2	95	0.00
35	Indeno(1,2,3-cd)pyrene	0.400	0.491	-22.7	112	0.00
36 I	Perylene-d12	0.400	0.400	0.0	119	0.00
37	Benzo(b)fluoranthene	0.400	0.331	17.3	95	0.00
38	Benzo(k)fluoranthene	0.400	0.422	-5.5	112	0.00
39 C	Benzo(a)pyrene	0.400	0.405	-1.3	119	0.00
40	Dibenzo(a,h)anthracene	0.400	0.354	11.5	110	0.00
41	Benzo(g,h,i)perylene	0.400	0.354	11.5	109	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN071020\
Data File : BN011138.D
Acq On : 09 Jul 2020 18:38
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jul 09 19:07:55 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN063020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 09 15:33:27 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0