

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN071120\
 Data File : BN011148.D
 Acq On : 10 Jul 2020 14:36
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN071120

Quant Time: Jul 10 15:06:59 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN071120.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 10 14:35:44 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	106	0.00
2	1,4-Dioxane	0.572	0.580	-1.4	100	0.00
3	n-Nitrosodimethylamine	0.288	0.267	7.3	104	0.00
4 S	2-Fluorophenol	0.986	0.934	5.3	96	0.00
5 S	Phenol-d6	1.115	1.096	1.7	102	0.00
6	bis(2-Chloroethyl)ether	1.021	0.930	8.9	104	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	100	0.00
8 S	Nitrobenzene-d5	0.355	0.358	-0.8	100	0.00
9	Naphthalene	1.190	1.200	-0.8	102	0.00
10	Hexachlorobutadiene	0.295	0.313	-6.1	103	0.00
11 SURR	2-Methylnaphthalene-d10	0.649	0.715	-10.2	105	0.00
12	2-Methylnaphthalene	0.749	0.813	-8.5	105	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	101	0.00
14 S	2,4,6-Tribromophenol	0.178	0.177	0.6	98	0.00
15 S	2-Fluorobiphenyl	1.874	1.906	-1.7	96	0.00
16	Acenaphthylene	2.036	2.006	1.5	102	0.00
17	Acenaphthene	1.358	1.373	-1.1	102	0.00
18	Fluorene	1.742	1.729	0.7	88	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	101	0.00
20	4,6-Dinitro-2-methylphenol	0.053	0.047	11.3	89	0.00
21	4-Bromophenyl-phenylether	0.272	0.264	2.9	98	0.00
22	Hexachlorobenzene	0.316	0.320	-1.3	101	0.00
23	Atrazine	0.191	0.189	1.0	100	0.00
24	Pentachlorophenol	0.094	0.082	12.8	95	0.00
25	Phenanthrene	1.285	1.276	0.7	101	0.00
26	Anthracene	1.067	1.045	2.1	100	0.00
27 SURR	Fluoranthene-d10	1.261	1.217	3.5	96	0.00
28	Fluoranthene	1.541	1.467	4.8	92	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	107	0.00
30	Pyrene	1.527	1.439	5.8	89	0.00
31 S	Terphenyl-d14	1.041	0.995	4.4	97	0.00
32	Benzo(a)anthracene	1.296	1.266	2.3	108	0.01
33	Chrysene	1.584	1.734	-9.5	118	0.01
34	Bis(2-ethylhexyl)phthalate	0.534	0.557	-4.3	113	0.01
35	Indeno(1,2,3-cd)pyrene	1.656	1.742	-5.2	97	0.01
36 I	Perylene-d12	1.000	1.000	0.0	98	0.00
37	Benzo(b)fluoranthene	1.607	1.491	7.2	100	0.00
38	Benzo(k)fluoranthene	1.816	1.925	-6.0	103	0.00
39 C	Benzo(a)pyrene	1.447	1.448	-0.1	97	0.00
40	Dibenzo(a,h)anthracene	1.533	1.558	-1.6	97	0.00
41	Benzo(g,h,i)perylene	1.683	1.619	3.8	91	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN071120\
Data File : BN011148.D
Acq On : 10 Jul 2020 14:36
Operator : CG/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
ICVBN071120

Quant Time: Jul 10 15:06:59 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN071120.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 10 14:35:44 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0