


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN071120\
 Data File : BN011139.D
 Acq On : 10 Jul 2020 08:59
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN071120.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Jul 10 14:35:44 2020



AutoFind: Scans 2476, 2477, 2478; Background Corrected with Scan 2470

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	25.7	41973	PASS
68	69	0.00	2	1.5	999	PASS
69	198	0.00	100	41.9	68538	PASS
70	69	0.00	2	0.4	272	PASS
127	198	10	80	50.4	82386	PASS
197	198	0.00	2	0.7	1080	PASS
198	198	100	100	100.0	163464	PASS
199	198	5	9	6.5	10625	PASS
275	198	10	60	27.8	45384	PASS
365	198	1	100	5.2	8550	PASS
441	198	0.01	100	10.5	17128	PASS
442	198	50	100	73.6	120336	PASS
443	442	15	24	18.8	22601	PASS