

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN071320\  
 Data File : BN011155.D  
 Acq On : 13 Jul 2020 16:23  
 Operator : CG/JU  
 Sample : PB130053BL  
 Misc :  
 ALS Vial : 3 Sample Multiplier: 1

Instrument :  
 BNA\_N  
 ClientSampleId :  
 PB130053BL

Quant Time: Jul 13 16:58:39 2020  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN071120.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Mon Jul 13 13:53:03 2020  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc | Units | Dev(Min) |
|-----------------------------|-------|------|----------|------|-------|----------|
| 1) 1,4-Dichlorobenzene-d4   | 7.52  | 152  | 1402     | 0.40 | ng    | 0.00     |
| 7) Naphthalene-d8           | 10.28 | 136  | 3966     | 0.40 | ng    | 0.03     |
| 13) Acenaphthene-d10        | 14.14 | 164  | 2279     | 0.40 | ng    | 0.01     |
| 19) Phenanthrene-d10        | 16.91 | 188  | 4514     | 0.40 | ng    | 0.02     |
| 29) Chrysene-d12            | 21.11 | 240  | 4415     | 0.40 | ng    | 0.01     |
| 36) Perylene-d12            | 23.25 | 264  | 4033     | 0.40 | ng    | 0.01     |
| System Monitoring Compounds |       |      |          |      |       |          |
| 4) 2-Fluorophenol           | 5.18  | 112  | 1740     | 0.50 | ng    | 0.00     |
| 5) Phenol-d6                | 6.73  | 99   | 1637     | 0.42 | ng    | 0.00     |
| 8) Nitrobenzene-d5          | 8.68  | 82   | 2065     | 0.59 | ng    | 0.03     |
| 11) 2-Methylnaphthalene-d10 | 11.89 | 152  | 2196     | 0.34 | ng    | 0.03     |
| 14) 2,4,6-Tribromophenol    | 15.68 | 330  | 406      | 0.40 | ng    | 0.04     |
| 15) 2-Fluorobiphenyl        | 12.77 | 172  | 4700     | 0.44 | ng    | 0.03     |
| 27) Fluoranthene-d10        | 18.94 | 212  | 5425     | 0.38 | ng    | 0.01     |
| 31) Terphenyl-d14           | 19.55 | 244  | 8197     | 0.71 | ng    | 0.00     |

| Target Compounds | Qvalue |
|------------------|--------|
|------------------|--------|

-----

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN071320\  
Data File : BN011155.D  
Acq On : 13 Jul 2020 16:23  
Operator : CG/JU  
Sample : PB130053BL  
Misc :  
ALS Vial : 3 Sample Multiplier: 1

Instrument :  
BNA\_N  
ClientSampleId :  
PB130053BL

Quant Time: Jul 13 16:58:39 2020  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN071120.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Mon Jul 13 13:53:03 2020  
Response via : Initial Calibration

