

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN071320\
 Data File : BN011158.D
 Acq On : 13 Jul 2020 18:11
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jul 13 18:43:44 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN071120.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 13 11:25:24 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	83	0.00
2	1,4-Dioxane	0.400	0.397	0.8	77	0.00
3	n-Nitrosodimethylamine	0.400	0.415	-3.7	91	0.00
4 S	2-Fluorophenol	0.400	0.388	3.0	78	0.00
5 S	Phenol-d6	0.400	0.435	-8.7	88	0.00
6	bis(2-Chloroethyl)ether	0.400	0.396	1.0	89	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	88	0.00
8 S	Nitrobenzene-d5	0.400	0.382	4.5	84	0.00
9	Naphthalene	0.400	0.401	-0.3	89	0.00
10	Hexachlorobutadiene	0.400	0.445	-11.2	95	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.382	4.5	81	0.00
12	2-Methylnaphthalene	0.400	0.386	3.5	82	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	84	0.00
14 S	2,4,6-Tribromophenol	0.400	0.443	-10.7	90	0.00
15 S	2-Fluorobiphenyl	0.400	0.408	-2.0	79	0.00
16	Acenaphthylene	0.400	0.393	1.8	84	0.00
17	Acenaphthene	0.400	0.401	-0.3	83	0.00
18	Fluorene	0.400	0.439	-9.7	80	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	85	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.499	-24.7	107	0.00
21	4-Bromophenyl-phenylether	0.400	0.464	-16.0	98	0.00
22	Hexachlorobenzene	0.400	0.460	-15.0	96	0.00
23	Atrazine	0.400	0.453	-13.2	95	0.00
24	Pentachlorophenol	0.400	0.429	-7.2	98	0.00
25	Phenanthrene	0.400	0.402	-0.5	85	0.00
26	Anthracene	0.400	0.391	2.3	83	0.00
27 SURR	Fluoranthene-d10	0.400	0.392	2.0	82	0.00
28	Fluoranthene	0.400	0.386	3.5	78	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	84	0.00
30	Pyrene	0.400	0.405	-1.3	75	0.00
31 S	Terphenyl-d14	0.400	0.419	-4.7	83	0.00
32	Benzo(a)anthracene	0.400	0.386	3.5	83	0.00
33	Chrysene	0.400	0.404	-1.0	85	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.445	-11.2	95	0.00
35	Indeno(1,2,3-cd)pyrene	0.400	0.451	-12.7	80	0.00
36 I	Perylene-d12	0.400	0.400	0.0	87	0.00
37	Benzo(b)fluoranthene	0.400	0.359	10.3	86	0.00
38	Benzo(k)fluoranthene	0.400	0.422	-5.5	91	0.00
39 C	Benzo(a)pyrene	0.400	0.385	3.8	83	0.00
40	Dibenzo(a,h)anthracene	0.400	0.354	11.5	75	0.00
41	Benzo(g,h,i)perylene	0.400	0.365	8.8	77	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN071320\
Data File : BN011158.D
Acq On : 13 Jul 2020 18:11
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jul 13 18:43:44 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN071120.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jul 13 11:25:24 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area	% Dev(min)
----------	--------	-------	------	------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0