

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN071420\
 Data File : BN011160.D
 Acq On : 14 Jul 2020 14:27
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jul 14 15:13:17 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN071120.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 14 15:08:13 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	94	0.00
2	1,4-Dioxane	0.572	0.514	10.1	79	-0.04
3	n-Nitrosodimethylamine	0.288	0.294	-2.1	102	0.00
4 S	2-Fluorophenol	0.986	0.966	2.0	88	0.00
5 S	Phenol-d6	1.115	1.183	-6.1	97	0.00
6	bis(2-Chloroethyl)ether	1.021	1.033	-1.2	102	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	100	0.00
8 S	Nitrobenzene-d5	0.355	0.364	-2.5	101	0.00
9	Naphthalene	1.190	1.098	7.7	93	0.00
10	Hexachlorobutadiene	0.295	0.245	16.9	80	0.00
11 SURR	2-Methylnaphthalene-d10	0.649	0.635	2.2	93	0.00
12	2-Methylnaphthalene	0.749	0.735	1.9	95	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	112	0.00
14 S	2,4,6-Tribromophenol	0.178	0.199	-11.8	122	0.00
15 S	2-Fluorobiphenyl	1.874	1.711	8.7	95	0.00
16	Acenaphthylene	2.036	1.875	7.9	106	0.00
17	Acenaphthene	1.358	1.265	6.8	104	0.00
18	Fluorene	1.742	1.709	1.9	96	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	131	0.00
20	4,6-Dinitro-2-methylphenol	0.053	0.057	-7.5	139	0.00
21	4-Bromophenyl-phenylether	0.272	0.249	8.5	120	0.00
22	Hexachlorobenzene	0.316	0.262	17.1	107	0.00
23	Atrazine	0.191	0.192	-0.5	132	0.00
24	Pentachlorophenol	0.094	0.091	3.2	137	0.00
25	Phenanthrene	1.285	1.228	4.4	126	0.00
26	Anthracene	1.067	1.015	4.9	126	0.00
27 SURR	Fluoranthene-d10	1.261	1.245	1.3	128	0.00
28	Fluoranthene	1.541	1.490	3.3	121	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	145	0.00
30	Pyrene	1.527	1.411	7.6	118	0.00
31 S	Terphenyl-d14	1.041	1.006	3.4	133	0.00
32	Benzo(a)anthracene	1.296	1.225	5.5	141	0.00
33	Chrysene	1.584	1.433	9.5	132	0.00
34	Bis(2-ethylhexyl)phthalate	0.534	0.604	-13.1	167#	0.00
35	Indeno(1,2,3-cd)pyrene	1.656	1.431	13.6	108	0.00
36 I	Perylene-d12	1.000	1.000	0.0	136	0.00
37	Benzo(b)fluoranthene	1.607	1.396	13.1	131	0.00
38	Benzo(k)fluoranthene	1.816	1.672	7.9	125	0.00
39 C	Benzo(a)pyrene	1.447	1.279	11.6	119	0.00
40	Dibenzo(a,h)anthracene	1.533	1.245	18.8	107	0.00
41	Benzo(g,h,i)perylene	1.683	1.358	19.3	107	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN071420\
Data File : BN011160.D
Acq On : 14 Jul 2020 14:27
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jul 14 15:13:17 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN071120.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 14 15:08:13 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0