



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN071522\
 Data File : BN020745.D
 Acq On : 15 Jul 2022 13:13
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN062622.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Sat Jul 16 03:21:59 2022






AutoFind: Scans 2480, 2481, 2482; Background Corrected with Scan 2473

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	49.8	159350	PASS
68	69	0.00	2	1.9	2690	PASS
69	198	0.00	100	45.3	144987	PASS
70	69	0.00	2	0.5	747	PASS
127	198	10	80	55.2	176405	PASS
197	198	0.00	2	0.2	567	PASS
198	198	100	100	100.0	319744	PASS
199	198	5	9	6.8	21616	PASS
275	198	10	60	25.5	81592	PASS
365	198	1	100	3.1	10048	PASS
441	198	0.01	100	12.3	39437	PASS
442	442	50	100	100.0	247701	PASS
443	442	15	24	20.0	49485	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN071522\
 Data File : BN020745.D
 Acq On : 15 Jul 2022 13:13
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP

Quant Time: Jul 16 05:28:30 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jul 16 05:28:22 2022
 Response via : Initial Calibration



TIC: BN020745.D\data.ms

(70) Pentachlorophenol (C)

16.875min (0.000) 19316.90 ng




response 323502

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	68.00
264.00	61.60	67.11
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN071522\
Data File : BN020745.D
Acq On : 15 Jul 2022 13:13
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Jul 16 05:28:30 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Jul 16 05:28:22 2022
Response via : Initial Calibration



TIC: BN020745.D\data.ms

(77) Benzidine

19.475min (0.000) 0.00 ng

response 2475605

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	14.51
183.00	13.20	12.39
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
7/15/2022	BNA_N	<u>BN020745.D</u>
Compound Name	Response	Retention Time
DDT	1138294	20.722
DDD	26291	20.28
DDE	1899	19.775
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
28190	1166484	2.42

Instrument :
BNA_N
ClientSampleId :
DFTPP