

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN071624\
 Data File : BN032806.D
 Acq On : 18 Jul 2024 07:45
 Operator : MA/JU
 Sample : PB162054BS
 Misc :
 ALS Vial : 78 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SLCS054

Quant Time: Jul 18 08:12:09 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN071224.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 17 02:35:25 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.540	152	6373	0.400	ng/ul	-0.01
4) Naphthalene-d8	10.310	136	20135	0.400	ng/ul	-0.02
9) Acenaphthene-d10	14.193	164	10766	0.400	ng/ul	0.00
13) Phenanthrene-d10	16.961	188	19284	0.400	ng/ul	0.00
17) Chrysene-d12	21.170	240	18343	0.400	ng/ul	0.00
23) Perylene-d12	23.368	264	22341	0.400	ng/ul	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.097	96	4457	0.582	ng/ul	0.00
6) 2-Methylnaphthalene-d10	11.916	152	9824	0.323	ng/ul	-0.02
18) Fluoranthene-d10	18.999	212	18786	0.314	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.126	88	12044	1.382	ng/ul	91
5) Naphthalene	10.360	128	16716	0.309	ng/ul	98
7) 2-Methylnaphthalene	11.993	142	10344	0.306	ng/ul	99
8) 1-Methylnaphthalene	12.213	142	10431	0.299	ng/ul	100
10) Acenaphthylene	13.915	152	14264	0.328	ng/ul	99
11) Acenaphthene	14.253	153	11107	0.309	ng/ul	99
12) Fluorene	15.257	166	11360	0.276	ng/ul	99
14) Pentachlorophenol	16.640	266	2564	0.758	ng/ul	99
15) Phenanthrene	16.999	178	16577	0.316	ng/ul	100
16) Anthracene	17.096	178	15556	0.363	ng/ul	100
19) Fluoranthene	19.031	202	21034	0.282	ng/ul	99
20) Pyrene	19.398	202	22508	0.292	ng/ul	98
21) Benzo(a)anthracene	21.156	228	21598	0.370	ng/ul	99
22) Chrysene	21.208	228	24122	0.286	ng/ul	98
24) Benzo(b)fluoranthene	22.707	252	23547	0.299	ng/ul	98
25) Benzo(k)fluoranthene	22.751	252	26181	0.273	ng/ul	98
26) Benzo(a)pyrene	23.275	252	23513	0.345	ng/ul	96
27) Indeno(1,2,3-cd)pyrene	25.582	276	30619	0.331	ng/ul#	99
28) Dibenzo(a,h)anthracene	25.582	278	24071	0.336	ng/ul#	96
29) Benzo(g,h,i)perylene	26.267	276	24044	0.311	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN071624\
Data File : BN032806.D
Acq On : 18 Jul 2024 07:45
Operator : MA/JU
Sample : PB162054BS
Misc :
ALS Vial : 78 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SLCS054

Quant Time: Jul 18 08:12:09 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN071224.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jul 17 02:35:25 2024
Response via : Initial Calibration

