

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN071921\
 Data File : BN015615.D
 Acq On : 19 Jul 2021 19:23
 Operator : CG/JU
 Sample : SP5623
 Misc : SFAM SURROGATE
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SP5623

Manual Integrations
 APPROVED

mohammad
 7/20/2021 11:38:53 AM

Quant Time: Jul 20 01:11:01 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN062121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 19 13:54:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.775	152	2550	0.400	ng/ul	# 0.00
4) Naphthalene-d8	10.562	136	9239	0.400	ng/ul	# 0.00
9) Acenaphthene-d10	14.418	164	6166	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.166	188	12526m	0.400	ng/ul	0.00
17) Chrysene-d12	21.366	240	9405	0.400	ng/ul	# 0.00
23) Perylene-d12	23.702	264	9952	0.400	ng/ul	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.264	96	37765	13.139	ng/ul	-0.02
6) 2-Methylnaphthalene-d10	12.157	152	9112	0.629	ng/ul	0.00
18) Fluoranthene-d10	19.202	212	20273	0.656	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN071921\
Data File : BN015615.D
Acq On : 19 Jul 2021 19:23
Operator : CG/JU
Sample : SP5623
Misc : SFAM SURROGATE
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SP5623

Manual Integrations
APPROVED

mohammad
7/20/2021 11:38:53 AM

Quant Time: Jul 20 01:11:01 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN062121.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jul 19 13:54:50 2021
Response via : Initial Calibration

