

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN071922\
 Data File : BN020786.D
 Acq On : 19 Jul 2022 22:45
 Operator : CG/JU
 Sample : N3753-07
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 GBHP8

Quant Time: Jul 20 01:05:50 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN062422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 20 01:02:30 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.817	152	2627	0.400	ng/ul	0.00
4) Naphthalene-d8	10.617	136	9109	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.469	164	5029	0.400	ng/ul	# 0.00
13) Phenanthrene-d10	17.221	188	9918	0.400	ng/ul	0.00
17) Chrysene-d12	21.427	240	10805	0.400	ng/ul	0.00
23) Perylene-d12	23.768	264	12106	0.400	ng/ul	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.218	96	10320	3.391	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.223	152	5067	0.339	ng/ul	0.00
18) Fluoranthene-d10	19.257	212	14390	0.404	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN071922\
Data File : BN020786.D
Acq On : 19 Jul 2022 22:45
Operator : CG/JU
Sample : N3753-07
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
GBHP8

Quant Time: Jul 20 01:05:50 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN062422.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jul 20 01:02:30 2022
Response via : Initial Calibration

