

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN071924\
 Data File : BN032833.D
 Acq On : 19 Jul 2024 09:59
 Operator : MA/JU
 Sample : PB162090BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK090

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 07/20/2024
 Supervised By :mohammad ahmed 07/22/2024

Quant Time: Jul 19 10:55:57 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN071224.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 17 02:35:25 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.532	152	4529	0.400	ng/u1	-0.02
4) Naphthalene-d8	10.299	136	14182	0.400	ng/u1	-0.03
9) Acenaphthene-d10	14.179	164	8023	0.400	ng/u1	-0.02
13) Phenanthrene-d10	16.944	188	15815m	0.400	ng/u1	-0.02
17) Chrysene-d12	21.173	240	13790m	0.400	ng/u1	0.00
23) Perylene-d12	23.354	264	14121m	0.400	ng/u1	-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.093	96	26735	4.916	ng/u1	0.00
6) 2-Methylnaphthalene-d10	11.922	152	4964	0.232	ng/u1	-0.01
18) Fluoranthene-d10	18.989	212	10452	0.233	ng/u1	-0.01

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN071924\
Data File : BN032833.D
Acq On : 19 Jul 2024 09:59
Operator : MA/JU
Sample : PB162090BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK090

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/20/2024
Supervised By :mohammad ahmed 07/22/2024

Quant Time: Jul 19 10:55:57 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN071224.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jul 17 02:35:25 2024
Response via : Initial Calibration

