

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN072020\
 Data File : BN011207.D
 Acq On : 20 Jul 2020 13:12
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SICV10

Quant Time: Jul 20 14:05:49 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN072020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 20 12:53:55 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.49	152	1198	0.40	ng/ul	0.00
2) Naphthalene-d8	10.24	136	4135	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.12	164	2403	0.40	ng/ul	0.00
10) Phenanthrene-d10	16.88	188	4990	0.40	ng/ul	0.00
16) Chrysene-d12	21.10	240	3356	0.40	ng/ul	0.00
20) Perylene-d12	23.24	264	2954	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	11.84	152	4286	0.42	ng/ul	0.00
14) Fluoranthene-d10	18.92	212	8937	0.42	ng/ul	0.00
Target Compounds				Ovalue		
3) Naphthalene	10.28	128	5177	0.418	ng/ul	100
5) 2-Methylnaphthalene	11.91	142	3402	0.421	ng/ul	99
7) Acenaphthylene	13.83	152	4876	0.419	ng/ul	99
8) Acenaphthene	14.19	153	3851	0.421	ng/ul	99
9) Fluorene	15.18	166	4338	0.418	ng/ul	98
11) Pentachlorophenol	16.55	266	435	0.408	ng/ul	97
12) Phenanthrene	16.92	178	6861	0.417	ng/ul	99
13) Anthracene	17.01	178	5585	0.412	ng/ul	99
15) Fluoranthene	18.95	202	7561	0.411	ng/ul	99
17) Pyrene	19.32	202	7481	0.418	ng/ul	100
18) Benzo(a)anthracene	21.08	228	5011	0.422	ng/ul	99
19) Chrysene	21.14	228	6771	0.429	ng/ul	100
21) Benzo(b)fluoranthene	22.61	252	5329	0.418	ng/ul	99
22) Benzo(k)fluoranthene	22.65	252	6473	0.442	ng/ul	98
23) Benzo(a)pyrene	23.14	252	4686	0.437	ng/ul	99
24) Indeno(1,2,3-cd)pyrene	25.33	276	5805	0.459	ng/ul#	94
25) Dibenzo(a,h)anthracene	25.34	278	4604	0.483	ng/ul	96
26) Benzo(g,h,i)perylene	25.95	276	5330	0.424	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN072020\
Data File : BN011207.D
Acq On : 20 Jul 2020 13:12
Operator : CG/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SICV10

Quant Time: Jul 20 14:05:49 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN072020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jul 20 12:53:55 2020
Response via : Initial Calibration

