

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN072020\
 Data File : BN011209.D
 Acq On : 20 Jul 2020 15:16
 Operator : CG/JU
 Sample : L3260-08
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 X2M78

Quant Time: Jul 20 15:47:42 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN072020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 20 12:53:55 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.49	152	1089	0.40	ng/ul	0.00
2) Naphthalene-d8	10.24	136	3975	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.12	164	2434	0.40	ng/ul	0.00
10) Phenanthrene-d10	16.88	188	5473	0.40	ng/ul	0.00
16) Chrysene-d12	21.10	240	4379	0.40	ng/ul	0.00
20) Perylene-d12	23.23	264	3924	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	11.84	152	1576	0.16	ng/ul	0.00
14) Fluoranthene-d10	18.92	212	5088	0.22	ng/ul	0.00
Target Compounds				Ovalue		
3) Naphthalene	10.29	128	4769	0.401	ng/ul	99
5) 2-Methylnaphthalene	11.91	142	3425	0.440	ng/ul	100
7) Acenaphthylene	13.83	152	5621	0.476	ng/ul	99
8) Acenaphthene	14.18	153	4699	0.507	ng/ul	99
11) Pentachlorophenol	16.54	266	1172	1.003	ng/ul	98
13) Anthracene	17.01	178	6547	0.440	ng/ul	100
17) Pyrene	19.31	202	15283	0.654	ng/ul	99
19) Chrysene	21.13	228	10204	0.496	ng/ul	99
21) Benzo(b)fluoranthene	22.60	252	10619	0.627	ng/ul	93
22) Benzo(k)fluoranthene	22.64	252	18934	0.974	ng/ul	93
25) Dibenzo(a,h)anthracene	25.33	278	9117	0.720	ng/ul	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN072020\
Data File : BN011209.D
Acq On : 20 Jul 2020 15:16
Operator : CG/JU
Sample : L3260-08
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
X2M78

Quant Time: Jul 20 15:47:42 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN072020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jul 20 12:53:55 2020
Response via : Initial Calibration

