

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN072022\
 Data File : BN020799.D
 Acq On : 20 Jul 2022 10:13
 Operator : CG/JU
 Sample : N3755-01
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 GWOW-BR-03-394-414-071422

Quant Time: Jul 20 10:45:54 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN062622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jul 16 03:21:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.811	152	3110	0.400	ng	0.00
7) Naphthalene-d8	10.615	136	9764	0.400	ng	# 0.00
13) Acenaphthene-d10	14.463	164	6263	0.400	ng	0.00
19) Phenanthrene-d10	17.215	188	14303	0.400	ng	0.00
29) Chrysene-d12	21.422	240	13630	0.400	ng	0.00
35) Perylene-d12	23.773	264	14870	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.406	112	878	0.102	ng	0.00
5) Phenol-d6	7.016	99	644	0.062	ng	0.00
8) Nitrobenzene-d5	8.982	82	1696	0.214	ng	0.01
11) 2-Methylnaphthalene-d10	12.212	152	4342	0.259	ng	0.00
14) 2,4,6-Tribromophenol	15.964	330	558	0.236	ng	-0.01
15) 2-Fluorobiphenyl	13.095	172	5722	0.221	ng	0.00
27) Fluoranthene-d10	19.257	212	14568	0.343	ng	0.00
31) Terphenyl-d14	19.860	244	10020	0.308	ng	0.00
Target Compounds						
24) Pentachlorophenol	16.887	266	178	0.244	ng	Qvalue 87
25) Phenanthrene	17.256	178	1067	0.022	ng	# 90
28) Fluoranthene	19.286	202	1160	0.022	ng	# 92
34) Bis(2-ethylhexyl)phtha...	21.351	149	2557	0.093	ng	# 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN072022\
Data File : BN020799.D
Acq On : 20 Jul 2022 10:13
Operator : CG/JU
Sample : N3755-01
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
GWOW-BR-03-394-414-071422

Quant Time: Jul 20 10:45:54 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN062622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Jul 16 03:21:59 2022
Response via : Initial Calibration

