




Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN072022\
 Data File : BN020796.D
 Acq On : 20 Jul 2022 08:00
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jul 20 10:52:38 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 20 10:52:33 2022
 Response via : Initial Calibration



TIC: BN020796.D\data.ms

(70) Pentachlorophenol (C)




16.875min (0.000) 22241.57 ng

response 386969

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	60.78
264.00	61.60	62.93
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN072022\
Data File : BN020796.D
Acq On : 20 Jul 2022 08:00
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jul 20 10:52:38 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jul 20 10:52:33 2022
Response via : Initial Calibration



TIC: BN020796.D\data.ms

(77) Benzidine

19.474min (0.000) 0.00 ng

response 2303598

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	14.28
183.00	13.20	11.95
0.00	0.00	0.00