


Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN072120\
 Data File : BN011211.D
 Acq On : 21 Jul 2020 08:32
 Operator : CG/JU
 Sample : DFTPP57
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP57

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN072020.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Jul 21 11:43:34 2020



AutoFind: Scans 2440, 2441, 2442; Background Corrected with Scan 2434

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	26.5	119653	PASS
68	69	0.00	2	1.9	3787	PASS
69	198	0.00	100	43.6	196800	PASS
70	69	0.00	2	0.5	975	PASS
127	198	10	80	51.2	231146	PASS
197	198	0.00	2	0.3	1487	PASS
198	198	100	100	100.0	451882	PASS
199	198	5	9	6.8	30701	PASS
275	198	10	60	26.0	117632	PASS
365	198	1	100	4.3	19305	PASS
441	198	0.01	100	9.8	44437	PASS
442	198	50	100	63.8	288104	PASS
443	442	15	24	19.4	55952	PASS

Scan 2434 (17.304 min): BN011211.D
DFTPP57

m/z	Abundance
39.90	383.0
44.00	1053.0
73.00	729.0
147.00	447.0
190.80	393.0
207.10	1159.0
208.00	470.0
220.90	314.0
281.00	687.0

Instrument :
BNA_N
ClientSampleId :
DFTPP57

m/z	Abundance
36.90	847.0
38.00	1627.0
39.00	7414.0
39.90	803.0
44.00	1106.0
49.00	923.0
50.00	25704.0
51.00	91624.0
52.00	4943.0
54.90	669.0
56.00	2825.0
57.00	6564.0
60.90	2310.0
62.00	2449.0
62.90	4904.0
64.00	934.0
64.90	2156.0
66.90	316.0
68.00	3370.0
68.90	145472.0
70.00	522.0
73.00	2185.0
74.00	18504.0
75.00	27920.0
76.00	8833.0
77.00	174144.0
78.00	12869.0
79.00	14349.0
80.00	10100.0
80.90	13564.0
82.00	2925.0
83.00	2185.0
83.80	356.0
85.00	2370.0
86.00	5145.0
87.00	1494.0
87.90	378.0
91.00	2942.0
91.90	4228.0
92.90	24736.0
94.00	1620.0
94.80	458.0
96.00	1060.0
97.20	688.0
98.00	19640.0
99.00	17320.0
100.00	1270.0
101.00	6818.0
103.00	2244.0
104.00	4595.0
105.00	4184.0
105.90	1527.0
107.00	52400.0
108.00	6889.0
109.10	1284.0
110.00	79344.0
111.00	13987.0
112.00	1607.0
112.90	707.0
115.10	331.0
116.00	3327.0
116.90	62984.0
118.00	4554.0
118.90	429.0
120.00	816.0
122.00	3813.0
123.00	5588.0
123.90	2724.0
124.90	1674.0
126.00	760.0
127.00	156224.0
128.00	11900.0
129.00	64848.0
130.00	6447.0
130.90	1035.0
131.90	779.0
132.80	715.0
134.00	2352.0
135.00	5977.0
136.00	2866.0
137.00	4382.0
138.00	1135.0
139.00	459.0
139.90	1250.0
141.00	9778.0
141.90	2861.0
143.00	2138.0
144.10	365.0
144.90	554.0
145.90	1544.0
147.00	5270.0
148.00	14904.0

Instrument :
BNA_N
ClientSampleId :
DFTPP57

148.90	3281.0
150.00	722.0
151.10	1147.0
152.00	809.0
152.90	3320.0
154.00	2464.0
155.00	4813.0
156.00	7293.0
156.80	1382.0
158.00	1704.0
159.00	1238.0
160.00	3799.0
161.00	3857.0
161.90	1129.0
162.90	578.0
164.00	329.0
165.00	4311.0
166.00	2755.0
167.00	22112.0
168.00	15998.0
169.10	1823.0
169.90	581.0
171.00	913.0
171.90	1747.0
172.90	1536.0
174.00	4454.0
175.10	7184.0
175.90	1721.0
177.00	3266.0
178.00	2036.0
179.00	12786.0
180.00	10224.0
181.00	4613.0
181.90	912.0
182.50	393.0
183.90	1192.0
185.00	6052.0
186.00	47200.0
187.00	14542.0
188.00	1883.0
189.00	3487.0
190.00	438.0
190.90	2414.0
192.00	5202.0
193.00	5377.0
194.00	1395.0
195.10	364.0
195.90	7247.0
197.90	286400.0
199.00	20048.0
199.90	1936.0
200.60	394.0
201.20	1191.0
203.00	2551.0
204.00	14168.0
205.00	22808.0
206.00	83280.0
207.00	12630.0
207.90	4045.0
209.00	1509.0
210.00	1890.0
210.90	2637.0
214.80	1445.0
216.00	2218.0
216.90	25240.0
217.90	3561.0
218.90	471.0
220.00	460.0
221.10	9347.0
221.80	3297.0
223.00	5924.0
224.00	53128.0
225.00	13309.0
226.10	2199.0
227.00	24200.0
228.00	4030.0
229.00	4132.0
230.00	918.0
231.00	1695.0
232.90	473.0
234.00	1680.0
235.00	1685.0
235.90	1181.0
236.90	3146.0
237.90	658.0
238.90	975.0
239.90	950.0
240.90	1364.0
242.00	2162.0
242.90	2614.0
244.00	32528.0
245.00	5637.0
246.00	8255.0
246.90	2169.0
248.90	1106.0
252.00	324.0

Instrument :
BNA_N
ClientSampleId :
DFTPP57

252.90	711.0
254.10	1431.0
255.00	165120.0
256.00	24832.0
256.90	1928.0
258.00	11257.0
259.00	1730.0
260.10	391.0
260.80	557.0
265.00	4808.0
266.10	900.0
269.90	382.0
271.00	543.0
272.00	734.0
272.90	4950.0
274.00	13269.0
274.90	66560.0
276.00	9142.0
277.00	6966.0
278.00	1302.0
279.00	449.0
280.80	584.0
282.00	315.0
282.90	887.0
283.90	1118.0
285.00	1038.0
286.10	326.0
293.00	1873.0
294.00	655.0
296.00	22672.0
296.90	3356.0
297.80	384.0
301.00	453.0
302.00	496.0
302.90	2612.0
303.90	903.0
307.80	542.0
313.80	1096.0
315.00	2445.0
315.90	1418.0
320.90	617.0
323.00	6232.0
324.00	1490.0
327.00	1300.0
327.80	937.0
332.00	796.0
332.90	643.0
334.00	4094.0
335.00	1496.0
341.00	524.0
341.80	492.0
345.90	1217.0
346.90	582.0
351.90	2255.0
353.00	1124.0
353.90	2003.0
354.90	591.0
364.90	9571.0
366.00	1608.0
370.90	489.0
372.00	2610.0
372.90	612.0
382.90	831.0
384.90	319.0
401.90	711.0
403.00	1294.0
404.00	696.0
420.90	1234.0
421.90	1222.0
422.90	7120.0
424.00	2558.0
424.90	469.0
441.00	19752.0
442.00	126904.0
442.90	26904.0
443.90	2701.0

Instrument :
BNA_N
ClientSampleId :
DFTPP57

m/z	Abundance
37.10	868.0
38.00	2931.0
39.00	12504.0
40.00	813.0
44.10	1468.0
44.80	402.0
49.20	1395.0
50.00	41088.0
51.00	143040.0
52.00	7378.0
54.90	1342.0
56.00	4558.0
57.00	9530.0
57.80	320.0
58.10	381.0
59.10	309.0
61.00	2752.0
62.00	3430.0
63.00	7541.0
63.90	1402.0
64.90	3960.0
66.00	387.0
68.10	3991.0
68.90	234304.0
70.00	1264.0
71.10	427.0
73.00	4099.0
74.00	30992.0
75.00	43592.0
76.00	16173.0
77.00	279232.0
78.00	20728.0
78.90	19776.0
79.90	15744.0
81.00	21808.0
82.00	5489.0
83.00	3394.0
83.80	721.0
84.90	3627.0
85.90	6525.0
87.00	2771.0
87.90	1011.0
88.90	428.0
91.00	4072.0
92.00	5876.0
93.00	41176.0
94.00	2478.0
95.10	512.0
96.00	1269.0
97.00	928.0
97.90	33480.0
99.00	23800.0
99.90	1433.0
101.00	11209.0
101.90	854.0
103.00	3769.0
104.00	6633.0
105.00	7847.0
106.00	2646.0
107.00	84776.0
108.00	12601.0
109.00	1994.0
110.00	133888.0
111.00	22064.0
111.90	2995.0
113.00	916.0
116.00	5236.0
117.00	103080.0
118.00	7966.0
119.00	830.0
119.90	859.0
120.90	394.0
122.00	5386.0
123.00	9057.0
123.90	3475.0
125.00	3827.0
126.10	1319.0
127.00	275712.0
128.00	19696.0
129.00	115816.0
129.90	9962.0
130.90	1807.0
131.90	968.0
134.00	5002.0
135.00	9152.0
136.00	3918.0
137.00	6226.0
137.90	1301.0
138.80	418.0
139.10	572.0
140.00	1407.0
141.00	15870.0

Instrument :
BNA_N
ClientSampleId :
DFTPP57

141.90	4657.0
143.00	2959.0
144.10	972.0
144.90	903.0
146.00	3228.0
147.00	7826.0
148.00	24472.0
149.00	4667.0
149.90	1259.0
151.00	2222.0
151.80	884.0
152.90	5151.0
154.00	3202.0
155.00	8140.0
156.00	13471.0
157.00	2436.0
158.00	2630.0
159.00	2022.0
160.00	4969.0
161.00	5982.0
161.90	1725.0
162.80	687.0
163.90	768.0
165.00	6792.0
166.00	4931.0
167.00	39184.0
168.00	25296.0
169.00	2888.0
169.90	995.0
170.90	2015.0
171.90	3011.0
173.00	3728.0
174.00	5291.0
175.00	13933.0
175.90	2820.0
177.00	5248.0
177.90	1979.0
178.90	23592.0
180.00	18304.0
180.90	6471.0
181.90	1350.0
183.00	620.0
183.90	1818.0
185.00	12969.0
186.00	87936.0
187.00	26880.0
188.10	3216.0
189.00	6199.0
189.90	1166.0
191.00	2983.0
192.00	8419.0
193.10	10380.0
194.00	1878.0
194.80	878.0
196.00	14292.0
196.90	4462.0
197.90	538240.0
198.90	35496.0
200.00	2484.0
201.30	2156.0
203.00	6233.0
204.00	25656.0
205.00	35640.0
206.00	157568.0
207.00	23320.0
208.00	6470.0
209.00	2866.0
210.10	2549.0
210.90	7086.0
212.10	1362.0
213.00	639.0
215.00	2773.0
216.00	4098.0
216.90	47832.0
218.00	6066.0
218.80	600.0
221.00	18152.0
221.90	5889.0
223.00	10029.0
224.00	94712.0
225.00	24728.0
226.00	2044.0
227.00	46176.0
227.90	6718.0
229.00	8432.0
229.90	1152.0
231.00	3817.0
232.00	466.0
232.90	658.0
233.90	3094.0
235.00	3493.0
236.00	1872.0
237.00	5368.0
238.00	909.0
238.90	1675.0
240.00	1946.0

Instrument :
BNA_N
ClientSampleId :
DFTPP57

241.00	2749.0
242.00	6439.0
243.00	4612.0
244.00	67952.0
245.00	7838.0
245.90	15706.0
246.90	2465.0
248.00	880.0
249.00	2585.0
249.70	480.0
250.80	800.0
252.00	1055.0
252.90	1867.0
255.00	339392.0
256.00	53128.0
257.00	4548.0
258.00	26264.0
259.00	4414.0
260.00	710.0
261.00	724.0
263.00	371.0
264.10	342.0
264.90	9230.0
266.00	1119.0
269.80	637.0
270.90	878.0
271.80	1146.0
273.00	9129.0
274.00	28128.0
275.00	143488.0
276.00	17200.0
277.00	15324.0
277.90	2816.0
278.80	376.0
279.30	312.0
281.10	625.0
283.00	1794.0
283.90	1259.0
285.00	3108.0
285.90	594.0
289.00	554.0
289.90	431.0
290.80	458.0
292.00	674.0
292.90	3191.0
294.10	980.0
295.00	1444.0
295.90	44720.0
297.00	7805.0
301.10	541.0
302.00	950.0
302.90	5479.0
304.00	1624.0
304.90	523.0
308.00	511.0
308.90	384.0
309.90	504.0
312.80	326.0
313.90	2388.0
315.00	5577.0
316.00	2844.0
317.00	715.0
320.90	1030.0
322.00	1054.0
323.00	13657.0
324.00	2352.0
326.00	327.0
327.00	3022.0
327.90	1273.0
332.00	1166.0
332.90	1818.0
334.00	9492.0
335.00	2832.0
336.00	314.0
340.90	1623.0
341.90	690.0
345.90	3138.0
346.90	630.0
351.00	416.0
352.00	4305.0
353.00	2088.0
354.00	5435.0
355.10	778.0
358.80	799.0
364.90	21800.0
365.90	2837.0
369.90	381.0
371.00	910.0
372.00	7185.0
372.90	1686.0
383.00	1646.0
383.90	412.0
389.90	976.0
391.10	399.0
391.90	460.0
400.80	428.0

Instrument :
BNA_N
ClientSampleId :
DFTPP57

401.90	2441.0
402.90	4513.0
403.90	1515.0
404.90	529.0
421.00	2597.0
422.00	2787.0
423.00	19080.0
424.00	5579.0
425.00	744.0
440.00	427.0
441.00	49752.0
442.00	347328.0
443.00	66640.0
444.00	6116.0

Instrument :
BNA_N
ClientSampleId :
DFTPP57

m/z	Abundance
37.00	989.0
38.00	2402.0
39.00	9789.0
40.00	1022.0
43.90	1087.0
44.90	354.0
49.00	1194.0
50.00	35248.0
51.00	124296.0
52.00	6819.0
52.90	608.0
54.90	663.0
55.90	3824.0
57.00	8894.0
61.00	2490.0
62.00	2828.0
63.00	8452.0
64.20	1114.0
65.00	3337.0
67.00	359.0
68.00	4000.0
68.90	210624.0
70.00	1139.0
73.00	2880.0
74.00	27608.0
75.00	37344.0
76.00	12862.0
77.00	250560.0
78.00	17968.0
79.00	19032.0
80.00	12655.0
81.00	18352.0
81.90	4364.0
83.00	4148.0
83.90	407.0
84.90	3237.0
86.00	7713.0
86.90	2363.0
87.90	988.0
90.90	3852.0
92.00	5608.0
93.00	37864.0
94.00	2696.0
94.80	466.0
96.10	1397.0
96.80	786.0
98.00	29232.0
99.00	20304.0
100.00	2341.0
101.00	9974.0
101.80	540.0
102.90	3566.0
104.00	7117.0
105.00	6725.0
106.00	2470.0
107.00	84912.0
108.00	13069.0
109.10	2143.0
109.90	127736.0
111.00	21088.0
112.00	2416.0
113.00	596.0
114.80	476.0
115.90	4825.0
117.00	99440.0
118.00	5681.0
119.00	874.0
120.00	1017.0
122.00	5853.0
123.00	8072.0
124.00	4221.0
125.00	4537.0
127.00	261504.0
128.00	19504.0
129.00	106408.0
130.00	9765.0
131.00	2441.0
131.90	901.0
132.90	670.0
134.00	3931.0
135.00	9206.0
136.00	3546.0
136.90	5087.0
137.90	943.0
139.00	369.0
140.00	1348.0
141.00	15087.0
142.10	4270.0
142.90	3795.0
143.90	959.0
144.80	497.0
146.00	2944.0

Instrument :
BNA_N
ClientSampleId :
DFTPP57

147.00	8881.0
148.00	24664.0
149.00	4283.0
149.90	912.0
151.00	1637.0
151.80	1565.0
153.00	5171.0
153.90	3636.0
155.00	7911.0
156.00	13505.0
157.10	2487.0
157.90	2894.0
159.00	2752.0
160.00	4159.0
161.00	6239.0
162.00	1954.0
163.00	756.0
164.00	595.0
165.00	6638.0
166.00	4675.0
167.00	36664.0
168.00	22032.0
169.00	2374.0
170.00	1382.0
170.90	1540.0
171.90	2935.0
172.90	4050.0
174.00	7459.0
175.00	12913.0
176.00	3374.0
177.00	6343.0
178.10	2402.0
178.90	25720.0
180.00	17736.0
181.00	6032.0
181.90	824.0
182.80	329.0
184.00	2112.0
185.00	11421.0
186.00	87160.0
187.00	25784.0
188.00	3340.0
189.00	6920.0
189.90	1103.0
190.90	2536.0
191.90	8625.0
193.00	9298.0
193.90	2155.0
195.00	813.0
196.00	15415.0
197.90	531008.0
199.00	36560.0
199.90	3309.0
201.40	2247.0
203.00	4779.0
204.00	25896.0
205.00	40544.0
206.00	159680.0
207.00	24680.0
208.00	6825.0
208.90	2580.0
210.10	2687.0
211.00	7279.0
211.90	1530.0
214.10	473.0
215.00	1574.0
216.00	4342.0
216.90	50400.0
218.00	5976.0
219.10	442.0
219.90	772.0
221.00	18440.0
222.00	6518.0
223.00	11041.0
224.00	101592.0
225.00	24264.0
226.00	3290.0
227.00	46048.0
228.00	6131.0
229.00	9233.0
229.90	1164.0
230.90	3454.0
231.90	692.0
233.00	849.0
233.90	2698.0
234.90	3255.0
235.90	2104.0
237.00	3649.0
238.00	665.0
239.00	2272.0
239.90	1535.0
240.90	2543.0
242.00	5386.0
243.00	5744.0
244.00	69136.0
245.00	8982.0

Instrument :
BNA_N
ClientSampleId :
DFTPP57

246.00	16227.0
246.90	3309.0
247.80	764.0
248.90	2516.0
250.00	536.0
251.00	797.0
252.10	474.0
253.00	1642.0
254.10	3211.0
255.00	374720.0
256.00	52896.0
256.90	4487.0
257.90	25904.0
259.00	4737.0
259.90	942.0
260.90	764.0
263.90	604.0
265.00	10863.0
266.00	1658.0
269.90	467.0
270.90	1133.0
271.90	1516.0
273.00	11424.0
274.00	27560.0
275.00	142848.0
276.00	21960.0
277.00	17984.0
277.90	2779.0
278.80	401.0
279.20	402.0
281.00	588.0
281.80	612.0
283.00	1966.0
284.10	1216.0
285.00	3237.0
285.90	602.0
289.00	401.0
290.00	589.0
291.10	448.0
292.00	834.0
292.90	3485.0
293.90	1012.0
296.00	58056.0
297.00	7891.0
298.00	560.0
300.90	769.0
302.00	994.0
303.00	6775.0
304.10	1589.0
307.90	922.0
308.80	742.0
309.90	883.0
312.90	691.0
314.00	3134.0
315.00	6202.0
316.00	3157.0
317.10	589.0
319.90	380.0
321.00	1444.0
321.90	1040.0
323.00	15268.0
324.00	2391.0
324.90	326.0
325.90	324.0
327.00	3358.0
328.00	1425.0
332.10	1146.0
332.90	1722.0
334.00	11074.0
335.00	3069.0
340.90	1825.0
341.80	454.0
345.90	3762.0
346.90	762.0
352.00	5030.0
352.90	3648.0
354.00	4880.0
355.00	1448.0
358.90	687.0
365.00	26544.0
365.90	4157.0
370.10	578.0
371.00	1314.0
372.00	6978.0
373.00	1898.0
382.90	2594.0
384.00	1135.0
390.00	1800.0
391.00	647.0
392.10	575.0
400.80	530.0
402.00	3373.0
403.00	4568.0
403.90	1896.0
404.90	416.0
421.00	4037.0

Instrument :
BNA_N
ClientSampleId :
DFTPP57

422.00	2947.0
422.90	22936.0
423.90	6806.0
425.10	688.0
441.00	63808.0
442.00	390080.0
443.00	74312.0
444.00	7520.0

Instrument :
BNA_N
ClientSampleId :
DFTPP57