


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN072121\
 Data File : BN015626.D
 Acq On : 20 Jul 2021 15:43
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP203

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN062121.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Jul 20 18:52:28 2021



AutoFind: Scans 2470, 2471, 2472; Background Corrected with Scan 2464

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	34.0	66013	PASS
68	69	0.00	2	1.8	1314	PASS
69	198	0.00	100	36.8	71488	PASS
70	69	0.00	2	0.4	256	PASS
127	198	10	80	48.6	94288	PASS
197	198	0.00	2	0.7	1411	PASS
198	198	100	100	100.0	194005	PASS
199	198	5	9	6.9	13337	PASS
275	198	10	60	24.3	47232	PASS
365	198	1	100	2.9	5694	PASS
441	198	0.01	100	11.8	22902	PASS
442	198	50	100	89.5	173613	PASS
443	442	15	24	18.5	32048	PASS

m/z	Abundance
39.90	501.0
44.00	847.0
48.90	309.0
73.00	495.0
83.80	319.0
206.90	646.0

Instrument :
BNA_N
ClientSampleId :
DFTPP203

m/z	Abundance
38.00	741.0
39.00	4222.0
39.90	574.0
43.90	703.0
49.00	753.0
50.00	13833.0
51.10	55584.0
52.00	3032.0
55.00	434.0
56.00	1872.0
57.00	4022.0
61.00	802.0
61.90	787.0
63.00	2495.0
65.00	1121.0
68.00	1146.0
69.00	59888.0
73.00	925.0
74.00	6193.0
75.00	10176.0
76.00	3680.0
77.00	70768.0
78.00	5211.0
79.00	4740.0
80.00	3528.0
81.00	5064.0
82.00	1270.0
82.90	1277.0
83.80	314.0
85.00	863.0
86.00	1419.0
86.90	624.0
91.00	1125.0
92.00	1329.0
93.00	7088.0
93.90	603.0
95.90	454.0
98.00	5621.0
99.00	5374.0
100.00	365.0
101.00	2490.0
103.00	899.0
104.00	1401.0
105.00	1469.0
106.00	516.0
107.00	22248.0
108.00	2998.0
109.00	568.0
110.00	39456.0
111.00	5619.0
112.00	983.0
116.00	1042.0
117.00	15070.0
118.00	1335.0
122.00	1361.0
123.00	2025.0
123.90	954.0
125.00	825.0
127.00	73904.0
128.00	5668.0
129.00	27216.0
130.00	2414.0
131.00	574.0
133.90	845.0
135.00	2145.0
136.00	733.0
137.00	1451.0
139.90	402.0
141.00	3416.0
142.00	1258.0
142.90	854.0
145.90	696.0
146.90	1807.0
147.90	4150.0
148.90	1071.0
151.00	533.0
152.90	1033.0
154.00	862.0
155.00	1948.0
156.00	3348.0
157.10	630.0
157.90	501.0
159.00	483.0
159.90	1038.0
160.90	1623.0
162.00	441.0
164.90	1264.0
166.00	1011.0
167.00	7005.0
168.00	5524.0
169.00	753.0
172.00	692.0

Instrument :
BNA_N
ClientSampleId :
DFTPP203

173.00	800.0
174.00	1380.0
175.00	2662.0
176.00	675.0
176.90	1188.0
177.90	482.0
178.90	5037.0
180.00	4190.0
181.00	1664.0
183.90	403.0
185.00	2579.0
186.00	19288.0
187.00	6144.0
188.00	567.0
189.00	1211.0
190.90	729.0
192.00	1632.0
193.00	1691.0
194.00	390.0
196.00	4334.0
197.00	948.0
197.90	146176.0
199.00	9755.0
200.00	814.0
201.50	663.0
202.90	924.0
204.00	4869.0
205.00	7625.0
206.00	37792.0
207.00	5559.0
207.90	1247.0
208.90	590.0
210.00	460.0
211.00	1354.0
211.90	310.0
214.90	343.0
216.00	899.0
217.00	8160.0
218.00	1196.0
221.00	5427.0
222.90	2049.0
224.00	19344.0
225.00	5234.0
226.00	709.0
226.90	7581.0
227.90	1294.0
229.00	1678.0
231.00	671.0
233.90	545.0
235.00	513.0
236.00	381.0
237.00	789.0
239.00	343.0
241.00	439.0
242.00	1139.0
243.00	1189.0
244.00	14459.0
245.00	2006.0
246.00	2487.0
246.90	665.0
248.90	505.0
255.00	77168.0
256.00	11727.0
256.90	1002.0
258.00	3900.0
258.90	772.0
264.90	1638.0
273.00	2043.0
274.00	5998.0
275.00	34248.0
276.00	4672.0
276.90	2133.0
277.90	514.0
285.10	405.0
293.00	478.0
296.00	8253.0
297.00	1335.0
303.00	1328.0
313.90	413.0
315.00	937.0
316.00	592.0
323.10	2698.0
323.90	508.0
326.90	418.0
333.00	332.0
334.00	1602.0
335.00	417.0
340.90	412.0
345.90	523.0
352.00	810.0
352.90	522.0
354.00	995.0
365.00	3059.0
365.90	547.0
372.00	1542.0
373.00	370.0

Instrument :
BNA_N
ClientSampleId :
DFTPP203

383.00	411.0
401.90	532.0
403.00	866.0
420.90	631.0
422.00	688.0
423.00	5076.0
424.00	2224.0
441.00	12531.0
442.00	95560.0
442.90	17184.0
444.00	1615.0

Instrument :
BNA_N
ClientSampleId :
DFTPP203

m/z	Abundance
37.00	339.0
38.00	986.0
39.00	5546.0
39.90	533.0
43.90	797.0
49.00	944.0
50.10	19632.0
51.00	78400.0
52.00	4198.0
54.90	575.0
56.00	2479.0
57.00	5420.0
61.00	1061.0
62.00	1056.0
63.00	3293.0
64.00	422.0
65.00	1877.0
68.00	1434.0
69.00	82800.0
70.00	457.0
73.00	966.0
74.00	8305.0
75.00	13568.0
76.00	4525.0
77.00	96248.0
78.00	7232.0
79.00	6229.0
80.00	4873.0
81.00	6653.0
81.90	1527.0
83.00	1594.0
83.90	353.0
85.00	1275.0
86.00	1989.0
87.00	950.0
88.00	350.0
91.00	1371.0
92.00	1759.0
93.00	10303.0
94.00	792.0
96.10	504.0
98.00	7787.0
99.00	7288.0
100.00	624.0
101.00	3672.0
103.00	1387.0
103.90	2583.0
104.90	2177.0
106.00	761.0
107.00	29800.0
108.00	4954.0
109.00	980.0
110.00	55704.0
111.00	8645.0
112.00	961.0
113.00	323.0
115.90	1581.0
117.00	22256.0
118.00	1736.0
119.90	348.0
122.00	2139.0
123.00	2973.0
123.90	1330.0
125.00	1381.0
127.00	111696.0
128.00	8265.0
129.00	38824.0
130.00	3391.0
131.00	618.0
132.00	326.0
134.00	1266.0
135.00	3779.0
136.00	1362.0
137.00	1830.0
137.80	396.0
139.90	410.0
141.00	4988.0
142.00	1818.0
142.90	1215.0
144.90	331.0
146.00	947.0
147.00	2747.0
148.00	5745.0
149.00	1518.0
149.90	364.0
151.20	468.0
151.80	483.0
153.00	1808.0
153.90	1245.0
155.00	2846.0
156.00	4289.0
157.00	915.0

Instrument :
BNA_N
ClientSampleId :
DFTPP203

157.90	1055.0
158.90	753.0
160.00	1724.0
161.00	2731.0
162.00	914.0
165.00	1803.0
165.90	1650.0
167.00	10592.0
168.00	6630.0
169.00	1085.0
169.90	326.0
171.00	455.0
172.00	1083.0
172.90	1070.0
174.00	2312.0
175.00	4535.0
176.00	1203.0
176.90	1730.0
178.00	726.0
179.00	7775.0
180.00	6091.0
181.00	2821.0
182.00	343.0
183.90	581.0
185.00	3803.0
186.00	31448.0
187.00	8868.0
188.00	1105.0
189.00	1517.0
189.90	318.0
190.90	872.0
192.00	2719.0
193.00	2915.0
194.00	636.0
196.00	6391.0
196.90	1513.0
198.00	222976.0
198.90	15630.0
200.00	1055.0
201.50	1048.0
203.00	1380.0
204.00	7755.0
205.00	13287.0
206.00	59080.0
207.00	8629.0
207.90	1889.0
208.90	725.0
211.00	2347.0
211.70	538.0
215.00	600.0
216.00	1279.0
216.90	14253.0
218.00	1963.0
221.00	9698.0
221.90	1763.0
223.00	3425.0
224.00	30832.0
225.00	8165.0
226.00	1135.0
227.00	12089.0
228.00	1869.0
229.00	2837.0
231.00	902.0
233.80	689.0
234.90	938.0
236.00	609.0
237.00	1361.0
238.90	530.0
240.00	510.0
240.90	822.0
242.10	1643.0
243.00	1716.0
244.00	24096.0
245.00	3699.0
246.00	4346.0
247.00	919.0
249.00	751.0
252.00	317.0
253.00	524.0
254.10	1107.0
255.00	128264.0
256.00	17776.0
257.00	1666.0
258.00	6100.0
259.00	1097.0
265.00	2468.0
266.00	376.0
271.90	326.0
273.00	3375.0
274.00	10371.0
275.00	53944.0
276.00	7735.0
277.00	4566.0
278.00	909.0
283.00	633.0
284.00	465.0

Instrument :
BNA_N
ClientSampleId :
DFTPP203

285.00	795.0
293.00	961.0
296.00	13801.0
297.00	1759.0
303.10	1571.0
304.00	583.0
314.00	626.0
314.90	1467.0
315.90	749.0
321.00	450.0
323.00	4414.0
324.00	983.0
326.90	850.0
327.80	475.0
333.00	479.0
334.00	3031.0
335.00	644.0
341.10	569.0
345.90	1079.0
352.00	1177.0
352.90	1197.0
354.00	1547.0
364.90	6315.0
366.00	871.0
372.00	2691.0
372.90	711.0
382.90	728.0
402.00	1180.0
403.00	1829.0
403.90	539.0
421.00	1597.0
422.00	1505.0
423.10	10049.0
424.10	2625.0
424.90	331.0
441.00	26128.0
442.00	207424.0
443.00	32776.0
444.00	3456.0

Instrument :
BNA_N
ClientSampleId :
DFTPP203

m/z	Abundance
36.90	318.0
38.00	666.0
39.10	4761.0
39.90	671.0
41.00	438.0
44.00	799.0
49.00	669.0
50.00	16238.0
51.00	64056.0
52.00	3431.0
55.00	446.0
56.00	2087.0
57.00	4692.0
61.00	844.0
62.00	814.0
63.00	2607.0
64.10	379.0
65.00	1388.0
68.00	1363.0
69.00	71776.0
70.00	312.0
73.00	819.0
74.00	7282.0
75.00	11315.0
76.00	4108.0
77.00	84272.0
78.00	5747.0
79.00	5637.0
79.90	3740.0
81.00	5878.0
82.00	1411.0
83.00	1519.0
83.90	366.0
85.00	1251.0
86.00	1840.0
87.00	728.0
88.00	325.0
90.90	1387.0
92.00	1568.0
93.00	9325.0
93.90	737.0
96.00	557.0
98.00	6877.0
99.00	5959.0
99.90	506.0
101.00	3368.0
102.90	1156.0
104.00	1956.0
104.90	1809.0
106.00	635.0
107.00	26648.0
108.00	4248.0
109.00	776.0
110.00	50136.0
111.00	7268.0
112.00	925.0
116.00	1463.0
117.00	19096.0
118.00	1367.0
120.00	453.0
121.90	1698.0
123.00	2374.0
124.00	1125.0
124.90	1328.0
126.10	739.0
127.00	97264.0
128.00	7378.0
129.00	34656.0
130.00	2658.0
131.00	615.0
133.90	1000.0
135.00	3328.0
136.00	1201.0
137.00	1408.0
140.10	439.0
141.00	5001.0
142.00	1632.0
142.90	1149.0
144.00	369.0
145.90	881.0
147.00	2318.0
147.90	5210.0
149.00	1360.0
150.00	425.0
151.00	548.0
151.80	316.0
152.90	1519.0
154.00	1247.0
155.00	2489.0
156.00	4087.0
157.00	732.0
158.00	891.0

Instrument :
BNA_N
ClientSampleId :
DFTPP203

158.90	647.0
160.00	1599.0
161.00	2220.0
161.90	753.0
164.10	323.0
164.90	1937.0
166.00	1428.0
167.00	9692.0
168.00	5919.0
169.00	868.0
169.90	386.0
170.90	418.0
171.90	861.0
173.00	1200.0
173.90	1978.0
175.00	3801.0
176.00	1197.0
177.00	2355.0
178.00	633.0
178.90	7096.0
180.00	5153.0
181.00	2391.0
182.10	465.0
183.90	471.0
185.00	3573.0
186.00	28272.0
187.00	7699.0
188.00	888.0
189.00	1662.0
189.80	353.0
190.90	842.0
192.00	2173.0
193.00	2775.0
193.90	548.0
195.10	339.0
196.00	6186.0
197.10	1773.0
198.00	212864.0
199.00	14625.0
200.00	1101.0
200.80	481.0
201.60	915.0
203.00	1558.0
204.00	7152.0
205.00	13553.0
206.00	53416.0
207.10	7661.0
208.00	1734.0
209.00	665.0
210.10	921.0
210.90	2109.0
215.00	543.0
216.00	1072.0
217.00	14443.0
217.90	1818.0
221.00	9146.0
221.70	2156.0
223.00	2661.0
224.00	30096.0
225.00	7275.0
226.00	904.0
227.00	12105.0
228.00	1478.0
229.00	2295.0
229.90	338.0
231.00	1204.0
233.90	882.0
235.00	986.0
235.90	450.0
237.00	1283.0
239.00	349.0
239.80	372.0
240.90	690.0
242.00	1878.0
242.90	1498.0
244.00	24752.0
245.00	2924.0
246.00	4611.0
247.00	748.0
248.90	902.0
253.00	656.0
255.00	121248.0
256.00	17384.0
257.00	1378.0
257.90	5744.0
259.00	1289.0
265.00	2549.0
266.00	381.0
271.90	369.0
272.90	3288.0
274.00	10521.0
275.00	53504.0
276.00	8508.0
277.00	4568.0
278.00	630.0
282.90	492.0

Instrument :
BNA_N
ClientSampleId :
DFTPP203

284.00	490.0
285.10	784.0
292.00	330.0
293.00	1106.0
295.10	328.0
296.00	13743.0
297.00	2088.0
303.00	1555.0
303.90	329.0
314.00	724.0
315.00	1687.0
315.90	981.0
321.00	473.0
323.00	5323.0
324.00	874.0
326.90	962.0
328.10	494.0
332.00	334.0
333.00	572.0
334.00	3203.0
335.00	881.0
341.00	491.0
346.00	991.0
351.90	1574.0
352.90	1123.0
354.00	1432.0
365.00	7707.0
365.90	825.0
371.00	554.0
372.00	2956.0
372.90	505.0
383.00	986.0
391.00	308.0
402.00	1056.0
403.00	1933.0
404.00	726.0
421.00	1580.0
422.00	1681.0
423.00	9185.0
424.00	3150.0
425.00	449.0
441.00	30048.0
442.00	217856.0
443.00	46184.0
444.00	3481.0

Instrument :
BNA_N
ClientSampleId :
DFTPP203