

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN072122\
 Data File : BN020808.D
 Acq On : 21 Jul 2022 13:50
 Operator : CG/JU
 Sample : PB146317BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK317

Quant Time: Jul 21 14:18:15 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN062422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 20 01:02:30 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.817	152	2795	0.400	ng/ul	0.00
4) Naphthalene-d8	10.612	136	9425	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.470	164	5946	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.217	188	12354	0.400	ng/ul	0.00
17) Chrysene-d12	21.438	240	9496	0.400	ng/ul	0.00
23) Perylene-d12	23.776	264	14368	0.400	ng/ul	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.218	96	18568	5.734	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.223	152	4789	0.310	ng/ul	0.00
18) Fluoranthene-d10	19.258	212	12174	0.389	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN072122\
Data File : BN020808.D
Acq On : 21 Jul 2022 13:50
Operator : CG/JU
Sample : PB146317BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK317

Quant Time: Jul 21 14:18:15 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN062422.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jul 20 01:02:30 2022
Response via : Initial Calibration

