

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN072219\
 Data File : BN006941.D
 Acq On : 22 Jul 2019 09:31
 Operator : HP/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD0.441

Quant Time: Jul 22 12:02:03 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN062719.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 19 23:29:55 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.62	152	2193	0.40	ng/ul	0.00
2) Naphthalene-d8	10.37	136	8652	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.24	164	4495	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.01	188	9085	0.40	ng/ul	0.00
16) Chrysene-d12	21.24	240	6604	0.40	ng/ul	0.00
20) Perylene-d12	23.47	264	7508	0.40	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Methylnaphthalene-d10	11.99	152	5324	0.40	ng/ul	0.00
14) Fluoranthene-d10	19.05	212	10123	0.38	ng/ul	0.00

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
3) Naphthalene	10.42	128	10535	0.424	ng/ul	97
5) 2-Methylnaphthalene	12.06	142	6423	0.383	ng/ul	99
7) Acenaphthylene	13.97	152	8810	0.380	ng/ul	99
8) Acenaphthene	14.31	153	7400	0.414	ng/ul	99
9) Fluorene	15.31	166	7461	0.381	ng/ul	98
11) Pentachlorophenol	16.74	266	428	0.316	ng/ul	100
12) Phenanthrene	17.06	178	11850	0.405	ng/ul	99
13) Anthracene	17.16	178	8798	0.348	ng/ul	97
15) Fluoranthene	19.09	202	13349	0.369	ng/ul	100
17) Pyrene	19.45	202	14198	0.470	ng/ul	99
18) Benzo(a)anthracene	21.23	228	9605	0.354	ng/ul	99
19) Chrysene	21.28	228	14314	0.473	ng/ul	99
21) Benzo(b)fluoranthene	22.80	252	11987	0.390	ng/ul	98
22) Benzo(k)fluoranthene	22.85	252	13690	0.425	ng/ul	99
23) Benzo(a)pyrene	23.38	252	11832	0.393	ng/ul	98
24) Indeno(1,2,3-cd)pyrene	25.70	276	13070	0.428	ng/ul#	95
25) Dibenzo(a,h)anthracene	25.70	278	9450	0.390	ng/ul	97
26) Benzo(g,h,i)perylene	26.38	276	12345	0.484	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN072219\
 Data File : BN006941.D
 Acq On : 22 Jul 2019 09:31
 Operator : HP/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 Client Sampled :
 SSTD0.441

Quant Time: Jul 22 12:02:03 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN062719.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 19 23:29:55 2019
 Response via : Initial Calibration

