

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN072221\
 Data File : BN015641.D
 Acq On : 21 Jul 2021 12:19
 Operator : CG/JU
 Sample : SSTDICC0.1
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1

Quant Time: Jul 21 14:14:46 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN072221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 21 14:10:52 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.744	152	12080	0.400	ng	0.00
7) Naphthalene-d8	10.521	136	53976	0.400	ng	0.00
13) Acenaphthene-d10	14.370	164	29262	0.400	ng	0.00
19) Phenanthrene-d10	17.127	188	56942	0.400	ng	0.00
29) Chrysene-d12	21.323	240	53623	0.400	ng	0.00
35) Perylene-d12	23.638	264	52571	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.346	112	3848	0.111	ng	0.00
5) Phenol-d6	6.914	99	4558	0.111	ng	0.00
8) Nitrobenzene-d5	8.886	82	3857	0.099	ng	0.00
11) 2-Methylnaphthalene-d10	12.118	152	9711	0.109	ng	0.00
14) 2,4,6-Tribromophenol	15.867	330	1270	0.120	ng	0.00
15) 2-Fluorobiphenyl	13.000	172	11899	0.101	ng	0.00
27) Fluoranthene-d10	19.162	212	19169	0.111	ng	0.00
31) Terphenyl-d14	19.768	244	13222	0.104	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.290	88	393	0.067	ng	# 52
3) n-Nitrosodimethylamine	3.650	42	1007	0.101	ng	98
6) bis(2-Chloroethyl)ether	7.172	93	3934	0.107	ng	99
9) Naphthalene	10.571	128	17452	0.111	ng	99
10) Hexachlorobutadiene	10.863	225	3026	0.103	ng	# 100
12) 2-Methylnaphthalene	12.189	142	10477	0.105	ng	99
16) Acenaphthylene	14.091	152	12604	0.093	ng	99
17) Acenaphthene	14.434	154	9235	0.102	ng	99
18) Fluorene	15.423	166	11308	0.104	ng	99
20) 4,6-Dinitro-2-methylph...	15.512	198	333	0.145	ng	# 67
21) 4-Bromophenyl-phenylether	16.324	248	3582	0.099	ng	99
22) Hexachlorobenzene	16.433	284	3876	0.095	ng	99
23) Atrazine	16.591	200	2619	0.100	ng	99
24) Pentachlorophenol	16.786	266	598	0.069	ng	97
25) Phenanthrene	17.163	178	18507	0.101	ng	99
26) Anthracene	17.249	178	15353	0.097	ng	100
28) Fluoranthene	19.192	202	20808	0.104	ng	100
30) Pyrene	19.554	202	21646	0.105	ng	100
32) Benzo(a)anthracene	21.312	228	19511	0.104	ng	97
33) Chrysene	21.365	228	20099	0.101	ng	97
34) Bis(2-ethylhexyl)phtha...	21.249	149	11697	0.146	ng	100
36) Indeno(1,2,3-cd)pyrene	26.010	276	23021	0.123	ng	96
37) Benzo(b)fluoranthene	22.985	252	21382	0.104	ng	95
38) Benzo(k)fluoranthene	22.985	252	21382	0.103	ng	95
39) Benzo(a)pyrene	23.539	252	18029	0.102	ng	# 86
40) Dibenzo(a,h)anthracene	26.028	278	18685	0.127	ng	95
41) Benzo(g,h,i)perylene	26.736	276	19124	0.120	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN072221\
Data File : BN015641.D
Acq On : 21 Jul 2021 12:19
Operator : CG/JU
Sample : SSTDIC0.1
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1

Quant Time: Jul 21 14:14:46 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN072221.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jul 21 14:10:52 2021
Response via : Initial Calibration




Abundance Scan 506 (7.744 min): BN015643.D\data.ms (-501) (#1)





1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.744 min Scan# 506
Delta R.T. -0.000 min
Lab File: BN015641.D
Acq: 21 Jul 2021 12:19


Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1



Tgt Ion:152 Resp: 12080
Ion Ratio Lower Upper
152 100
150 153.1 123.4 185.2
115 59.1 47.0 70.6







Abundance Scan 22 (3.281 min): BN015643.D\data.ms (-18) (#2)



1,4-Dioxane
Concen: 0.067 ng
RT: 3.290 min Scan# 23
Delta R.T. 0.009 min
Lab File: BN015641.D
Acq: 21 Jul 2021 12:19

Tgt Ion: 88 Resp: 393
Ion Ratio Lower Upper
88 100
43 276.6 144.0 216.0#
58 98.7 72.6 108.8









n-Nitrosodimethylamine
 Concen: 0.101 ng
 RT: 3.650 min Scan# 62
 Delta R.T. -0.000 min
 Lab File: BN015641.D
 Acq: 21 Jul 2021 12:19

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1

Tgt Ion: 42 Resp: 1007




Ion	Ratio	Lower	Upper
42	100		
74	157.6	123.8	185.8
44	8.8	5.9	8.9




2-Fluorophenol
 Concen: 0.111 ng
 RT: 5.346 min Scan# 246
 Delta R.T. -0.000 min
 Lab File: BN015641.D
 Acq: 21 Jul 2021 12:19

Tgt Ion: 112 Resp: 3848

Ion	Ratio	Lower	Upper
112	100		
64	55.3	44.2	66.4
63	28.2	22.1	33.1









Phenol-d6
 Concen: 0.111 ng
 RT: 6.914 min Scan# 416
 Delta R.T. -0.000 min
 Lab File: BN015641.D
 Acq: 21 Jul 2021 12:19




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1

Tgt Ion:	99	Resp:	4558
Ion	Ratio	Lower	Upper
99	100		
42	15.9	13.0	19.6
71	29.9	24.4	36.6




bis(2-Chloroethyl)ether
 Concen: 0.107 ng
 RT: 7.172 min Scan# 444
 Delta R.T. -0.000 min
 Lab File: BN015641.D
 Acq: 21 Jul 2021 12:19

Tgt Ion:	93	Resp:	3934
Ion	Ratio	Lower	Upper
93	100		
63	73.3	59.1	88.7
95	33.8	26.3	39.5






Abundance Scan 742 (10.521 min): BN015643.D\data.ms (-737) (-)




Naphthalene-d8
Concen: 0.400 ng
RT: 10.521 min Scan# 742
Delta R.T. -0.000 min
Lab File: BN015641.D
Acq: 21 Jul 2021 12:19

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1

Tgt Ion:	136	Resp:	53976
Ion	Ratio	Lower	Upper
136	100		
137	11.0	8.8	13.2
54	6.9	5.7	8.5
68	6.0	4.8	7.2





Abundance Scan 613 (8.886 min): BN015643.D\data.ms (-610) (-)





Nitrobenzene-d5
Concen: 0.099 ng
RT: 8.886 min Scan# 613
Delta R.T. -0.000 min
Lab File: BN015641.D
Acq: 21 Jul 2021 12:19

Tgt Ion:	82	Resp:	3857
Ion	Ratio	Lower	Upper
82	100		
128	40.5	32.2	48.4
54	49.9	39.0	58.6



Abundance Scan 613 (8.886 min): BN015641.D\data.ms (-597) (-)









Scan 746 (10.571 min): BN015643.D\data.ms (-746)
 Naphthalene
 Concen: 0.111 ng
 RT: 10.571 min Scan# 746
 Delta R.T. -0.000 min
 Lab File: BN015641.D
 Acq: 21 Jul 2021 12:19




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1


Tgt Ion:	128	Resp:	17452
Ion	Ratio	Lower	Upper
128	100		
129	11.3	8.7	13.1
127	13.2	10.4	15.6



Scan 769 (10.863 min): BN015643.D\data.ms (-769)
 Hexachlorobutadiene
 Concen: 0.103 ng
 RT: 10.863 min Scan# 769
 Delta R.T. -0.000 min
 Lab File: BN015641.D
 Acq: 21 Jul 2021 12:19

Tgt Ion:	225	Resp:	3026
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	63.7	51.1	76.7









2-Methylnaphthalene-d10
Concen: 0.109 ng
RT: 12.118 min Scan# 957
Delta R.T. -0.000 min
Lab File: BN015641.D
Acq: 21 Jul 2021 12:19




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1

Tgt Ion:152 Resp: 9711
Ion Ratio Lower Upper
152 100
151 20.7 16.2 24.4



2-Methylnaphthalene
Concen: 0.105 ng
RT: 12.189 min Scan# 973
Delta R.T. -0.000 min
Lab File: BN015641.D
Acq: 21 Jul 2021 12:19


Tgt Ion:142 Resp: 10477
Ion Ratio Lower Upper
142 100
141 89.5 71.1 106.7
115 34.1 26.8 40.2






Abundance Scan 1191 (14.370 min): BN015643.D\data.ms (-1112) (-)

Acenaphthene-d10
Concen: 0.400 ng
RT: 14.370 min Scan# 1191
Delta R.T. -0.000 min
Lab File: BN015641.D
Acq: 21 Jul 2021 12:19

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1







Tgt Ion:	164	Resp:	29262
Ion	Ratio	Lower	Upper
164	100		
162	102.5	82.4	123.6
160	47.2	37.3	55.9




Abundance Scan 1309 (15.867 min): BN015643.D\data.ms (-1305) (-)

2,4,6-Tribromophenol
Concen: 0.120 ng
RT: 15.867 min Scan# 1309
Delta R.T. -0.000 min
Lab File: BN015641.D
Acq: 21 Jul 2021 12:19

Tgt Ion:	330	Resp:	1270
Ion	Ratio	Lower	Upper
330	100		
332	0.0	0.0	0.0
141	37.3	29.0	43.6





Abundance Scan 1083 (13.000 min): BN015643.D\data.ms (-1075) (-)




2-Fluorobiphenyl
Concen: 0.101 ng
RT: 13.000 min Scan# 1083
Delta R.T. -0.000 min
Lab File: BN015641.D
Acq: 21 Jul 2021 12:19

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:	172	Resp:	11899
Ion	Ratio	Lower	Upper
172	100		
171	34.2	27.2	40.8
170	22.7	17.8	26.6



Abundance Scan 1083 (13.000 min): BN015641.D\data.ms (-1075) (-)





Abundance Scan 1169 (14.091 min): BN015643.D\data.ms (-1153) (-)




Acenaphthylene
Concen: 0.093 ng
RT: 14.091 min Scan# 1169
Delta R.T. -0.000 min
Lab File: BN015641.D
Acq: 21 Jul 2021 12:19

Tgt Ion:	152	Resp:	12604
Ion	Ratio	Lower	Upper
152	100		
151	18.9	15.6	23.4
153	12.9	10.3	15.5




Abundance Scan 1169 (14.091 min): BN015641.D\data.ms (-1153) (-)






Abundance Scan 1196 (14.434 min): BN015643.D\data.ms (-1197) (-)

Acenaphthene
Concen: 0.102 ng
RT: 14.434 min Scan# 1196
Delta R.T. -0.000 min
Lab File: BN015641.D
Acq: 21 Jul 2021 12:19

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Tgt Ion:	154	Resp:	9235
Ion	Ratio	Lower	Upper
154	100		
153	111.5	90.2	135.2
152	54.8	43.7	65.5






Abundance Scan 1274 (15.423 min): BN015643.D\data.ms (-1275) (-)

Fluorene
Concen: 0.104 ng
RT: 15.423 min Scan# 1274
Delta R.T. -0.000 min
Lab File: BN015641.D
Acq: 21 Jul 2021 12:19




Tgt Ion:	166	Resp:	11308
Ion	Ratio	Lower	Upper
166	100		
165	96.9	78.6	118.0
167	13.5	10.7	16.1





Abundance Scan 1412 (17.127 min): BN015643.D\data.ms (-1416) (-)

Phenanthrene-d10
Concen: 0.400 ng
RT: 17.127 min Scan# 1412
Delta R.T. -0.000 min
Lab File: BN015641.D
Acq: 21 Jul 2021 12:19


Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1



Tgt Ion:	188	Resp:	56942
Ion	Ratio	Lower	Upper
188	100		
94	0.0	0.0	0.0
80	6.7	5.3	7.9





Abundance Scan 1412 (17.127 min): BN015641.D\data.ms (-1396) (-)




Abundance Scan 1281 (15.512 min): BN015643.D\data.ms (-1276) (-)

4,6-Dinitro-2-methylphenol
Concen: 0.145 ng
RT: 15.512 min Scan# 1281
Delta R.T. -0.000 min
Lab File: BN015641.D
Acq: 21 Jul 2021 12:19


Tgt Ion:	198	Resp:	333
Ion	Ratio	Lower	Upper
198	100		
182	31.8	14.0	21.0#
77	0.0	0.0	0.0



Abundance Scan 1281 (15.512 min): BN015641.D\data.ms (-1273) (-)






Abundance Scan 1346 (16.324 min): BN015643.D\data.ms (-1321) (-)




4-Bromophenyl-phenylether
Concen: 0.099 ng
RT: 16.324 min Scan# 1346
Delta R.T. -0.000 min
Lab File: BN015641.D
Acq: 21 Jul 2021 12:19

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1

Tgt Ion:	248	Resp:	3582
Ion	Ratio	Lower	Upper
248	100		
250	102.5	82.1	123.1
141	37.0	28.5	42.7






Abundance Scan 1355 (16.433 min): BN015643.D\data.ms (-1320) (-)




Hexachlorobenzene
Concen: 0.095 ng
RT: 16.433 min Scan# 1355
Delta R.T. -0.000 min
Lab File: BN015641.D
Acq: 21 Jul 2021 12:19

Tgt Ion:	284	Resp:	3876
Ion	Ratio	Lower	Upper
284	100		
142	34.5	27.1	40.7
249	31.1	24.7	37.1





Abundance Scan 1368 (16.592 min): BN015643.D\data.ms (-1327) (-)




Atrazine
Concen: 0.100 ng
RT: 16.591 min Scan# 1368
Delta R.T. -0.000 min
Lab File: BN015641.D
Acq: 21 Jul 2021 12:19

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:	200	Resp:	2619
Ion	Ratio	Lower	Upper
200	100		
173	27.4	21.2	31.8
215	48.8	38.8	58.2



Abundance Scan 1368 (16.591 min): BN015641.D\data.ms (-1327) (-)





Abundance Scan 1384 (16.786 min): BN015643.D\data.ms (-1329) (-)




Pentachlorophenol
Concen: 0.069 ng
RT: 16.786 min Scan# 1384
Delta R.T. -0.000 min
Lab File: BN015641.D
Acq: 21 Jul 2021 12:19


Tgt Ion:	266	Resp:	598
Ion	Ratio	Lower	Upper
266	100		
264	61.4	50.2	75.4
268	67.6	51.6	77.4



Abundance Scan 1384 (16.786 min): BN015641.D\data.ms (-1376) (-)





Abundance Scan 1415 (17.164 min): BN015643.D\data.ms (-1425) (-)




Phenanthrene
Concen: 0.101 ng
RT: 17.163 min Scan# 1415
Delta R.T. -0.000 min
Lab File: BN015641.D
Acq: 21 Jul 2021 12:19

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:178 Resp: 18507
Ion Ratio Lower Upper
178 100
176 18.8 15.0 22.6
179 16.0 12.3 18.5



Abundance Scan 1415 (17.163 min): BN015641.D\data.ms (-1390) (-)





Abundance Scan 1423 (17.261 min): BN015643.D\data.ms (-1425) (-)




Anthracene
Concen: 0.097 ng
RT: 17.249 min Scan# 1422
Delta R.T. -0.012 min
Lab File: BN015641.D
Acq: 21 Jul 2021 12:19


Tgt Ion:178 Resp: 15353
Ion Ratio Lower Upper
178 100
176 18.1 14.5 21.7
179 15.3 12.2 18.4



Abundance Scan 1422 (17.249 min): BN015641.D\data.ms (-1415) (-)





Abundance Scan 1716 (19.162 min): BN015643.D\data.ms (-1#27) (-)




Fluoranthene-d10
Concen: 0.111 ng
RT: 19.162 min Scan# 1716
Delta R.T. -0.000 min
Lab File: BN015641.D
Acq: 21 Jul 2021 12:19

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:212 Resp: 19169
Ion Ratio Lower Upper
212 100
106 13.9 11.3 16.9
104 7.8 6.1 9.1



Abundance Scan 1716 (19.162 min): BN015641.D\data.ms (-1696) (-)





Abundance Scan 1722 (19.192 min): BN015643.D\data.ms (-1#28) (-)




Fluoranthene
Concen: 0.104 ng
RT: 19.192 min Scan# 1722
Delta R.T. -0.000 min
Lab File: BN015641.D
Acq: 21 Jul 2021 12:19

Tgt Ion:202 Resp: 20808
Ion Ratio Lower Upper
202 100
101 11.4 9.3 13.9
203 17.3 13.8 20.6




Abundance Scan 1722 (19.192 min): BN015641.D\data.ms (-1682) (-)






Abundance Scan 2010 (21.323 min): BN015643.D\data.ms (-1736) (-)

Chrysene-d12
Concen: 0.400 ng
RT: 21.323 min Scan# 2010
Delta R.T. -0.000 min
Lab File: BN015641.D
Acq: 21 Jul 2021 12:19

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Tgt Ion:	240	Resp:	53623
Ion	Ratio	Lower	Upper
240	100		
120	10.6	7.1	10.7
236	26.1	20.7	31.1






Abundance Scan 1795 (19.554 min): BN015643.D\data.ms (-1736) (-)

Pyrene
Concen: 0.105 ng
RT: 19.554 min Scan# 1795
Delta R.T. -0.000 min
Lab File: BN015641.D
Acq: 21 Jul 2021 12:19




Tgt Ion:	202	Resp:	21646
Ion	Ratio	Lower	Upper
202	100		
200	20.6	16.4	24.6
203	17.5	13.9	20.9






Abundance Scan 1838 (19.768 min): BN015643.D\data.ms (-1838) (-)

Terphenyl-d14
Concen: 0.104 ng
RT: 19.768 min Scan# 1838
Delta R.T. -0.000 min
Lab File: BN015641.D
Acq: 21 Jul 2021 12:19

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1




Tgt Ion	Ratio	Lower	Upper
244	100		
212	8.1	6.1	9.1
122	13.1	9.4	14.2






Abundance Scan 2008 (21.302 min): BN015643.D\data.ms (-2008) (-)


Benzo(a)anthracene
Concen: 0.104 ng
RT: 21.312 min Scan# 2009
Delta R.T. 0.010 min
Lab File: BN015641.D
Acq: 21 Jul 2021 12:19



Tgt Ion	Ratio	Lower	Upper
228	100		
226	26.1	22.1	33.1
229	20.3	15.3	22.9






Abundance Scan 2013 (21.355 min): BN015643.D\data.ms (-2034) (-)




Chrysene
Concen: 0.101 ng
RT: 21.365 min Scan# 2014
Delta R.T. 0.011 min
Lab File: BN015641.D
Acq: 21 Jul 2021 12:19

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1

Tgt Ion:	228	Resp:	20099
Ion	Ratio	Lower	Upper
228	100		
226	28.9	24.3	36.5
229	20.4	15.5	23.3






Abundance Scan 2003 (21.249 min): BN015643.D\data.ms (-1994) (-)



Bis(2-ethylhexyl)phthalate
Concen: 0.146 ng
RT: 21.249 min Scan# 2003
Delta R.T. -0.000 min
Lab File: BN015641.D
Acq: 21 Jul 2021 12:19



Tgt Ion:	149	Resp:	11697
Ion	Ratio	Lower	Upper
149	100		
167	28.5	23.0	34.4
279	4.0	2.9	4.3





Abundance Scan 2471 (23.635 min): BN015643.D\data.ms (-2435) (-)

Perylene-d12
Concen: 0.400 ng
RT: 23.638 min Scan# 2472
Delta R.T. 0.003 min
Lab File: BN015641.D
Acq: 21 Jul 2021 12:19

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1





Tgt Ion:	264	Resp:	52571
Ion	Ratio	Lower	Upper
264	100		
260	24.6	19.7	29.5
265	24.9	20.0	30.0





Abundance Scan 3164 (26.007 min): BN015643.D\data.ms (-3146) (-)


Indeno(1,2,3-cd)pyrene
Concen: 0.123 ng
RT: 26.010 min Scan# 3165
Delta R.T. 0.003 min
Lab File: BN015641.D
Acq: 21 Jul 2021 12:19



Tgt Ion:	276	Resp:	23021
Ion	Ratio	Lower	Upper
276	100		
138	31.0	21.8	32.6
277	25.2	20.0	30.0



Abundance Scan 2267 (22.937 min): BN015643.D\data.ms (-230) (-)





Benzo(b)fluoranthene
Concen: 0.104 ng
RT: 22.985 min Scan# 2281
Delta R.T. 0.048 min
Lab File: BN015641.D
Acq: 21 Jul 2021 12:19


Tgt Ion:	252	Resp:	21382
Ion	Ratio	Lower	Upper
252	100		
253	24.5	17.8	26.8
125	13.3	8.9	13.3

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Abundance Scan 2281 (22.985 min): BN015641.D\data.ms



Abundance Scan 2281 (22.985 min): BN015641.D\data.ms (-2209) (-)





Abundance Scan 2280 (22.981 min): BN015643.D\data.ms (-230) (-)




Benzo(k)fluoranthene
Concen: 0.103 ng
RT: 22.985 min Scan# 2281
Delta R.T. 0.003 min
Lab File: BN015641.D
Acq: 21 Jul 2021 12:19

Tgt Ion:	252	Resp:	21382
Ion	Ratio	Lower	Upper
252	100		
253	24.5	17.8	26.6
125	13.3	8.9	13.3


Abundance Scan 2281 (22.985 min): BN015641.D\data.ms



Abundance Scan 2281 (22.985 min): BN015641.D\data.ms (-2192) (-)





Abundance Scan 2441 (23.532 min): BN015643.D\data.ms (-2383) (-)




Benzo(a)pyrene
Concen: 0.102 ng
RT: 23.539 min Scan# 2443
Delta R.T. 0.007 min
Lab File: BN015641.D
Acq: 21 Jul 2021 12:19

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:252 Resp: 18029
Ion Ratio Lower Upper
252 100
253 25.9 17.6 26.4
125 22.6 10.1 15.1#



Abundance Scan 2443 (23.539 min): BN015641.D\data.ms (-2383) (-)





Abundance Scan 3169 (26.024 min): BN015643.D\data.ms (-3146) (-)




Dibenzo(a,h)anthracene
Concen: 0.127 ng
RT: 26.028 min Scan# 3170
Delta R.T. 0.003 min
Lab File: BN015641.D
Acq: 21 Jul 2021 12:19


Tgt Ion:278 Resp: 18685
Ion Ratio Lower Upper
278 100
139 21.5 14.7 22.1
279 25.3 19.0 28.6



Abundance Scan 3170 (26.028 min): BN015641.D\data.ms (-3146) (-)




Abundance Scan 3376 (26.733 min): BN015643.D\data.ms (-3349)




Benzo(g,h,i)perylene
Concen: 0.120 ng
RT: 26.736 min Scan# 3377
Delta R.T. 0.003 min
Lab File: BN015641.D
Acq: 21 Jul 2021 12:19

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1

Abundance Scan 3377 (26.736 min): BN015641.D\data.ms



Tgt Ion:	276	Resp:	19124
Ion	Ratio	Lower	Upper
276	100		
277	24.7	19.0	28.6
138	24.1	18.5	27.7



Abundance Scan 3377 (26.736 min): BN015641.D\data.ms (-3318)

